



**Department  
of Health**

**Wadsworth  
Center**

# **New York State Biomonitoring Program for Trace Elements**

## **Event #2, 2016**

### **Trace Elements in Whole Blood, Urine, and Serum**

### **September, 2016**

## **Wadsworth Center**

NEW YORK STATE DEPARTMENT OF HEALTH

*Trace Elements Laboratory*



**Event #2, 2016:  
Trace Elements in Whole Blood, Urine, and Serum**

9/6/2016

Dear Laboratory Director,

This report summarizes performance for the second biomonitoring proficiency test (PT) event of 2016 for **Trace Elements in Whole Blood, Urine, and Serum**. One of the key goals of this PT program is to achieve harmonization of biomonitoring data for trace elements.

**Target Value Assignment and Performance Evaluation**

For these PT materials, target values have been assigned for a limited number of trace elements that are gradable under criteria set by the NYS DOH Biomonitoring PT program. See assay-specific narratives for details. Data for additional trace elements are reported and are included here in order to characterize the PT materials more completely. Participant data and descriptive statistics are provided for educational purposes. No target value or acceptable range is implied.

Where the data permit, robust statistics were used to assign target values based on Algorithm A as defined by ISO 13528:2005E "*Statistical methods for use in proficiency testing by inter-laboratory comparisons*" [1]. Acceptable ranges for the "graded elements" are based on consensus criteria and/or those set by the NYS DOH's PT program. For example, some are fixed based on US regulatory guidelines (Pb, Cd) while for other elements the criteria are based on a consensus of the Network of PT scheme organizers for trace elements in occupational and environmental laboratory medicine [2]. Quality specifications are element and matrix specific; full details are provided under each element specific narrative.

A confidential, three-digit code number assigned by PT program staff identifies all laboratory participants.

Samples for the next PT event (Event #3 of 2016) will be shipped October 12<sup>th</sup> 2016. Comments about this report may be directed to [trel@health.ny.gov](mailto:trel@health.ny.gov)

Sincerely,

Patrick J. Parsons, PhD  
Chief, Inorganic and Nuclear Chemistry  
Division of Environmental Sciences  
Wadsworth Center

Aubrey L. Galusha, PhD  
Coordinator, Biomonitoring PT Program,  
Division of Environmental Sciences  
Wadsworth Center



**Department  
of Health**

**Wadsworth  
Center**

**Event #2, 2016**

**Trace Elements in  
Whole Blood**

**Wadsworth Center**  
NEW YORK STATE DEPARTMENT OF HEALTH  
*Trace Elements Laboratory*



## Event #2, 2016: Trace Elements in Whole Blood

### PT Materials

Caprine (goat) whole blood was collected in Hospira “empty container” blood bags and preserved with K<sub>2</sub>EDTA. Each unit of whole blood was transferred into polypropylene containers and supplemented with arsenic (As), cadmium (Cd), cobalt (Co), chromium (Cr), mercury (Hg), manganese (Mn), lead (Pb), silver (Ag), barium (Ba), beryllium (Be), copper (Cu), molybdenum (Mo), nickel (Ni), platinum (Pt), antimony (Sb), selenium (Se), tin (Sn), titanium (Ti), thallium (Tl), uranium (U), vanadium (V), tungsten (W), and zinc (Zn). Whole blood samples were homogenized overnight prior to aliquoting 2-mL into polypropylene vials. PT samples were stored at 4°C prior to circulation to laboratories for analysis.

### Graded Elements

Seven elements in whole blood are formally graded: As, Cd, Co, Cr, Hg, Mn and Pb. Target values for the graded elements are assigned to these pools based on (a) the robust mean calculated from data reported by all laboratories, or (b) where a robust mean is not possible, the arithmetic mean after outlier deletion.

### Additional Elements

An additional 24 elements (beyond the seven graded) were reported by at least one participant: Ag, Al, Ba, Be, Bi, Cs, Cu, I, Li, Mo, Ni, Pt, Sb, Se, Sn, Sr, Te, Th, Ti, Tl, U, V, W, and Zn. These data are included here to provide a more complete characterization of the PT materials. All results reported by participant laboratories are tabulated and organized by lab code. The PT data are graphed for visual comparison purposes for all elements where at least five laboratories reported a value greater than the LOD. A statistical summary table is provided for samples where at least two comparable values were reported as above the LOD.

The summary statistics for the additional elements are provided for educational purposes only, i.e., no acceptable response is implied. However, it is expected that each laboratory would wish to investigate a potential source of bias if warranted by these data. Future events might result in additional elements becoming graded if a consensus can be reached regarding desired quality specifications.



# Results for Event #2, 2016 Whole Blood Arsenic (As) Summary Statistics

	Whole Blood As (µg/L)				
	BE16-06	BE16-07	BE16-08	BE16-09	BE16-10
<b>Target (Arithmetic Mean (<math>\bar{x}</math>))</b>	8.35	5.34	18.6	13.2	34.5
<b>Upper Limit</b>	14.35	11.34	24.6	19.2	41.4
<b>Lower Limit</b>	2.35	0.00	12.6	7.2	27.6
<b>Arithmetic SD (s)</b>	3.35	3.50	1.8	3.1	2.4
<b>Arithmetic RSD (%)</b>	40.1	65.5	9.91	23.3	7.19
<b>Number of Sample Measurements (N)</b>	9	9	8	9	8

The acceptable range is based on quality specifications:  $\pm 6 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 6 \mu\text{g/L}$  at concentrations less than or equal to  $30 \mu\text{g/L}$ . These quality specifications were established by New York State Department of Health's Wadsworth Center, the PT Program organizer.








Results for Event #2, 2016
Whole Blood Arsenic (As)
Performance of Participating Laboratories

Table with 7 columns: Lab Code, Method, BE16-06, BE16-07, BE16-08, BE16-09, BE16-10. Includes a Target row and data for labs 103, 110, 147, 200, 293, 391, 596, 597, and 598. Red arrows indicate values above target.

Based on the grading criteria for As in Whole Blood, 88% of results were satisfactory, with 1 of the 9 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #2, 2016: Whole Blood As



**Legend:**

○ CHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the arithmetic mean of all laboratories. Gray area = acceptable range based on quality specifications:

±6 µg/L or ±20% around the target value, whichever is greater; thus, it is fixed at ±6 µg/L at concentrations less than or equal to 30 µg/L.



# Results for Event #2, 2016 Whole Blood Cadmium (Cd) Summary Statistics

## Whole Blood Cd (µg/L)

	BE16-06	BE16-07	BE16-08	BE16-09	BE16-10
<b>Target (Robust Mean (x*))</b>	7.59	5.10	3.91	0.891	1.69
<b>Upper Limit</b>	8.73	6.10	4.91	1.891	2.69
<b>Lower Limit</b>	6.45	4.10	2.91	0.000	0.69
<b>Robust SD (s*)</b>	0.57	0.33	0.25	0.161	0.16
<b>Robust RSD (%)</b>	7.62	6.52	6.62	18.0	9.86
<b>Number of Sample Measurements (N)</b>	15	15	15	15	15
<b>Standard Uncertainty (u)</b>	0.186	0.107	0.083	0.052	0.053

The acceptable range is based on quality specifications:  $\pm 1 \mu\text{g/L}$  or  $\pm 15\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1 \mu\text{g/L}$  at concentrations less than or equal to  $6.7 \mu\text{g/L}$ . These quality specifications are based on those used by US OSHA for occupational exposure.










Results for Event #2, 2016
Whole Blood Cadmium (Cd)
Performance of Participating Laboratories

Table with 7 columns: Lab Code, Method, BE16-06, BE16-07, BE16-08, BE16-09, BE16-10. Includes a Target row and 15 data rows for various lab codes and methods.

Based on the grading criteria for Cd in Whole Blood, 98% of results were satisfactory, with 0 of the 15 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #2, 2016: Whole Blood Cd



### Legend:

○ CHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

$\pm 1 \mu\text{g/L}$  or  $\pm 15\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1 \mu\text{g/L}$  at concentrations less than or equal to  $6.7 \mu\text{g/L}$ .



# Results for Event #2, 2016 Whole Blood Cobalt (Co) Summary Statistics

Whole Blood Co (µg/L)					
	BE16-06	BE16-07	BE16-08	BE16-09	BE16-10
<b>Target (Arithmetic Mean (<math>\bar{x}</math>))</b>	10.1	4.06	6.73	15.3	9.66
<b>Upper Limit</b>	12.2	5.56	8.23	18.4	11.59
<b>Lower Limit</b>	8.1	2.56	5.23	12.2	7.73
<b>Arithmetic SD (s)</b>	1.1	0.31	0.46	1.0	0.67
<b>Arithmetic RSD (%)</b>	11.1	7.75	6.90	6.83	7.02
<b>Number of Sample Measurements (N)</b>	8	8	8	8	8

The acceptable range is based on quality specifications:  $\pm 1.5 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1.5 \mu\text{g/L}$  at concentrations less than or equal to  $7.5 \mu\text{g/L}$ . These quality specifications were established based on discussions with the US FDA, and represent a consensus from a network of Trace Element PT program organizers



Results for Event #2, 2016
Whole Blood Cobalt (Co)
Performance of Participating Laboratories


Table with 7 columns: Lab Code, Method, BE16-06, BE16-07, BE16-08, BE16-09, BE16-10. Includes a Target row and 8 laboratory rows with their respective results.

Based on the grading criteria for Co in Whole Blood, 100% of results were satisfactory, with 0 of the 8 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.




# Results for Event #2, 2016: Whole Blood Co


BE16-06




BE16-07




BE16-08



BE16-09



BE16-10



### Legend:

○ CHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the arithmetic mean of all laboratories. Gray area = acceptable range based on quality specifications:

±1.5 µg/L or ±20% around the target value, whichever is greater; thus, it is fixed at ±1.5 µg/L at concentrations less than or equal to 7.5 µg/L.



# Results for Event #2, 2016 Whole Blood Chromium (Cr) Summary Statistics

	Whole Blood Cr (µg/L)				
	BE16-06	BE16-07	BE16-08	BE16-09	BE16-10
<b>Target (Arithmetic Mean (<math>\bar{x}</math>))</b>	4.75	7.00	11.6	3.09	1.39
<b>Upper Limit</b>	6.75	9.00	13.9	5.09	3.39
<b>Lower Limit</b>	2.75	5.00	9.2	1.09	0.00
<b>Arithmetic SD (s)</b>	1.01	0.80	1.6	1.22	1.25
<b>Arithmetic RSD (%)</b>	21.3	11.5	14.4	39.4	89.9
<b>Number of Sample Measurements (N)</b>	8	8	8	7	7

The acceptable range is based on quality specifications:  $\pm 2 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 2 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ . These quality specifications were established based on discussions with the US FDA, and represent a consensus from a network of Trace Element PT program organizers








### Results for Event #2, 2016 Whole Blood Chromium (Cr) Performance of Participating Laboratories

Whole Blood Cr (µg/L)						
Lab Code	Method	BE16-06	BE16-07	BE16-08	BE16-09	BE16-10
	<b>Target</b>	<b>4.75</b>	<b>7.00</b>	<b>11.6</b>	<b>3.09</b>	<b>1.39</b>
103	DRC/CC-ICP-MS	4.50	6.36	10.7	2.34	0.773
110	DRC/CC-ICP-MS	4.7	6.9	11.0	4.5	1.0
147	DRC/CC-ICP-MS	4.12	6.45	11.1	2.29	0.889
293	DRC/CC-ICP-MS	5.72	7.28	13.36	3.28	2.3
391	DRC/CC-ICP-MS	3.715	6.383	12.505	1.897	0.014
596	ICP-MS	4.98	8.52	9.29	<LOD	<LOD
597	DRC/CC-ICP-MS	3.66	6.35	10.5	2.35	0.98
598	DRC/CC-ICP-MS	6.62	7.8	14.4 ↑	5.02	3.78 ↑

Based on the grading criteria for Cr in Whole Blood, 95% of results were satisfactory, with 1 of the 8 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #2, 2016: Whole Blood Cr



### Legend:

○ CHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the arithmetic mean of all laboratories.  
Gray area = acceptable range based on quality specifications:

±2 µg/L or ±20% around the target value, whichever is greater; thus, it is fixed at ±2 µg/L at concentrations less than or equal to 10 µg/L.





# Results for Event #2, 2016 Whole Blood Mercury (Hg) Summary Statistics

## Whole Blood Hg (µg/L)

	BE16-06	BE16-07	BE16-08	BE16-09	BE16-10
<b>Target (Robust Mean (x*))</b>	0.949	25.6	5.83	16.0	2.40
<b>Upper Limit</b>	3.949	33.2	8.83	20.8	5.40
<b>Lower Limit</b>	0.000	17.9	2.83	11.2	0.00
<b>Robust SD (s*)</b>	0.205	2.5	0.29	2.0	0.27
<b>Robust RSD (%)</b>	21.6	9.98	5.03	12.9	11.6
<b>Number of Sample Measurements (N)</b>	14	14	14	14	14
<b>Standard Uncertainty (u)</b>	0.068	0.853	0.098	0.694	0.093

The acceptable range is based on quality specifications:  $\pm 3 \mu\text{g/L}$  or  $\pm 30\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 3 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ . These quality specifications were established by New York State Department of Health's Wadsworth Center, the PT Program organizer.








Results for Event #2, 2016
Whole Blood Mercury (Hg)
Performance of Participating Laboratories

Table with 7 columns: Lab Code, Method, BE16-06, BE16-07, BE16-08, BE16-09, BE16-10. Includes a Target row and 14 laboratory rows with numerical results and red arrows indicating high values.

Based on the grading criteria for Hg in Whole Blood, 95% of results were satisfactory, with 1 of the 14 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #2, 2016: Whole Blood Hg



### Legend:

○ CHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

$\pm 3 \mu\text{g/L}$  or  $\pm 30\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 3 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ .



Results for Event #2, 2016  
Whole Blood Manganese (Mn)  
Summary Statistics

	Whole Blood Mn ( $\mu\text{g/L}$ )				
	BE16-06	BE16-07	BE16-08	BE16-09	BE16-10
<b>Target (Robust Mean (<math>x^*</math>))</b>	19.4	15.3	16.9	23.6	29.9
<b>Upper Limit</b>	22.7	18.3	19.9	27.6	35.0
<b>Lower Limit</b>	16.1	12.3	13.9	19.6	24.8
<b>Robust SD (<math>s^*</math>)</b>	2.1	1.1	1.2	1.9	2.7
<b>Robust RSD (%)</b>	11.2	7.78	7.5	8.42	9.16
<b>Number of Sample Measurements (N)</b>	10	10	10	10	10
<b>Standard Uncertainty (<math>u</math>)</b>	0.861	0.473	0.503	0.786	1.08

The acceptable range is based on quality specifications:

$\pm 3 \mu\text{g/L}$  or  $\pm 17\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 3 \mu\text{g/L}$  at concentrations less than or equal to  $17 \mu\text{g/L}$ . These quality specifications were recently proposed by a network of Trace Element PT program organizers (Praamsma M, et al. An assessment of clinical laboratory performance for the determination of manganese in blood and urine. Clinical Chemistry and Laboratory Medicine. 2016 In press.)








Results for Event #2, 2016
Whole Blood Manganese (Mn)
Performance of Participating Laboratories

Table with 7 columns: Lab Code, Method, BE16-06, BE16-07, BE16-08, BE16-09, BE16-10. Includes a Target row and 10 laboratory rows with numerical values and red arrows indicating performance relative to targets.

Based on the grading criteria for Mn in Whole Blood, 88% of results were satisfactory, with 1 of the 10 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #2, 2016: Whole Blood Mn



### Legend:

○ CHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

$\pm 3 \mu\text{g/L}$  or  $\pm 17\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 3 \mu\text{g/L}$  at concentrations less than or equal to  $17 \mu\text{g/L}$ .



# Results for Event #2, 2016 Whole Blood Lead (Pb) Summary Statistics

## Whole Blood Pb (µg/dL)

	BE16-06	BE16-07	BE16-08	BE16-09	BE16-10
<b>Target (Robust Mean (x*))</b>	0.420	0.680	0.940	2.40	1.40
<b>Upper Limit</b>	2.420	2.680	2.940	4.40	3.40
<b>Lower Limit</b>	0.000	0.000	0.000	0.40	0.00
<b>Robust SD (s*)</b>	0.090	0.130	0.140	0.20	0.10
<b>Robust RSD (%)</b>	21.6	19.8	15.6	10.3	12.3
<b>Number of Sample Measurements (N)</b>	13	13	14	15	15
<b>Standard Uncertainty (u)</b>	0.031	0.046	0.049	0.082	0.057

The acceptable range is based on quality specifications:  $\pm 2 \mu\text{g/dL}$  or  $\pm 10\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 2 \mu\text{g/L}$  at concentrations less than or equal to  $20 \mu\text{g/dL}$ . These quality specifications are recommended by the Clinical Laboratory Standards Institute (CLSI, C40-A2) and have been proposed for use in proficiency testing programs approved under CLIA by the Centers for Medicare and Medicaid Services (CMS) in the USA. (<http://shop.clsi.org/C40.html>)



Results for Event #2, 2016
Whole Blood Lead (Pb)
Performance of Participating Laboratories

Table with 7 columns: Lab Code, Method, BE16-06, BE16-07, BE16-08, BE16-09, BE16-10. Includes a Target row and 15 data rows for various lab codes and methods.


Based on the grading criteria for Pb in Whole Blood, 100% of results were satisfactory, with 0 of the 15 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.






# Results for Event #2, 2016: Whole Blood Pb


BE16-06




BE16-07




BE16-08



BE16-09



BE16-10



### Legend:

○ CHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

$\pm 2 \mu\text{g/dL}$  or  $\pm 10\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 2 \mu\text{g/dL}$  at concentrations less than or equal to  $20 \mu\text{g/dL}$ .



## Results for Event #2, 2016 Additional Elements in Whole Blood: Selenium (Se)






Whole Blood Se (µg/L)						
Lab Code	Method	BE16-06	BE16-07	BE16-08	BE16-09	BE16-10
103	DRC/CC-ICP-MS	277	359	338	269	330
107	DRC/CC-ICP-MS	330	400	380	330	380
110	DRC/CC-ICP-MS	293	370	339	292	342
147	ICP-MS	288	361	360	284	339
597	DRC/CC-ICP-MS	272	442	423	360	395

Summary Statistics					
	BE16-06	BE16-07	BE16-08	BE16-09	BE16-10
Arithmetic Mean ( $\bar{x}$ )	292	386	368	307	357
Arithmetic SD (s)	22	35	35	37	28
Arithmetic RSD (%)	7.82	9.09	9.58	12.1	7.98
Number of Sample Measurements (N)	5	5	5	5	5

\*Denotes a statistical Outlier.



# Results for Event #2, 2016: Whole Blood Se



### Legend:

○ CHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #2, 2016 Additional Elements in Whole Blood: Vanadium (V)






Whole Blood V (µg/L)						
Lab Code	Method	BE16-06	BE16-07	BE16-08	BE16-09	BE16-10
110	DRC/CC-ICP-MS	23.7	8.3	4.2	1.3	16
147	DRC/CC-ICP-MS	23.1	7.81	4.27	1.19	15.9
596	HR-ICP-MS	18.5	6.18	3.20	0.835	11.0
597	DRC/CC-ICP-MS	16.8	6.90	3.58	0.99	13.3
599	DRC/CC-ICP-MS	35.8	12.4	6.37	1.20	25.4

Summary Statistics					
	BE16-06	BE16-07	BE16-08	BE16-09	BE16-10
Arithmetic Mean ( $\bar{x}$ )	23.5	8.31	4.32	1.10	16.3
Arithmetic SD (s)	7.4	2.42	1.22	0.18	5.4
Arithmetic RSD (%)	31.5	29.1	28.3	16.9	33.5
Number of Sample Measurements (N)	5	5	5	5	5

\*Denotes a statistical Outlier.



# Results for Event #2, 2016: Whole Blood V



### Legend:

○ CHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #2, 2016 Additional Elements in Whole Blood: Barium (Ba)

Whole Blood Ba (µg/L)						
Lab Code	Method	BE16-06	BE16-07	BE16-08	BE16-09	BE16-10
110	ICP-MS	30.5	18.9	26.6	24.0	16.9
147	ICP-MS	27.9	17.6	26.1	22.9	16.6
596	HR-ICP-MS	25.1	15.1	*11.0	19.4	13.4
599	DRC/CC-ICP-MS	27.8	16.9	25.3	19.0	15.6

Summary Statistics						
	BE16-06	BE16-07	BE16-08	BE16-09	BE16-10	
Arithmetic Mean ( $\bar{x}$ )	27.8	17.1	26.0	21.3	15.6	
Arithmetic SD (s)	2.2	1.5	0.6	2.4	1.5	
Arithmetic RSD (%)	7.92	9.24	2.52	11.7	10.1	
Number of Sample Measurements (N)	4	4	3	4	4	

\*Denotes a statistical Outlier.



## Results for Event #2, 2016 Additional Elements in Whole Blood: Beryllium (Be)

Whole Blood Be (µg/L)						
Lab Code	Method	BE16-06	BE16-07	BE16-08	BE16-09	BE16-10
110	ICP-MS	2.2	3.2	0.2	0.8	1.2
147	ICP-MS	2.21	2.83	<1.17	<1.17	<1.17
599	DRC/CC-ICP-MS	*3.14	4.23	0.235	0.742	1.51

Summary Statistics						
	BE16-06	BE16-07	BE16-08	BE16-09	BE16-10	
Arithmetic Mean ( $\bar{x}$ )	2.20	3.42	0.217	0.771	1.35	
Arithmetic SD (s)	0.00	0.72	0.024	0.041	0.21	
Arithmetic RSD (%)	0.32	21.2	11.3	5.31	16.1	
Number of Sample Measurements (N)	2	3	2	2	2	

\*Denotes a statistical Outlier.



## Results for Event #2, 2016 Additional Elements in Whole Blood: Cesium (Cs)

Whole Blood Cs (µg/L)						
Lab Code	Method	BE16-06	BE16-07	BE16-08	BE16-09	BE16-10
110	ICP-MS	0.3	0.3	0.3	0.3	0.3
599	DRC/CC-ICP-MS	0.395	0.219	0.245	0.131	0.175

Summary Statistics						
	BE16-06	BE16-07	BE16-08	BE16-09	BE16-10	
Arithmetic Mean ( $\bar{x}$ )	0.347	0.259	0.272	0.215	0.237	
Arithmetic SD (s)	0.067	0.057	0.038	0.119	0.088	
Arithmetic RSD (%)	19.3	22.0	14.2	55.4	37.2	
Number of Sample Measurements (N)	2	2	2	2	2	

\*Denotes a statistical Outlier.





## Results for Event #2, 2016 Additional Elements in Whole Blood: Copper (Cu)

### Whole Blood Cu (µg/L)

Lab Code	Method	BE16-06	BE16-07	BE16-08	BE16-09	BE16-10
110	ICP-MS	1304	1507	1564	1472	1466
147	ICP-MS	1264	1461	1620	1423	1436
596	ICP-AES/OES	1210	1410	1800	1600	1640
597	DRC/CC-ICP-MS	947	1257	1397	1230	1189

### Summary Statistics

	BE16-06	BE16-07	BE16-08	BE16-09	BE16-10
Arithmetic Mean ( $\bar{x}$ )	1181	1408	1595	1431	1432
Arithmetic SD (s)	160	108	166	153	185
Arithmetic RSD (%)	13.6	7.71	10.4	10.7	12.9
Number of Sample Measurements (N)	4	4	4	4	4

\*Denotes a statistical Outlier.



## Results for Event #2, 2016 Additional Elements in Whole Blood: Molybdenum (Mo)

Whole Blood Mo (µg/L)						
Lab Code	Method	BE16-06	BE16-07	BE16-08	BE16-09	BE16-10
103	DRC/CC-ICP-MS	28.4	28.8	30.5	78.3	75.2
147	ICP-MS	27.7	28.7	31.3	78.8	76.8
596	HR-ICP-MS	26.8	26.3	29.2	72.6	64.4
599	DRC/CC-ICP-MS	31.7	31.7	35	76.3	85.5

Summary Statistics						
	BE16-06	BE16-07	BE16-08	BE16-09	BE16-10	
Arithmetic Mean ( $\bar{x}$ )	28.6	28.8	31.5	76.5	75.4	
Arithmetic SD (s)	2.1	2.2	2.4	2.8	8.6	
Arithmetic RSD (%)	7.45	7.65	7.90	3.68	11.4	
Number of Sample Measurements (N)	4	4	4	4	4	

\*Denotes a statistical Outlier.



## Results for Event #2, 2016 Additional Elements in Whole Blood: Nickel (Ni)

Whole Blood Ni ( $\mu\text{g/L}$ )						
Lab Code	Method	BE16-06	BE16-07	BE16-08	BE16-09	BE16-10
110	DRC/CC-ICP-MS	2.6	3.9	6.5	14.4	2.6
147	ICP-MS	7.4	14.7	3.96	1.87	6.05
597	DRC/CC-ICP-MS	6.68	14	4.16	2.41	5.54

Summary Statistics						
	BE16-06	BE16-07	BE16-08	BE16-09	BE16-10	
Arithmetic Mean ( $\bar{x}$ )	5.56	10.8	4.87	6.22	4.73	
Arithmetic SD (s)	2.58	6.0	1.41	7.08	1.86	
Arithmetic RSD (%)	46.5	55.6	28.9	113	39.3	
Number of Sample Measurements (N)	3	3	3	3	3	

\*Denotes a statistical Outlier.



## Results for Event #2, 2016 Additional Elements in Whole Blood: Platinum (Pt)

Whole Blood Pt ( $\mu\text{g/L}$ )						
Lab Code	Method	BE16-06	BE16-07	BE16-08	BE16-09	BE16-10
110	ICP-MS	2.91	1.15	1.80	0.28	0.85
596	HR-ICP-MS	3.50	1.31	2.05	*0.338	0.896
599	DRC/CC-ICP-MS	2.94	1.23	1.63	0.281	0.78

Summary Statistics						
	BE16-06	BE16-07	BE16-08	BE16-09	BE16-10	
Arithmetic Mean ( $\bar{x}$ )	3.11	1.23	1.82	0.28	0.841	
Arithmetic SD (s)	0.33	0.08	0.21	0.00	0.058	
Arithmetic RSD (%)	10.6	6.50	11.5	0.252	6.93	
Number of Sample Measurements (N)	3	3	3	2	3	

\*Denotes a statistical Outlier.



## Results for Event #2, 2016 Additional Elements in Whole Blood: Antimony (Sb)

Whole Blood Sb ( $\mu\text{g/L}$ )						
Lab Code	Method	BE16-06	BE16-07	BE16-08	BE16-09	BE16-10
103	DRC/CC-ICP-MS	2.24	0.504	1.05	0.390	0.0880
110	ICP-MS	1.92	0.54	1.00	0.35	0.14
147	ICP-MS	1.86	0.548	1.07	0.36	0.164

Summary Statistics						
	BE16-06	BE16-07	BE16-08	BE16-09	BE16-10	
Arithmetic Mean ( $\bar{x}$ )	2.0	0.530	1.04	0.366	0.130	
Arithmetic SD (s)	0.2	0.023	0.03	0.020	0.038	
Arithmetic RSD (%)	10.1	4.41	3.46	5.67	29.7	
Number of Sample Measurements (N)	3	3	3	3	3	

\*Denotes a statistical Outlier.



## Results for Event #2, 2016 Additional Elements in Whole Blood: Tin (Sn)

Whole Blood Sn ( $\mu\text{g/L}$ )						
Lab Code	Method	BE16-06	BE16-07	BE16-08	BE16-09	BE16-10
110	ICP-MS	10.4	6.1	2.8	5.0	1.2
147	ICP-MS	9.43	5.87	2.64	4.77	1.06
596	ICP-MS	8.63	5.11	2.40	4.24	0.834

Summary Statistics						
	BE16-06	BE16-07	BE16-08	BE16-09	BE16-10	
Arithmetic Mean ( $\bar{x}$ )	9.48	5.69	2.61	4.67	1.03	
Arithmetic SD (s)	0.88	0.51	0.20	0.38	0.18	
Arithmetic RSD (%)	9.34	9.10	7.70	8.34	17.9	
Number of Sample Measurements (N)	3	3	3	3	3	

\*Denotes a statistical Outlier.



## Results for Event #2, 2016 Additional Elements in Whole Blood: Thallium (TI)

Whole Blood TI (µg/L)						
Lab Code	Method	BE16-06	BE16-07	BE16-08	BE16-09	BE16-10
103	DRC/CC-ICP-MS	6.41	11.3	1.05	0.560	2.19
110	ICP-MS	6.3	10.8	1.0	0.6	2.2
147	ICP-MS	5.29	9.59	0.981	0.487	2.03
596	HR-ICP-MS	5.48	8.98	0.936	0.467	1.86

Summary Statistics						
	BE16-06	BE16-07	BE16-08	BE16-09	BE16-10	
Arithmetic Mean ( $\bar{x}$ )	5.87	10.1	0.991	0.528	2.06	
Arithmetic SD (s)	0.56	1.0	0.047	0.062	0.16	
Arithmetic RSD (%)	9.66	10.5	4.75	11.7	7.73	
Number of Sample Measurements (N)	4	4	4	4	4	

\*Denotes a statistical Outlier.



## Results for Event #2, 2016 Additional Elements in Whole Blood: Uranium (U)

Whole Blood U ( $\mu\text{g/L}$ )						
Lab Code	Method	BE16-06	BE16-07	BE16-08	BE16-09	BE16-10
103	DRC/CC-ICP-MS	0.0200	0.0222	0.0520	<0.00748	0.0421
110	ICP-MS	0.03	0.03	0.06	<0.02	0.05
147	ICP-MS	0.0142	0.0226	0.0486	<0.0136	0.04612
596	HR-ICP-MS	0.0355	0.0396	0.0705	0.0181	0.0593

Summary Statistics						
	BE16-06	BE16-07	BE16-08	BE16-09	BE16-10	
Arithmetic Mean ( $\bar{x}$ )	0.024	0.028	0.057	0.018	0.049	
Arithmetic SD (s)	0.009	0.008	0.009	NA	0.007	
Arithmetic RSD (%)	38.5	28.5	16.8	NA	14.8	
Number of Sample Measurements (N)	4	4	4	1	4	

\*Denotes a statistical Outlier.





## Results for Event #2, 2016 Additional Elements in Whole Blood: Tungsten (W)

### Whole Blood W (µg/L)

Lab Code	Method	BE16-06	BE16-07	BE16-08	BE16-09	BE16-10
110	ICP-MS	1.0	00.5	9.4	2.8	5.7
200	ICP-MS	*1.3	0.6	10.8	3.4	6.6
596	HR-ICP-MS	1.00	0.395	9.12	2.57	5.22

### Summary Statistics

	BE16-06	BE16-07	BE16-08	BE16-09	BE16-10
Arithmetic Mean ( $\bar{x}$ )	1.00	0.498	9.77	2.92	5.84
Arithmetic SD (s)	0.00	0.102	0.90	0.42	0.70
Arithmetic RSD (%)	0.00	20.5	9.2	14.6	11.9
Number of Sample Measurements (N)	2	3	3	3	3

\*Denotes a statistical Outlier.



## Results for Event #2, 2016 Additional Elements in Whole Blood: Zinc (Zn)

Whole Blood Zn (µg/L)						
Lab Code	Method	BE16-06	BE16-07	BE16-08	BE16-09	BE16-10
110	ICP-MS	1992	2207	2522	2624	2907
147	ICP-MS	2000	2268	2706	2686	3039
596	ICP-AES/OES	1950	2130	2490	2450	2860
597	DRC/CC-ICP-MS	*1479	1893	2237	2316	2523

Summary Statistics						
	BE16-06	BE16-07	BE16-08	BE16-09	BE16-10	
Arithmetic Mean ( $\bar{x}$ )	1980	2124	2488	2519	2832	
Arithmetic SD (s)	26	164	192	168	219	
Arithmetic RSD (%)	1.35	7.73	7.75	6.67	7.75	
Number of Sample Measurements (N)	3	4	4	4	4	

\*Denotes a statistical Outlier.



## Results for Event #2, 2016

### Additional Elements in Whole Blood

#### Whole Blood Ag (µg/L)

Lab Code	Method	BE16-06	BE16-07	BE16-08	BE16-09	BE16-10
147	ICP-MS	9.02	0.454	1.92	7.49	3.75

#### Whole Blood Al (µg/L)

Lab Code	Method	BE16-06	BE16-07	BE16-08	BE16-09	BE16-10
147	ICP-MS	<9.17	<9.17	<9.17	<9.17	<9.17
596	ICP-AES/OES	263	263	288	338	400

#### Whole Blood Bi (µg/L)

Lab Code	Method	BE16-06	BE16-07	BE16-08	BE16-09	BE16-10
147	ICP-MS	<0.0836	<0.0836	<0.0836	<0.0836	<0.0836

#### Whole Blood I (µg/L)

Lab Code	Method	BE16-06	BE16-07	BE16-08	BE16-09	BE16-10
147	ICP-MS	60.1	62.7	64.4	72.2	68.5

#### Whole Blood Li (µg/L)

Lab Code	Method	BE16-06	BE16-07	BE16-08	BE16-09	BE16-10
147	ICP-MS	0.874	0.93	0.978	1.02	1.15

#### Whole Blood Sr (µg/L)

Lab Code	Method	BE16-06	BE16-07	BE16-08	BE16-09	BE16-10
103	DRC/CC-ICP-MS	36.9	35.5	35.0	35.7	32.9

#### Whole Blood Te (µg/L)

Lab Code	Method	BE16-06	BE16-07	BE16-08	BE16-09	BE16-10
147	ICP-MS	<0.128	<0.128	<0.128	<0.128	<0.128

#### Whole Blood Th (µg/L)

Lab Code	Method	BE16-06	BE16-07	BE16-08	BE16-09	BE16-10
147	ICP-MS	<0.0148	<0.0148	<0.0148	<0.0148	<0.0148

#### Whole Blood Ti (µg/L)

Lab Code	Method	BE16-06	BE16-07	BE16-08	BE16-09	BE16-10
599	DRC/CC-ICP-MS	6.00	10.1	0.984	0.425	2.08



**Department  
of Health**

**Wadsworth  
Center**

**Event #2, 2016**

**Trace Elements in  
Urine**

**Wadsworth Center**  
NEW YORK STATE DEPARTMENT OF HEALTH  
*Trace Elements Laboratory*



## Event #2, 2016: Trace Elements in Urine

### PT Materials

Urine was collected from volunteer donors into polyethylene containers and stored at 4°C. Following collection, urine was acidified to 1% (v/v) with nitric acid and mixed with a sulfamic acid solution (stock solution contained 200 mg/mL sulfamic acid and 10% (v/v) Triton-X 100) to a final concentration of 1% (v/v) to stabilize Hg. Urine was stored frozen at -80°C pending further preparation. The urine was thawed at room temperature and precipitated salts removed by centrifugation. Urine supernatants were combined and subsequently separated into five pools. Each urine pool was supplemented with arsenic (As), barium (Ba), beryllium (Be), cadmium (Cd), mercury (Hg), manganese (Mn), lead (Pb), thallium (Tl), uranium (U), aluminum (Al), cobalt (Co), chromium (Cr), cesium (Cs), copper (Cu), molybdenum (Mo), nickel (Ni), platinum (Pt), antimony (Sb), selenium (Se), tin (Sn), tellurium (Te), vanadium (V), tungsten (W), and zinc (Zn) and stirred overnight to ensure thorough mixing prior to aliquoting 10-mL into polypropylene vials. PT samples were stored at -80°C until the week of the PT event, when they were thawed at 4°C prior to circulation to laboratories for analysis.

### Graded Elements

Nine elements in urine are formally graded: As, Ba, Be, Cd, Hg, Mn, Pb, Tl, and U. Target values for the graded elements are assigned to these pools based on the robust mean calculated from data reported by all laboratories.

### Additional Elements

An additional 23 elements (beyond the nine graded) were reported by at least one participant: Al, B, Bi, Co, Cr, Cs, Cu, Fe, I, Li, Mo, Ni, Pt, Sb, Se, Sn, Sr, Te, Th, Ti, V, W, Zn. These data are included here to provide a more complete characterization of the PT materials. All results reported by participant laboratories are tabulated and organized by lab code. The PT data are graphed for visual comparison purposes for all elements where at least five laboratories reported a value greater than the LOD. A statistical summary table is provided for samples where at least two comparable values were reported as above the LOD.

The summary statistics for the additional elements are provided for educational purposes only, i.e., no acceptable response is implied. However, it is expected that each laboratory would wish to investigate a potential source of bias if warranted by these data. Future events might result in additional elements becoming graded if a consensus can be reached regarding desired quality specifications.



# Results for Event #2, 2016 Urine Arsenic (As)

## Summary Statistics

	Urine As (µg/L)				
	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10
<b>Target (Robust Mean (x*))</b>	8.05	96.9	58.3	43.0	25.7
<b>Upper Limit</b>	14.05	116.3	70.0	51.6	31.7
<b>Lower Limit</b>	2.05	77.5	46.7	34.4	19.7
<b>Robust SD (s*)</b>	0.98	7.4	3.5	2.8	2.3
<b>Robust RSD (%)</b>	12.1	7.66	6.04	6.69	9.08
<b>Number of Sample Measurements (N)</b>	19	19	19	19	19
<b>Standard Uncertainty (u)</b>	0.281	2.13	1.01	0.825	0.670

The acceptable range is based on quality specifications:  $\pm 6 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 6 \mu\text{g/L}$  at concentrations less than or equal to  $30 \mu\text{g/L}$ . These quality specifications are based on the same criteria used by the US Centers for Disease Control Prevention (CDC) for public health labs participating in the Laboratory Response Network (LRN) PT program for Toxic Metals.



# Results for Event #2, 2016 Urine Arsenic (As)

## Performance of Participating Laboratories


Urine As (µg/L)						
Lab Code	Method	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10
	Target	8.05	96.9	58.3	43.0	25.7
103	DRC/CC-ICP-MS	7.57	97.1	57.7	42.9	25.6
107	DRC/CC-ICP-MS	7.3	90	54	39	24
110	DRC/CC-ICP-MS	9.6	97.9	58.3	43.3	26.4
116	DRC/CC-ICP-MS	7.37	93.5	55.6	40.6	24.4
147	ICP-MS	7.47	90.6	54.5	42.2	24.3
200	ICP-MS	9.0	99.8	60.0	43.5	27.0
293	DRC/CC-ICP-MS	7.47	94.38	56.0	42.55	24.87
324	HR-ICP-MS	7.66	108.399	64.83	46.928	28.408
391	DRC/CC-ICP-MS	8.24	112.091	66.6	47.839	29.839
399	DRC/CC-ICP-MS	7.95	102	60.9	44.4	26.3
596	ICP-MS	7.36	90.5	57.2	41.6	24.1
596	HR-ICP-MS	7.93	88.9	55.7	38.0	21.6
597	DRC/CC-ICP-MS	6.84	88.1	56.7	41.1	24.7
598	ICP-MS	8.52	97	57.5	43	23.8
599	DRC/CC-ICP-MS	8.80	114.5	68.72	51.03	28.41
604	DRC/CC-ICP-MS	6.28	90.8	54.2	39.3	22.8
605	ICP-MS	9.77	95.0	58.5	44.0	26.5
606	DRC/CC-ICP-MS	9.64	101	63.9	45.0	28.6
632	DRC/CC-ICP-MS	9.59	101	61.4	44.8	28

Based on the grading criteria for As in Urine, 100% of results were satisfactory, with 0 of the 19 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.




# Results for Event #2, 2016: Urine As


UE16-06




UE16-07




UE16-08



UE16-09



UE16-10



### Legend:

○ CHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

±6 µg/L or ±20% around the target value, whichever is greater; thus, it is fixed at ±6 µg/L at concentrations less than or equal to 30 µg/L.





## Results for Event #2, 2016 Urine Barium (Ba)

### Summary Statistics

	Urine Ba ( $\mu\text{g/L}$ )				
	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10
<b>Target (Robust Mean (<math>x^*</math>))</b>	1.27	5.07	1.05	3.31	4.77
<b>Upper Limit</b>	2.27	6.08	2.05	4.31	5.77
<b>Lower Limit</b>	0.27	4.05	0.05	2.31	3.77
<b>Robust SD (<math>s^*</math>)</b>	0.09	0.28	0.12	0.18	0.29
<b>Robust RSD (%)</b>	7.28	5.58	11.4	5.43	6.16
<b>Number of Sample Measurements (N)</b>	13	13	13	13	13
<b>Standard Uncertainty (<math>u</math>)</b>	0.032	0.098	0.041	0.062	0.102

The acceptable range is based on quality specifications:

$\pm 1 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1 \mu\text{g/L}$  at concentrations less than or equal to  $5 \mu\text{g/L}$ . These quality specifications are based on the same criteria used by the US Centers for Disease Control Prevention (CDC) for public health labs participating in the Laboratory Response Network (LRN) PT program for Toxic Metals.



# Results for Event #2, 2016 Urine Barium (Ba)

## Performance of Participating Laboratories


Urine Ba (µg/L)						
Lab Code	Method	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10
	<b>Target</b>	<b>1.27</b>	<b>5.07</b>	<b>1.05</b>	<b>3.31</b>	<b>4.77</b>
107	ICP-MS	1.2	4.8	1.0	3.2	4.6
110	ICP-MS	1.28	5.11	1.02	3.35	4.82
116	DRC/CC-ICP-MS	1.29	5.05	1.05	3.24	4.85
147	ICP-MS	1.31	5.21	1.05	3.54	4.88
399	ICP-MS	1.40	5.27	1.12	3.43	4.97
596	ICP-MS	1.26	5.04	1.25	3.27	5.52
597	DRC/CC-ICP-MS	1.10	4.35	0.90	3.10	4.32
598	ICP-MS	1.35	5.57	1.61	3.52	4.55
599	DRC/CC-ICP-MS	1.41	5.29	1.25	3.47	4.88
605	ICP-MS	1.21	4.76	0.959	3.12	4.56
606	ICP-MS	1.13	4.86	0.869	3.09	4.49
607	ICP-MS	1.29	5.09	1.05	3.30	4.74
632	ICP-MS	1.32	5.32	1.09	3.44	5.15

Based on the grading criteria for Ba in Urine, 100% of results were satisfactory, with 0 of the 13 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.




# Results for Event #2, 2016: Urine Ba


UE16-06




UE16-07




UE16-08



UE16-09



UE16-10



### Legend:

○ CHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

±1 µg/L or ±20% around the target value, whichever is greater; thus, it is fixed at ±1 µg/L at concentrations less than or equal to 5 µg/L.



# Results for Event #2, 2016 Urine Beryllium (Be)

## Summary Statistics

	Urine Be (µg/L)				
	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10
<b>Target (Robust Mean (x*))</b>	0.316	1.13	1.40	2.53	0.535
<b>Upper Limit</b>	1.316	2.13	2.40	3.53	1.535
<b>Lower Limit</b>	0.000	0.13	0.40	1.53	0.000
<b>Robust SD (s*)</b>	0.028	0.05	0.12	0.19	0.050
<b>Robust RSD (%)</b>	9.11	4.51	9.15	7.82	9.36
<b>Number of Sample Measurements (N)</b>	10	11	11	11	11
<b>Standard Uncertainty (u)</b>	0.011	0.019	0.048	0.074	0.018

The acceptable range is based on quality specifications:  $\pm 1 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1 \mu\text{g/L}$  at concentrations less than or equal to  $5 \mu\text{g/L}$ . These quality specifications are based on the same criteria used by the US Centers for Disease Control Prevention (CDC) for public health labs participating in the Laboratory Response Network (LRN) PT program for Toxic Metals.



Results for Event #2, 2016
Urine Beryllium (Be)
Performance of Participating Laboratories


Table with 7 columns: Lab Code, Method, UE16-06, UE16-07, UE16-08, UE16-09, UE16-10. Includes a Target row and 11 data rows for various lab codes and methods.

Based on the grading criteria for Be in Urine, 100% of results were satisfactory, with 0 of the 11 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.




# Results for Event #2, 2016: Urine Be


UE16-06




UE16-07




UE16-08



UE16-09



UE16-10



### Legend:

○ CHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

±1 µg/L or ±20% around the target value, whichever is greater; thus, it is fixed at ±1 µg/L at concentrations less than or equal to 5 µg/L.



# Results for Event #2, 2016 Urine Cadmium (Cd)

## Summary Statistics

	Urine Cd (µg/L)				
	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10
<b>Target (Robust Mean (x*))</b>	3.90	0.229	0.589	1.64	0.860
<b>Upper Limit</b>	4.90	1.229	1.589	2.64	1.860
<b>Lower Limit</b>	2.90	0.000	0.000	0.64	0.000
<b>Robust SD (s*)</b>	0.16	0.063	0.079	0.06	0.043
<b>Robust RSD (%)</b>	4.10	27.8	13.5	4.22	5.02
<b>Number of Sample Measurements (N)</b>	19	17	18	19	18
<b>Standard Uncertainty (u)</b>	0.050	0.019	0.023	0.019	0.012

The acceptable range is based on quality specifications:  $\pm 1 \mu\text{g/L}$  or  $\pm 15\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1 \mu\text{g/L}$  at concentrations less than or equal to  $6.6 \mu\text{g/L}$ . These quality specifications are based on the same criteria used by the US Centers for Disease Control Prevention (CDC) for public health labs participating in the Laboratory Response Network (LRN) PT program for Toxic Metals.



Results for Event #2, 2016
Urine Cadmium (Cd)
Performance of Participating Laboratories






Table with 7 columns: Lab Code, Method, UE16-06, UE16-07, UE16-08, UE16-09, UE16-10. Includes a Target row and 19 laboratory data rows.

Based on the grading criteria for Cd in Urine, 98% of results were satisfactory, with 0 of the 19 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.





# Results for Event #2, 2016: Urine Cd



### Legend:

○ CHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

$\pm 1 \mu\text{g/L}$  or  $\pm 15\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1 \mu\text{g/L}$  at concentrations less than or equal to  $6.6 \mu\text{g/L}$ .



# Results for Event #2, 2016 Urine Mercury (Hg)

## Summary Statistics

	Urine Hg (µg/L)				
	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10
<b>Target (Robust Mean (x*))</b>	1.74	3.31	9.80	6.86	15.9
<b>Upper Limit</b>	4.74	6.31	12.80	9.86	20.7
<b>Lower Limit</b>	0.00	0.31	6.80	3.86	11.1
<b>Robust SD (s*)</b>	0.43	0.38	0.63	0.65	1.3
<b>Robust RSD (%)</b>	24.9	11.6	6.47	9.49	8.54
<b>Number of Sample Measurements (N)</b>	12	14	14	14	14
<b>Standard Uncertainty (u)</b>	0.156	0.129	0.212	0.217	0.456

The acceptable range is based on quality specifications:  $\pm 3 \mu\text{g/L}$  or  $\pm 30\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 3 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ . These quality specifications were established by New York State Department of Health's Wadsworth Center, the PT Program organizer.



## Results for Event #2, 2016

### Urine Mercury (Hg)






#### Performance of Participating Laboratories

Urine Hg (µg/L)						
Lab Code	Method	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10
Target		1.74	3.31	9.80	6.86	15.9
103	DRC/CC-ICP-MS	1.46	3.14	9.60	6.68	15.6
107	DRC/CC-ICP-MS	1.3	2.8	8.8	6.1	14
110	ICP-MS	2.36	3.65	10.3	7.25	16.9
147	CV-AAS	1.39	3.12	9.39	6.5	15.5
200	ICP-MS	1.6	3.0	9.6	6.4	15.4
293	ICP-MS	1.7	3.29	9.46	6.73	14.97
391	DRC/CC-ICP-MS	2.509	4.056	11.13	8.53	17.474
596	ICP-MS	1.29	2.82	9.33	5.95	15.7
597	DRC/CC-ICP-MS	2.24	3.78	10.79	7.69	17.61
598	ICP-MS	1.97	3.14	9.2	6.52	13.68
604	DRC/CC-ICP-MS	1.89	3.74	10.6	7.69	17.3
605	ICP-MS	PLC	3.32	10.0	6.92	16.5
606	ICP-MS	<2.00	3.44	10.9	7.41	17.1
632	ICP-MS	1.50	3.32	9.60	6.65	15.9

Based on the grading criteria for Hg in Urine, 100% of results were satisfactory, with 0 of the 14 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #2, 2016: Urine Hg



### Legend:

○ CHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

±3 µg/L or ±30% around the target value, whichever is greater; thus, it is fixed at ±3 µg/L at concentrations less than or equal to 10 µg/L.



# Results for Event #2, 2016 Urine Manganese (Mn) Summary Statistics

	Urine Mn (µg/L)				
	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10
<b>Target (Robust Mean (x*))</b>	0.869	0.348	1.00	1.60	0.502
<b>Upper Limit</b>	1.249	0.728	1.38	1.98	0.882
<b>Lower Limit</b>	0.489	0.000	0.62	1.22	0.122
<b>Robust SD (s*)</b>	0.074	0.111	0.16	0.19	0.090
<b>Robust RSD (%)</b>	8.62	31.9	16.2	12.1	17.9
<b>Number of Sample Measurements (N)</b>	14	13	14	15	14
<b>Standard Uncertainty (u)</b>	0.025	0.038	0.054	0.063	0.030

The acceptable range is based on quality specifications:  $\pm 0.38 \mu\text{g/L}$  or  $\pm 10\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 0.38 \mu\text{g/L}$  at concentrations less than or equal to  $3.8 \mu\text{g/L}$ . These quality specifications were recently proposed by a network of Trace Element PT program organizers (Praamsma M, et al. An assessment of clinical laboratory performance for the determination of manganese in blood and urine. Clinical Chemistry and Laboratory Medicine. 2016 In press.)








Results for Event #2, 2016  
Urine Manganese (Mn)  
Performance of Participating Laboratories

Lab Code	Method	Urine Mn (µg/L)							
		UE16-06	UE16-07	UE16-08	UE16-09	UE16-10			
	<b>Target</b>	<b>0.869</b>	<b>0.348</b>	<b>1.00</b>	<b>1.60</b>	<b>0.502</b>			
103	DRC/CC-ICP-MS	0.781	0.254	0.899	1.52	0.416			
107	DRC/CC-ICP-MS	1.2	0.33	0.96	1.5	0.48			
110	DRC/CC-ICP-MS	0.88	0.31	1.00	1.63	0.50			
116	DRC/CC-ICP-MS	0.860	0.262	0.806	1.37	0.415			
147	DRC/CC-ICP-MS	0.834	0.309	0.89	1.53	0.513			
324	HR-ICP-MS	1.164	0.728	↑	1.123	1.646	0.625		
391	DRC/CC-ICP-MS	0.809	0.732	↑	1.692	↑	2.06	↑	0.553
399	DRC/CC-ICP-MS	0.86	0.332		0.963	1.62		0.544	
596	ICP-MS	0.645	<LOD		0.874	1.35		0.375	
597	DRC/CC-ICP-MS	0.81	0.27		0.83	1.42		0.46	
598	ICP-MS	0.9	0.81	↑	1.59	↑	1.81	0.78	
604	DRC/CC-ICP-MS	0.87	0.3		0.95	1.8		0.5	
605	ICP-MS	PLC	PLC		PLC	1.52		PLC	
606	DRC/CC-ICP-MS	0.892	0.203		1.13	1.68		0.453	
632	DRC/CC-ICP-MS	1.20	0.474		1.27	1.90		0.618	

Based on the grading criteria for Mn in Urine, 92% of results were satisfactory, with 2 of the 15 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #2, 2016: Urine Mn



### Legend:

○ CHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

$\pm 0.38 \mu\text{g/L}$  or  $\pm 10\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 0.38 \mu\text{g/L}$  at concentrations less than or equal to  $3.8 \mu\text{g/L}$ .



# Results for Event #2, 2016 Urine Lead (Pb) Summary Statistics

	Urine Pb (µg/L)				
	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10
<b>Target (Robust Mean (x*))</b>	1.73	8.38	2.48	1.04	4.02
<b>Upper Limit</b>	2.73	10.06	3.48	2.04	5.02
<b>Lower Limit</b>	0.73	6.70	1.48	0.04	3.02
<b>Robust SD (s*)</b>	0.18	0.52	0.21	0.14	0.41
<b>Robust RSD (%)</b>	10.3	6.20	8.77	13.6	10.1
<b>Number of Sample Measurements (N)</b>	18	18	18	18	18
<b>Standard Uncertainty (u)</b>	0.053	0.153	0.064	0.042	0.120

The acceptable range is based on quality specifications:  $\pm 1 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1 \mu\text{g/L}$  at concentrations less than or equal to  $5 \mu\text{g/L}$ . These quality specifications are based on the same criteria used by the US Centers for Disease Control Prevention (CDC) for public health labs participating in the Laboratory Response Network (LRN) PT program for Toxic Metals.





# Results for Event #2, 2016 Urine Lead (Pb)

## Performance of Participating Laboratories


Urine Pb (µg/L)						
Lab Code	Method	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10
	<b>Target</b>	<b>1.73</b>	<b>8.38</b>	<b>2.48</b>	<b>1.04</b>	<b>4.02</b>
103	DRC/CC-ICP-MS	1.74	8.68	2.50	1.01	4.08
107	ICP-MS	1.8	8.9	2.7	1.1	4.20
110	ICP-MS	1.7	8.5	2.5	1.0	4.1
116	ICP-MS	1.72	8.46	2.45	0.971	3.87
147	ICP-MS	1.6	8.12	2.36	0.976	3.75
200	ICP-MS	2.2	8.2	2.2	1.4	4.2
293	ICP-MS	1.6	7.71	2.22	0.58	3.17
324	HR-ICP-MS	1.944	8.024	2.542	1.225	3.81
391	DRC/CC-ICP-MS	1.896	9.606	2.816	1.173	4.466
399	ICP-MS	1.76	9.19	2.62	1.05	4.2
596	ICP-MS	2.00	9.84	3.02	1.13	4.58
597	DRC/CC-ICP-MS	1.46	7.3	2.10	0.92	3.59
598	ICP-MS	1.56	7.88	2.47	1.19	3.41
599	DRC/CC-ICP-MS	1.85	8.30	3.10	1.31	4.41
605	ICP-MS	1.65	8.38	2.43	0.969	3.92
606	ICP-MS	1.53	7.98	2.26	0.911	3.70
607	ICP-MS	1.73	8.60	2.52	1.02	4.01
632	ICP-MS	1.65	8.33	2.44	0.904	10.2 ↑

Based on the grading criteria for Pb in Urine, 98% of results were satisfactory, with 0 of the 18 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.




# Results for Event #2, 2016: Urine Pb


UE16-06




UE16-07




UE16-08



UE16-09



UE16-10



### Legend:

- CHEAR Labs
- Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

$\pm 1 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1 \mu\text{g/L}$  at concentrations less than or equal to  $5 \mu\text{g/L}$ .



# Results for Event #2, 2016 Urine Thallium (TI)

## Summary Statistics

	Urine TI (µg/L)				
	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10
<b>Target (Robust Mean (x*))</b>	2.19	0.475	0.778	1.65	0.227
<b>Upper Limit</b>	2.63	0.675	0.978	1.98	0.427
<b>Lower Limit</b>	1.75	0.275	0.578	1.32	0.027
<b>Robust SD (s*)</b>	0.10	0.024	0.033	0.03	0.016
<b>Robust RSD (%)</b>	4.82	5.20	4.24	2.18	7.27
<b>Number of Sample Measurements (N)</b>	13	13	13	13	13
<b>Standard Uncertainty (u)</b>	0.036	0.0085	0.011	0.012	0.0057

The acceptable range is based on quality specifications:  $\pm 0.2 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 0.2 \mu\text{g/L}$  at concentrations less than or equal to  $1 \mu\text{g/L}$ . These quality specifications are based on the same criteria used by the US Centers for Disease Control Prevention (CDC) for public health labs participating in the Laboratory Response Network (LRN) PT program for Toxic Metals.



# Results for Event #2, 2016 Urine Thallium (TI)






## Performance of Participating Laboratories

Urine TI (µg/L)						
Lab Code	Method	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10
Target		2.19	0.475	0.778	1.65	0.227
103	DRC/CC-ICP-MS	2.14	0.459	0.753	1.64	0.219
107	ICP-MS	2.3	0.50	0.82	1.8	0.25
110	ICP-MS	2.19	0.46	0.77	1.66	0.21
116	ICP-MS	2.19	0.481	0.755	1.66	0.228
147	ICP-MS	2.11	0.462	0.736	1.64	0.221
399	ICP-MS	2.28	0.489	0.830	1.77	0.209
596	ICP-MS	2.49	0.565	0.917	1.81	0.265
598	ICP-MS	2.04	0.54	0.75	1.54	0.26
599	DRC/CC-ICP-MS	2.30	0.474	0.795	1.63	0.227
605	ICP-MS	2.16	0.448	0.767	1.64	0.214
606	ICP-MS	2.08	0.442	0.749	1.57	0.218
607	ICP-MS	2.22	0.462	0.789	1.69	0.232
632	ICP-MS	2.17	0.486	0.801	1.65	0.237

Based on the grading criteria for TI in Urine, 100% of results were satisfactory, with 0 of the 13 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #2, 2016: Urine TI



### Legend:

○ CHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

±0.2 µg/L or ±20% around the target value, whichever is greater; thus, it is fixed at ±0.2 µg/L at concentrations less than or equal to 1 µg/L.



# Results for Event #2, 2016 Urine Uranium (U)

## Summary Statistics

	Urine U (µg/L)				
	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10
<b>Target (Robust Mean (x*))</b>	0.007	0.012	0.019	0.064	0.037
<b>Upper Limit</b>	0.037	0.042	0.049	0.094	0.067
<b>Lower Limit</b>	0.000	0.000	0.000	0.034	0.007
<b>Robust SD (s*)</b>	0.001	0.002	0.002	0.003	0.003
<b>Robust RSD (%)</b>	13.9	19.7	15.4	5.22	9.93
<b>Number of Sample Measurements (N)</b>	11	14	14	14	14
<b>Standard Uncertainty (u)</b>	0.00040	0.00083	0.00098	0.0011	0.0012

The acceptable range is based on quality specifications:  $\pm 0.03 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 0.03 \mu\text{g/L}$  at concentrations less than or equal to  $.15 \mu\text{g/L}$ . These quality specifications are based on the same criteria used by the US Centers for Disease Control Prevention (CDC) for public health labs participating in the Laboratory Response Network (LRN) PT program for Toxic Metals.



# Results for Event #2, 2016 Urine Uranium (U)






## Performance of Participating Laboratories

Urine U (µg/L)						
Lab Code	Method	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10
	Target	0.007	0.012	0.019	0.064	0.037
103	DRC/CC-ICP-MS	0.00739	0.0113	0.0161	0.0632	0.0361
107	ICP-MS	0.0095	0.015	0.022	0.076	0.043
110	ICP-MS	0.010	0.014	0.019	0.065	0.038
116	ICP-MS	0.00714	0.0115	0.0196	0.0601	0.0366
147	ICP-MS	<0.0105	0.0122	0.021	0.0674	0.0395
324	ICP-MS	0.010	0.014	0.022	0.069	0.041
399	ICP-MS	0.008	0.016	0.019	0.064	0.035
596	HR-ICP-MS	0.00290	0.00570	0.0107	0.0610	0.0288
598	ICP-MS	<0.1	0.12 ↑	0.12 ↑	0.15 ↑	0.1 ↑
599	DRC/CC-ICP-MS	<0.02	<0.02	<0.02	<0.02 ↓	<0.02
604	ICP-MS	0.007	0.010	0.017	0.063	0.034
605	ICP-MS	PLC	0.0123	0.0166	0.0630	0.0344
606	ICP-MS	0.007	0.011	0.017	0.064	0.036
607	ICP-MS	0.009	0.014	0.021	0.069	0.041
632	ICP-MS	0.0070	0.0113	0.0185	0.0642	0.0341

Based on the grading criteria for U in Urine, 92% of results were satisfactory, with 1 of the 15 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #2, 2016: Urine U



### Legend:

○ CHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

±0.03 µg/L or ±20% around the target value, whichever is greater; thus, it is fixed at ±0.03 µg/L at concentrations less than or equal to .15 µg/L.





## Results for Event #2, 2016

### Additional Elements in Urine: Cobalt (Co)






Urine Co ( $\mu\text{g/L}$ )						
Lab Code	Method	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10
103	DRC/CC-ICP-MS	2.05	0.780	0.572	1.18	1.61
107	ICP-MS	2.1	0.80	0.62	1.2	1.6
110	ICP-MS	2.05	0.78	0.59	1.17	1.55
116	DRC/CC-ICP-MS	1.94	0.791	0.564	1.09	1.49
147	ICP-MS	2.06	0.784	0.625	1.23	1.61
324	HR-ICP-MS	2.399	1.309	0.891	1.485	1.859
391	DRC/CC-ICP-MS	2.277	0.8	0.679	1.203	1.755
399	DRC/CC-ICP-MS	2.04	0.777	0.577	1.22	1.58
485	HR-ICP-MS	2.1	0.776	0.594	1.19	1.57
596	ICP-MS	1.94	0.775	0.634	1.14	1.49
597	DRC/CC-ICP-MS	1.73	0.66	0.46	1.04	1.42
605	ICP-MS	2.05	0.806	0.610	1.22	1.62
606	ICP-MS	1.95	0.700	0.516	1.11	1.49
632	ICP-MS	2.16	0.870	0.658	1.28	1.69

### Summary Statistics

	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10
<b>Robust Mean (<math>x^*</math>)</b>	2.05	0.785	0.603	1.19	1.58
<b>Robust SD (<math>s^*</math>)</b>	0.11	0.021	0.051	0.05	0.10
<b>Robust RSD (%)</b>	5.62	2.73	8.59	4.45	6.76
<b>Number of Sample Measurements (N)</b>	14	14	14	14	14
<b>Standard Uncertainty (<math>u</math>)</b>	0.038	0.0071	0.017	0.017	0.035



# Results for Event #2, 2016: Urine Co



### Legend:

- CHEAR Labs    ● Other Labs
- Horizontal purple line = robust mean of all laboratories.
- Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #2, 2016

### Additional Elements in Urine: Chromium (Cr)






Urine Cr ( $\mu\text{g/L}$ )						
Lab Code	Method	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10
103	DRC/CC-ICP-MS	11.7	2.87	14.7	6.87	0.746
107	DRC/CC-ICP-MS	10	2.6	13	6.2	0.95
110	DRC/CC-ICP-MS	12.1	3.0	14.9	7.2	1.1
116	DRC/CC-ICP-MS	11.1	2.65	13.7	6.39	0.836
147	DRC/CC-ICP-MS	10.6	2.49	12.8	6.08	0.738
324	HR-ICP-MS	11.445	3.098	14.428	6.906	0.985
391	DRC/CC-ICP-MS	12.662	2.999	15.905	7.267	0.813
485	HR-ICP-MS	11.8	2.80	14.4	6.36	0.805
596	ICP-MS	9.79	2.60	18.3	5.54	2.11
597	DRC/CC-ICP-MS	10.3	2.45	12.2	6.09	0.73
632	DRC/CC-ICP-MS	10.8	2.51	13.4	6.04	0.299

### Summary Statistics

	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10
<b>Robust Mean (<math>x^*</math>)</b>	11.1	2.72	14.1	6.42	0.838
<b>Robust SD (<math>s^*</math>)</b>	1.0	0.24	1.4	0.54	0.143
<b>Robust RSD (%)</b>	9.29	9.02	10.3	8.46	17.1
<b>Number of Sample Measurements (N)</b>	11	11	11	11	11
<b>Standard Uncertainty (<math>u</math>)</b>	0.389	0.092	0.552	0.205	0.054



# Results for Event #2, 2016: Urine Cr



### Legend:

○ CHEAR Labs    ● Other Labs

Horizontal purple line = robust mean of all laboratories.

Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #2, 2016 Additional Elements in Urine: Cesium (Cs)






Urine Cs (µg/L)						
Lab Code	Method	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10
107	ICP-MS	3.5	10	6.7	5.1	2.7
110	ICP-MS	3.7	11.1	7.1	5.4	2.8
147	ICP-MS	3.79	11.2	7.2	5.62	2.78
399	ICP-MS	3.77	11.4	7.33	5.48	2.85
597	DRC/CC-ICP-MS	2.99	9.09	5.91	4.52	2.36
599	DRC/CC-ICP-MS	4.33	11.6	7.52	5.57	2.86
605	ICP-MS	3.45	10.4	6.74	5.05	2.60
606	ICP-MS	3.42	10.5	6.56	4.95	2.56
632	ICP-MS	3.82	11.7	7.69	5.58	2.99

Summary Statistics						
	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10	
Arithmetic Mean ( $\bar{x}$ )	3.64	10.7	6.97	5.25	2.72	
Arithmetic SD (s)	0.36	0.8	0.55	0.37	0.19	
Arithmetic RSD (%)	10.0	7.94	7.91	7.08	6.98	
Number of Sample Measurements (N)	9	9	9	9	9	

\*Denotes a statistical Outlier.



# Results for Event #2, 2016: Urine Cs



### Legend:

○ CHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #2, 2016 Additional Elements in Urine: Copper (Cu)






Urine Cu (µg/L)						
Lab Code	Method	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10
110	ICP-MS	2.3	40.9	23.1	17.2	8.6
116	DRC/CC-ICP-MS	ND	34.2	16.6	10.0	ND
147	ICP-MS	2.88	43.9	25.5	18.5	9.59
324	HR-ICP-MS	2.177	37.793	21.762	14.821	7.373
391	DRC/CC-ICP-MS	2.967	41.613	23.002	21.242	13.66
596	ICP-AES/OES	*15.7	43.7	31.7	19.7	9.67
597	DRC/CC-ICP-MS	1.60	34.4	19.9	14.7	7.52
632	ICP-MS	4.78	46.9	29.4	20.0	13.0

Summary Statistics					
	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>	2.78	40.4	23.8	17.0	9.91
<b>Arithmetic SD (s)</b>	1.09	4.6	4.9	3.6	2.50
<b>Arithmetic RSD (%)</b>	39.4	11.3	20.5	21.7	25.2
<b>Number of Sample Measurements (N)</b>	6	8	8	8	7

\*Denotes a statistical Outlier.



# Results for Event #2, 2016: Urine Cu



### Legend:

○ CHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.





## Results for Event #2, 2016

### Additional Elements in Urine: Molybdenum (Mo)






Urine Mo ( $\mu\text{g/L}$ )						
Lab Code	Method	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10
103	DRC/CC-ICP-MS	133	18.9	67.2	27.4	43.0
107	ICP-MS	130	19	68	27	43
110	ICP-MS	137	19.5	68.7	27.5	43.7
147	ICP-MS	126	19.3	63.8	27.3	40.7
324	HR-ICP-MS	127.654	18.496	64.672	26.763	41.791
399	ICP-MS	139	19.9	70.5	28.7	44.9
485	HR-ICP-MS	143	19.6	69.6	28.6	45.1
596	HR-ICP-MS	123	18.8	63.4	25.1	36.9
599	DRC/CC-ICP-MS	142.3	19.8	70.3	29.0	44.6
605	ICP-MS	127	17.7	64.6	26.0	41.1
606	ICP-MS	135	18.4	67.6	26.9	42.9
632	ICP-MS	147	23.5	75.3	30.6	47.6

#### Summary Statistics

	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10
<b>Robust Mean (<math>x^*</math>)</b>	134	19.1	67.5	27.4	43.0
<b>Robust SD (<math>s^*</math>)</b>	8	0.7	3.3	1.3	2.3
<b>Robust RSD (%)</b>	6.47	4.12	4.95	5.00	5.36
<b>Number of Sample Measurements (N)</b>	12	12	12	12	12
<b>Standard Uncertainty (<math>u</math>)</b>	3.13	0.285	1.20	0.496	0.833



# Results for Event #2, 2016: Urine Mo



### Legend:

- CHEAR Labs    ● Other Labs
- Horizontal purple line = robust mean of all laboratories.
- Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #2, 2016 Additional Elements in Urine: Nickel (Ni)






Urine Ni (µg/L)						
Lab Code	Method	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10
107	DRC/CC-ICP-MS	7.2	5.9	4.9	3.8	2.0
110	ICP-MS	7.8	6.3	5.4	4.1	2.6
147	DRC/CC-ICP-MS	7.52	6.35	5.13	3.73	2.25
324	ICP-MS	7.260	6.380	5.140	3.870	2.250
391	DRC/CC-ICP-MS	8.55	7.435	6.136	5.02	3.156
485	HR-ICP-MS	7.79	5.95	4.84	3.73	1.96
596	ICP-AES/OES	6.00	8.00	*8.00	<LOD	4.00
597	DRC/CC-ICP-MS	6.7	5.18	4.4	3.48	2.10
599	DRC/CC-ICP-MS	6.47	5.1	4.03	3.02	1.70

Summary Statistics						
	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10	
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>	7.25	6.28	4.99	3.84	2.44	
<b>Arithmetic SD (s)</b>	0.77	0.94	0.63	0.57	0.71	
<b>Arithmetic RSD (%)</b>	10.7	15.0	12.7	14.8	29.3	
<b>Number of Sample Measurements (N)</b>	9	9	8	8	9	

\*Denotes a statistical Outlier.



# Results for Event #2, 2016: Urine Ni



### Legend:

○ CHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #2, 2016 Additional Elements in Urine: Platinum (Pt)






Urine Pt (µg/L)						
Lab Code	Method	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10
107	ICP-MS	0.53	0.29	0.14	0.46	0.61
110	ICP-MS	0.47	0.21	0.09	0.38	0.54
147	ICP-MS	0.690	0.365	0.175	0.488	0.579
399	ICP-MS	0.515	0.276	0.132	0.462	0.582
596	ICP-MS	0.646	0.395	0.18	0.528	0.703
605	ICP-MS	0.503	0.268	0.143	0.439	0.548
632	ICP-MS	0.559	0.320	0.179	0.472	0.617

Summary Statistics					
	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10
Arithmetic Mean ( $\bar{x}$ )	0.559	0.303	0.148	0.461	0.596
Arithmetic SD (s)	0.080	0.062	0.032	0.045	0.054
Arithmetic RSD (%)	14.3	20.5	22.0	9.84	9.17
Number of Sample Measurements (N)	7	7	7	7	7

\*Denotes a statistical Outlier.



# Results for Event #2, 2016: Urine Pt



### Legend:

○ CHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #2, 2016 Additional Elements in Urine: Antimony (Sb)






Urine Sb ( $\mu\text{g/L}$ )						
Lab Code	Method	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10
103	DRC/CC-ICP-MS	0.420	0.230	0.103	0.555	0.842
107	ICP-MS	0.37	0.22	0.11	0.55	0.84
110	ICP-MS	0.39	0.23	0.12	0.59	0.86
147	ICP-MS	0.363	0.219	0.105	0.538	0.849
399	ICP-MS	0.411	0.242	0.120	0.604	0.878
597	DRC/CC-ICP-MS	0.49	0.35	0.39	0.71	1.01
599	DRC/CC-ICP-MS	0.849	*0.712	0.370	0.683	0.985
605	ICP-MS	0.368	PLC	PLC	0.497	0.824
606	ICP-MS	0.407	0.181	<0.080	0.541	0.731
632	ICP-MS	0.510	0.324	0.200	0.691	1.02

Summary Statistics					
	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10
<b>Robust Mean (<math>\bar{x}^*</math>)</b>	0.421	0.250	0.190	0.581	0.863
<b>Robust SD (<math>s^*</math>)</b>	0.061	0.057	0.121	0.059	0.047
<b>Robust RSD (%)</b>	14.6	23.0	64.0	10.2	5.46
<b>Number of Sample Measurements (N)</b>	10	8	8	10	10
<b>Standard Uncertainty (<math>u</math>)</b>	0.024	-	-	0.023	0.018

\*Statistical Outlier. Arithmetic statistics were compiled for sample UE16-07 and UE16-08 due to an insufficient number of sample measurements to perform robust statistics.



# Results for Event #2, 2016: Urine Sb



**Legend:**

○ CHEAR Labs    ● Other Labs  
 Horizontal purple line = robust mean of all laboratories.  
 Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.





## Results for Event #2, 2016 Additional Elements in Urine: Selenium (Se)






Urine Se (µg/L)						
Lab Code	Method	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10
103	DRC/CC-ICP-MS	48.6	181	98.4	62.7	132
110	DRC/CC-ICP-MS	9.8	139	56.1	17.0	88.9
147	ICP-MS	10.5	133	54.3	20	88.5
596	ICP-MS	26.5	145	74.6	37.8	104
596	HR-ICP-MS	6.61	130	52.0	15.2	76.8
597	DRC/CC-ICP-MS	10.3	121	53.5	20.8	86.9
632	DRC/CC-ICP-MS	47.8	168	93.5	51.3	123

Summary Statistics					
	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10
Arithmetic Mean ( $\bar{x}$ )	22.8	145	68.9	32.1	100
Arithmetic SD (s)	18.4	21	20.0	18.8	20
Arithmetic RSD (%)	81.0	14.8	29	58.6	20.5
Number of Sample Measurements (N)	7	7	7	7	7

\*Denotes a statistical Outlier.



# Results for Event #2, 2016: Urine Se



**Legend:**

○ CHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #2, 2016 Additional Elements in Urine: Tin (Sn)






Urine Sn (µg/L)						
Lab Code	Method	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10
107	ICP-MS	2.3	12	0.94	6.3	1.1
110	ICP-MS	2.34	13.3	0.67	6.74	0.96
147	ICP-MS	2.32	12.9	0.659	6.71	1
399	ICP-MS	2.39	13.8	0.663	6.84	0.996
596	ICP-MS	1.97	11.6	0.45	6.00	0.720
599	DRC/CC-ICP-MS	2.81	13.5	0.925	6.96	1.24
605	ICP-MS	2.37	13.1	PLC	6.75	1.74

Summary Statistics					
	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10
Arithmetic Mean ( $\bar{x}$ )	2.35	12.8	0.717	6.61	1.10
Arithmetic SD (s)	0.24	0.8	0.185	0.33	0.31
Arithmetic RSD (%)	10.4	6.23	25.8	5.12	28.8
Number of Sample Measurements (N)	7	7	6	7	7

\*Denotes a statistical Outlier.



# Results for Event #2, 2016: Urine Sn



### Legend:

○ CHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #2, 2016 Additional Elements in Urine: Strontium (Sr)

Urine Sr ( $\mu\text{g/L}$ )						
Lab Code	Method	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10
103	DRC/CC-ICP-MS	26.9	25.6	25.8	25.8	25.8
107	ICP-MS	29	28	28	28	28
200	ICP-MS	23.7	27.2	26.3	28.0	27.2
399	DRC/CC-ICP-MS	29.7	29.7	29.6	30	29.6
605	ICP-MS	28.8	28.2	28.4	28.3	28.5


Summary Statistics					
	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10
Arithmetic Mean ( $\bar{x}$ )	27.6	27.7	27.6	28.0	27.8
Arithmetic SD (s)	2.4	1.4	1.5	1.4	1.4
Arithmetic RSD (%)	8.77	5.40	5.64	5.33	5.12
Number of Sample Measurements (N)	5	5	5	5	5

\*Denotes a statistical Outlier.




# Results for Event #2, 2016: Urine Sr


UE16-06




UE16-07




UE16-08



UE16-09



UE16-10



### Legend:

○ CHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #2, 2016 Additional Elements in Urine: Vanadium (V)






Urine V (µg/L)						
Lab Code	Method	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10
116	DRC/CC-ICP-MS	0.132	1.46	ND	0.214	0.788
147	DRC/CC-ICP-MS	0.098	1.43	<0.0408	0.223	0.791
485	HR-ICP-MS	0.146	1.63	<0.012	0.245	0.923
596	HR-ICP-MS	<LOD	1.21	<LOD	<LOD	0.491
597	DRC/CC-ICP-MS	0.13	1.31	0.02	0.22	0.75
599	DRC/CC-ICP-MS	0.189	1.49	<0.02	0.242	0.818

Summary Statistics						
	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10	
Arithmetic Mean ( $\bar{x}$ )	0.139	1.42	0.02	0.228	0.760	
Arithmetic SD (s)	0.033	0.14	NA	0.013	0.144	
Arithmetic RSD (%)	23.7	10.2	NA	5.95	18.9	
Number of Sample Measurements (N)	5	6	1	5	6	

\*Denotes a statistical Outlier.



# Results for Event #2, 2016: Urine V



### Legend:

○ CHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.





## Results for Event #2, 2016 Additional Elements in Urine: Tungsten (W)






Urine W (µg/L)						
Lab Code	Method	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10
107	ICP-MS	0.34	0.16	1.1	0.59	0.33
110	ICP-MS	0.36	0.09	0.98	0.50	0.25
147	ICP-MS	0.333	0.132	1.05	0.556	0.331
200	ICP-MS	1.8	0.3	1.2	0.7	0.4
324	ICP-MS	<1	<1	1.000	<1	<1
399	ICP-MS	0.309	0.167	1.08	0.499	0.343
596	ICP-MS	0.341	0.153	1.1	0.558	1.02
599	DRC/CC-ICP-MS	5.72	*2.01	2.21	1.34	0.903
605	ICP-MS	0.317	PLC	1.05	0.572	0.323
606	ICP-MS	0.314	0.083	1.05	0.466	0.255
632	ICP-MS	0.700	0.379	1.29	0.733	0.459

Summary Statistics					
	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10
<b>Robust Mean (x*)</b>	0.350	0.183	1.08	0.586	0.368
<b>Robust SD (s*)</b>	0.039	0.104	0.05	0.103	0.103
<b>Robust RSD (%)</b>	11.3	57.0	5.31	17.7	28.1
<b>Number of Sample Measurements (N)</b>	10	8	11	10	10
<b>Standard Uncertainty (u)</b>	0.015	-	0.021	0.041	0.041

\*Statistical Outlier. Arithmetic statistics were compiled for sample UE16-07 due to an insufficient number of sample measurements to perform robust statistics.



# Results for Event #2, 2016: Urine W



### Legend:

- CHEAR Labs    ● Other Labs
- Horizontal purple line = robust mean of all laboratories.
- Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #2, 2016 Additional Elements in Urine: Zinc (Zn)






Urine Zn (µg/L)						
Lab Code	Method	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10
110	ICP-MS	468	95	209	742	134
147	ICP-MS	527	125	235	856	192
324	HR-ICP-MS	450.334	84.625	189.056	682.683	130.966
596	ICP-AES/OES	407	50.7	159	687	86.7
597	DRC/CC-ICP-MS	397	76.8	173	657	121
632	ICP-MS	511	106	232	814	150

Summary Statistics					
	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10
Arithmetic Mean ( $\bar{x}$ )	460	89.6	199	739	135
Arithmetic SD (s)	52	25.5	31	79	34
Arithmetic RSD (%)	11.5	28.4	15.6	10.8	25.5
Number of Sample Measurements (N)	6	6	6	6	6

\*Denotes a statistical Outlier.



# Results for Event #2, 2016: Urine Zn



### Legend:

○ CHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #2, 2016 Additional Elements in Urine: Aluminum (Al)

Urine Al (µg/L)						
Lab Code	Method	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10
147	DRC/CC-ICP-MS	<13.8	<13.8	<13.8	<13.8	<13.8
324	HR-ICP-MS	9.552	19.139	9.179	7.420	12.533
485	HR-ICP-MS	7.72	15.8	9.01	NR	11.1

Summary Statistics						
	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10	
Arithmetic Mean ( $\bar{x}$ )	8.63	17.4	9.09	7.42	11.8	
Arithmetic SD (s)	1.29	2.3	0.11	NA	1.0	
Arithmetic RSD (%)	15.0	13.5	1.25	NA	8.57	
Number of Sample Measurements (N)	2	2	2	1	2	

\*Denotes a statistical Outlier.



## Results for Event #2, 2016 Additional Elements in Urine: Tellurium (Te)

Urine Te ( $\mu\text{g/L}$ )						
Lab Code	Method	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10
110	ICP-MS	1.18	1.33	0.40	0.88	<MDL
596	ICP-MS	0.99	1.23	0.470	0.733	0.373
599	DRC/CC-ICP-MS	1.33	1.57	0.656	0.902	0.571

Summary Statistics						
	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10	
Arithmetic Mean ( $\bar{x}$ )	1.16	1.37	0.508	0.838	0.471	
Arithmetic SD (s)	0.17	0.17	0.132	0.091	0.140	
Arithmetic RSD (%)	14.6	12.6	26.0	10.9	29.6	
Number of Sample Measurements (N)	3	3	3	3	2	

\*Denotes a statistical Outlier.



## Results for Event #2, 2016 Additional Elements in Urine

### Urine B (µg/L)

Lab Code	Method	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10
200	ICP-MS	257	198	220	206	275

### Urine Bi (µg/L)

Lab Code	Method	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10
147	ICP-MS	<0.230	<0.230	<0.230	<0.230	<0.230

### Urine Fe (µg/L)

Lab Code	Method	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10
324	HR-ICP-MS	2.874	2.698	2.821	2.246	1.596

### Urine I (µg/L)

Lab Code	Method	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10
107	ICP-MS	59	57	57	56	57

### Urine Li (µg/L)

Lab Code	Method	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10
147	ICP-MS	8.61	8.54	8.61	9.02	8.47

### Urine Th (µg/L)

Lab Code	Method	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10
147	ICP-MS	<0.00557	<0.00557	<0.00557	<0.00557	<0.00557

### Urine Ti (µg/L)

Lab Code	Method	UE16-06	UE16-07	UE16-08	UE16-09	UE16-10
485	HR-ICP-MS	<0.49	<0.49	<0.49	<0.49	<0.49



**Department  
of Health**

**Wadsworth  
Center**

**Event #2, 2016**

**Trace Elements in  
Serum**

**Wadsworth Center**  
NEW YORK STATE DEPARTMENT OF HEALTH  
*Trace Elements Laboratory*





## Event #2, 2016: Trace Elements in Serum

### PT Materials

Test materials were prepared from human serum obtained from Tennessee Blood Services, Inc. The company certifies that these materials were tested by FDA approved methods and found to be negative for HIV 1<sup>2</sup> and HIV-1 RNA, and non-reactive to HBsAg, HCV3 and STS. Units of serum were filtered into polypropylene containers through cheesecloth to remove particulates and supplemented with aluminum (Al), copper (Cu), selenium (Se), zinc (Zn), arsenic (As), beryllium (Be), cadmium (Cd), cobalt (Co), chromium (Cr), mercury (Hg), manganese (Mn), molybdenum (Mo), nickel (Ni), lead (Pb), platinum (Pt), antimony (Sb), titanium (Ti), thallium (Tl), uranium (U), vanadium (V), and tungsten (W). Serum units were homogenized overnight prior to aliquoting 2-mL into polypropylene vials. PT samples were stored at -80°C until the week of the PT event, when they were thawed at 4°C prior to circulation to laboratories for analysis.

### Graded Elements

Three elements in serum are formally graded: Cu, Se, and Zn. Target values for the graded elements are assigned to these pools based on the robust mean calculated from data reported by all laboratories.

### Additional Elements

An additional 30 elements (beyond the three graded) were reported by at least one participant: Ag, Al, As, B, Ba, Be, Bi, Cd, Co, Cr, Cs, Fe, Hg, I, Li, Mn, Mo, Ni, Pb, Pt, Sb, Sn, Sr, Te, Th, Ti, Tl, U, V, and W. These data are included here to provide a more complete characterization of the PT materials. All results reported by participant laboratories are tabulated and organized by lab code. The PT data are graphed for visual comparison purposes for all elements where at least five laboratories reported a value greater than the LOD. A statistical summary table is provided for samples where at least two comparable values were reported as above the LOD.

The summary statistics for the additional elements are provided for educational purposes only, i.e., no acceptable response is implied. However, it is expected that each laboratory would wish to investigate a potential source of bias if warranted by these data. Future events might result in additional elements becoming graded if a consensus can be reached regarding desired quality specifications.



# Results for Event #2, 2016 Serum Copper (Cu) Summary Statistics

	Serum Cu (µg/L)				
	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10
<b>Target (Robust Mean (x*))</b>	1948	1320	741	1072	1365
<b>Upper Limit</b>	2240	1518	852	1233	1570
<b>Lower Limit</b>	1656	1122	630	911	1160
<b>Robust SD (s*)</b>	80	54	47	56	85
<b>Robust RSD (%)</b>	4.12	4.12	6.34	5.24	6.27
<b>Number of Sample Measurements (N)</b>	10	10	10	10	10
<b>Standard Uncertainty (u)</b>	31.7	21.5	18.5	22.2	33.8

The acceptable range is based on quality specifications: ±95 µg/L or ±15% around the target value, whichever is greater; thus, it is fixed at ±95 µg/L at concentrations less than or equal to 635 µg/L. These quality specifications were established by New York State Department of Health's Wadsworth Center, the PT Program organizer.



# Results for Event #2, 2016 Serum Copper (Cu)






## Performance of Participating Laboratories

Serum Cu (µg/L)						
Lab Code	Method	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10
	Target	1948	1320	741	1072	1365
107	DRC/CC-ICP-MS	1900	1300	740	1100	1300
110	ICP-MS	1924	1355	752	1072	1373
147	ICP-MS	1938	1366	781	1112	1366
200	ICP-MS	2045	1435	806	1143	1492
293	ICP-MS	1811.82	1246.03	692.94	1010.81	1290.53
457	ICP-AES/OES	1980	1306	782	1118	1428
483	DRC/CC-ICP-MS	1990	1310	735	1050	1390
596	ICP-AES/OES	2130	1350	738	1100	1350
597	DRC/CC-ICP-MS	1744	1163	657	980	1171
598	ICP-MS	1960	1330	717	998	1430

Based on the grading criteria for Cu in Serum, 100% of results were satisfactory, with 0 of the 10 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #2, 2016: Serum Cu



### Legend:

○ CHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

±95 µg/L or ±15% around the target value, whichever is greater; thus, it is fixed at ±95 µg/L at concentrations less than or equal to 635 µg/L.



# Results for Event #2, 2016 Serum Selenium (Se)

## Summary Statistics

	Serum Se ( $\mu\text{g/L}$ )				
	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10
<b>Target (Robust Mean (<math>x^*</math>))</b>	146	160	103	212	122
<b>Upper Limit</b>	176	192	124	255	147
<b>Lower Limit</b>	117	128	82	170	98
<b>Robust SD (<math>s^*</math>)</b>	14	12	5	12	7
<b>Robust RSD (%)</b>	10.0	7.90	5.02	5.91	5.84
<b>Number of Sample Measurements (N)</b>	10	10	10	10	10
<b>Standard Uncertainty (<math>u</math>)</b>	5.83	5.02	2.05	4.97	2.82

The acceptable range is based on quality specifications:  $\pm 2 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 2 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ . These quality specifications were established by New York State Department of Health's Wadsworth Center, the PT Program organizer.








Results for Event #2, 2016
Serum Selenium (Se)
Performance of Participating Laboratories

Table with 7 columns: Lab Code, Method, SE16-06, SE16-07, SE16-08, SE16-09, SE16-10. Includes a Target row and data for 10 laboratories. Red arrows indicate values outside acceptable ranges for Lab 598.

Based on the grading criteria for Se in Serum, 92% of results were satisfactory, with 1 of the 10 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #2, 2016: Serum Se



### Legend:

○ CHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

$\pm 2 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 2 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ .



# Results for Event #2, 2016 Serum Zinc (Zn) Summary Statistics

	Serum Zn (µg/L)				
	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10
<b>Target (Robust Mean (x*))</b>	965	1364	1077	689	546
<b>Upper Limit</b>	1110	1569	1238	793	628
<b>Lower Limit</b>	820	1159	915	586	464
<b>Robust SD (s*)</b>	59	91	49	25	33
<b>Robust RSD (%)</b>	6.20	6.74	4.63	3.72	6.05
<b>Number of Sample Measurements (N)</b>	10	10	10	10	10
<b>Standard Uncertainty (u)</b>	23.6	36.3	19.7	10.1	13.0

The acceptable range is based on quality specifications: ±15 µg/L or ±15% around the target value, whichever is greater; thus, it is fixed at ±15 µg/L at concentrations less than or equal to 100 µg/L. These quality specifications were established by New York State Department of Health's Wadsworth Center, the PT Program organizer.





# Results for Event #2, 2016 Serum Zinc (Zn)






## Performance of Participating Laboratories

Serum Zn (µg/L)						
Lab Code	Method	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10
	<b>Target</b>	<b>965</b>	<b>1364</b>	<b>1077</b>	<b>689</b>	<b>546</b>
107	DRC/CC-ICP-MS	960	1300	1000	690	520
110	ICP-MS	972	1390	1116	696	541
147	ICP-MS	933	1366	1090	703	532
200	ICP-MS	1020	1491	1177	746	602
293	ICP-MS	915.03	1320.26	1026.140	679.74	533.33
457	ICP-AES/OES	944	1294	1079	691	537
483	DRC/CC-ICP-MS	1020	1390	1100	674	566
596	ICP-AES/OES	1140	↑ 1480	1110	739	564
597	DRC/CC-ICP-MS	871	1218	977	636	469
598	ICP-MS	966	1380	1060	660	586

Based on the grading criteria for Zn in Serum, 98% of results were satisfactory, with 0 of the 10 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #2, 2016: Serum Zn



### Legend:

○ CHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

±15 µg/L or ±15% around the target value, whichever is greater; thus, it is fixed at ±15 µg/L at concentrations less than or equal to 100 µg/L.



## Results for Event #2, 2016 Additional Elements in Serum: Aluminum (Al)

Serum Al (µg/L)						
Lab Code	Method	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10
147	ETAAS-Z	40.60	82.8	53.1	17.6	36
200	DRC/CC-ICP-MS	50.5	73.7	52.1	25.4	36.2
293	ICP-MS	43.17	83.92	60.71	15.92	35.35
485	HR-ICP-MS	39.9	84.5	58.4	15.6	34.4
596	ICP-AES/OES	62.50	150	100	37.5	100
598	ICP-MS	93.30	116	95.3	51.3	116

Summary Statistics						
	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10	
Arithmetic Mean ( $\bar{x}$ )	44.7	80.1	55.3	19.6	35.8	
Arithmetic SD (s)	5.1	5.6	4.7	5.0	0.4	
Arithmetic RSD (%)	11.4	6.99	8.51	25.7	1.23	
Number of Sample Measurements (N)	3	3	3	3	3	


\*Denotes a statistical Outlier.

Serum Al was not graded for Event #2, 2016 due to lack of a consensus value. The mean values reported here are based on three laboratories (147, 200, and 293) that have a long history of successful proficiency testing in this scheme.




# Results for Event #2, 2016: Serum AI


SE16-06




SE16-07




SE16-08



SE16-09



SE16-10



### Legend:

○ CHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #2, 2016 Additional Elements in Serum: Arsenic (As)






Serum As (µg/L)						
Lab Code	Method	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10
103	DRC/CC-ICP-MS	2.97	5.27	2.96	9.41	15.7
110	DRC/CC-ICP-MS	3.0	5.0	2.9	9.0	14.8
147	ICP-MS	3.28	5.77	3.76	9.36	14.5
596	ICP-MS	*5.03	*9.53	*8.23	*14.3	*19.2
597	DRC/CC-ICP-MS	2.82	5.21	3.14	9.34	14.8

Summary Statistics					
	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10
Arithmetic Mean ( $\bar{x}$ )	3.01	5.31	3.19	9.27	14.9
Arithmetic SD (s)	0.19	0.32	0.39	0.18	0.5
Arithmetic RSD (%)	6.35	6.14	12.3	2.01	3.47
Number of Sample Measurements (N)	4	4	4	4	4

\*Denotes a statistical Outlier.



# Results for Event #2, 2016: Serum As



### Legend:

○ CHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #2, 2016 Additional Elements in Serum: Cadmium (Cd)






Serum Cd (µg/L)						
Lab Code	Method	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10
103	DRC/CC-ICP-MS	0.353	1.47	3.38	0.668	0.811
110	ICP-MS	0.3	1.5	3.5	0.7	0.9
147	ICP-MS	0.263	1.63	3.53	0.722	0.864
596	HR-ICP-MS	0.117	*0.937	*2.27	*0.277	*0.450
597	DRC/CC-ICP-MS	0.31	1.51	3.49	0.68	0.86

Summary Statistics					
	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10
Arithmetic Mean ( $\bar{x}$ )	0.268	1.52	3.47	0.692	0.858
Arithmetic SD (s)	0.090	0.07	0.06	0.023	0.036
Arithmetic RSD (%)	33.7	4.60	1.88	3.42	4.25
Number of Sample Measurements (N)	5	4	4	4	4

\*Denotes a statistical Outlier.



# Results for Event #2, 2016: Serum Cd



### Legend:

○ CHEAR Labs

● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.





# Results for Event #2, 2016 Additional Elements in Serum: Cobalt (Co)






Serum Co (µg/L)						
Lab Code	Method	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10
103	DRC/CC-ICP-MS	1.35	4.49	0.936	3.70	0.0922
110	ICP-MS	1.5	4.5	0.9	3.6	0.2
147	ICP-MS	1.31	4.08	0.897	3.39	0.201
485	HR-ICP-MS	1.43	4.44	0.909	3.86	0.106
596	ICP-MS	1.72	4.67	1.09	3.74	0.370
597	DRC/CC-ICP-MS	1.18	*2.98	0.63	*2.58	0.08

Summary Statistics						
	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10	
Arithmetic Mean ( $\bar{x}$ )	1.42	4.43	0.894	3.65	0.175	
Arithmetic SD (s)	0.18	0.21	0.148	0.17	0.110	
Arithmetic RSD (%)	12.7	4.89	16.6	4.82	62.9	
Number of Sample Measurements (N)	6	5	6	5	6	

\*Denotes a statistical Outlier.



# Results for Event #2, 2016: Serum Co



### Legend:

○ CHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #2, 2016 Additional Elements in Serum: Chromium (Cr)






Serum Cr (µg/L)						
Lab Code	Method	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10
103	DRC/CC-ICP-MS	0.546	00.312	3.13	0.349	2.37
110	DRC/CC-ICP-MS	0.6	0.4	3.2	*0.7	2.3
147	DRC/CC-ICP-MS	0.671	0.372	2.92	0.405	2.17
485	HR-ICP-MS	0.576	0.341	3.08	0.402	2.34
596	ICP-MS	<LOD	<LOD	*17.0	<LOD	<LOD
597	DRC/CC-ICP-MS	0.63	0.33	3.09	0.44	2.20

Summary Statistics						
	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10	
Arithmetic Mean ( $\bar{x}$ )	0.604	0.351	3.08	0.399	2.27	
Arithmetic SD (s)	0.048	0.035	0.10	0.037	0.08	
Arithmetic RSD (%)	7.96	9.97	3.34	9.40	3.83	
Number of Sample Measurements (N)	5	5	5	4	5	

\*Denotes a statistical Outlier.



# Results for Event #2, 2016: Serum Cr



### Legend:

○ CHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #2, 2016 Additional Elements in Serum: Mercury (Hg)






Serum Hg ( $\mu\text{g/L}$ )						
Lab Code	Method	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10
103	DRC/CC-ICP-MS	1.77	3.04	0.781	2.31	2.18
110	ICP-MS	1.8	3.1	0.7	2.3	2.0
147	ICP-MS	1.58	2.69	0.702	2.13	1.98
596	ICP-MS	1.85	2.62	0.765	2.1	1.75
597	DRC/CC-ICP-MS	1.73	3.05	0.92	2.51	2.18

Summary Statistics						
	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10	
Arithmetic Mean ( $\bar{x}$ )	1.74	2.90	0.773	2.27	2.01	
Arithmetic SD (s)	0.10	0.22	0.089	0.16	0.17	
Arithmetic RSD (%)	5.87	7.79	11.5	7.25	8.79	
Number of Sample Measurements (N)	5	5	5	5	5	

\*Denotes a statistical Outlier.



# Results for Event #2, 2016: Serum Hg



**Legend:**

○ CHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #2, 2016 Additional Elements in Serum: Manganese (Mn)






Serum Mn (µg/L)						
Lab Code	Method	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10
103	DRC/CC-ICP-MS	6.91	2.31	5.69	3.33	10.4
110	ICP-MS	6.8	2.6	5.7	3.5	10.0
147	ICP-MS	7.09	2.80	5.77	3.43	9.95
596	ICP-MS	6.97	2.12	4.94	2.85	9.44
597	DRC/CC-ICP-MS	6.52	2.27	5.12	3.08	8.71

Summary Statistics						
	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10	
Arithmetic Mean ( $\bar{x}$ )	6.85	2.42	5.44	3.23	9.69	
Arithmetic SD (s)	0.21	0.27	0.38	0.26	0.65	
Arithmetic RSD (%)	3.15	11.3	7.06	8.30	6.70	
Number of Sample Measurements (N)	5	5	5	5	5	

\*Denotes a statistical Outlier.



# Results for Event #2, 2016: Serum Mn



### Legend:

○ CHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.





## Results for Event #2, 2016 Additional Elements in Serum: Molybdenum (Mo)






Serum Mo (µg/L)						
Lab Code	Method	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10
103	DRC/CC-ICP-MS	0.998	1.17	2.15	4.83	2.35
110	ICP-MS	*1.3	1.5	2.5	5.2	2.4
147	ICP-MS	0.950	1.23	2.16	4.78	2.12
485	HR-ICP-MS	0.980	1.09	2.34	5.09	2.35
596	HR-ICP-MS	0.92	1.41	1.81	4.07	2.00

Summary Statistics						
	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10	
Arithmetic Mean ( $\bar{x}$ )	0.961	1.28	2.19	4.79	2.24	
Arithmetic SD (s)	0.034	0.17	0.25	0.44	0.17	
Arithmetic RSD (%)	3.56	13.3	11.7	9.20	7.77	
Number of Sample Measurements (N)	4	5	5	5	5	

\*Denotes a statistical Outlier.



# Results for Event #2, 2016: Serum Mo



### Legend:

○ CHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #2, 2016 Additional Elements in Serum: Nickel (Ni)






Serum Ni (µg/L)						
Lab Code	Method	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10
110	DRC/CC-ICP-MS	2.9	1.7	7.4	5.8	00.7
147	ICP-MS	2.66	1.21	7.40	5.93	1.01
485	HR-ICP-MS	2.55	1.06	7.56	6.29	0.732
596	ICP-MS	3.03	1.68	7.59	5.96	1.46
597	DRC/CC-ICP-MS	3.24	1.59	6.93	5.95	0.66

Summary Statistics						
	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10	
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>	2.87	1.44	7.37	5.98	0.912	
<b>Arithmetic SD (s)</b>	0.27	0.29	0.26	0.18	0.335	
<b>Arithmetic RSD (%)</b>	9.67	20.2	3.58	3.03	36.7	
<b>Number of Sample Measurements (N)</b>	5	5	5	5	5	

\*Denotes a statistical Outlier.



# Results for Event #2, 2016: Serum Ni



### Legend:

○ CHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #2, 2016 Additional Elements in Serum: Lead (Pb)






Serum Pb (µg/L)						
Lab Code	Method	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10
103	DRC/CC-ICP-MS	2.71	10.0	1.10	8.10	5.35
110	ICP-MS	2.8	10	1.1	8.2	5.3
147	ICP-MS	3.21	10.5	*12.1	8.81	5.45
596	ICP-MS	2.78	9.98	1.02	7.84	5.19
597	DRC/CC-ICP-MS	2.42	8.69	0.96	7.29	*4.47

Summary Statistics						
	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10	
Arithmetic Mean ( $\bar{x}$ )	2.78	9.83	1.04	8.04	5.32	
Arithmetic SD (s)	0.28	0.67	0.06	0.55	0.10	
Arithmetic RSD (%)	10.1	6.87	6.51	6.87	2.03	
Number of Sample Measurements (N)	5	5	4	5	4	

\*Denotes a statistical Outlier.



# Results for Event #2, 2016: Serum Pb



### Legend:

○ CHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #2, 2016 Additional Elements in Serum: Barium (Ba)

Serum Ba (µg/L)						
Lab Code	Method	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10
110	ICP-MS	0.4	0.4	0.9	1.1	0.4
147	ICP-MS	0.385	0.412	0.971	1.15	0.451
596	HR-ICP-MS	0.368	0.334	0.764	0.848	0.365
597	DRC/CC-ICP-MS	0.38	0.43	0.88	0.97	0.39

Summary Statistics						
	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10	
Arithmetic Mean ( $\bar{x}$ )	0.383	0.394	0.878	1.01	0.401	
Arithmetic SD (s)	0.012	0.041	0.085	0.13	0.036	
Arithmetic RSD (%)	3.35	10.6	9.77	13.3	8.99	
Number of Sample Measurements (N)	4	4	4	4	4	

\*Denotes a statistical Outlier.



## Results for Event #2, 2016 Additional Elements in Serum: Beryllium (Be)

Serum Be (µg/L)						
Lab Code	Method	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10
110	ICP-MS	7.2	1.2	0.4	4.3	2.1
147	ICP-MS	6.54	1.13	0.572	4.31	1.75

Summary Statistics						
	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10	
Arithmetic Mean ( $\bar{x}$ )	6.87	1.16	0.485	4.3	1.92	
Arithmetic SD (s)	0.46	0.04	0.121	0.0	0.24	
Arithmetic RSD (%)	6.79	4.24	25	0.164	12.8	
Number of Sample Measurements (N)	2	2	2	2	2	

\*Denotes a statistical Outlier.





## Results for Event #2, 2016 Additional Elements in Serum: Cesium (Cs)

Serum Cs ( $\mu\text{g/L}$ )						
Lab Code	Method	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10
110	ICP-MS	0.9	0.9	0.3	0.3	0.7
597	DRC/CC-ICP-MS	0.79	0.77	0.25	0.25	0.52

Summary Statistics						
	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10	
Arithmetic Mean ( $\bar{x}$ )	0.844	0.834	0.275	0.275	0.610	
Arithmetic SD (s)	0.077	0.091	0.035	0.035	0.127	
Arithmetic RSD (%)	9.20	11.0	12.8	12.8	20.8	
Number of Sample Measurements (N)	2	2	2	2	2	

\*Denotes a statistical Outlier.



## Results for Event #2, 2016 Additional Elements in Serum: Iron (Fe)

Serum Fe ( $\mu\text{g/L}$ )						
Lab Code	Method	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10
457	ICP-AES/OES	1026	974	620	579	290
483	DRC/CC-ICP-MS	860	860	540	540	300

Summary Statistics						
	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10	
Arithmetic Mean ( $\bar{x}$ )	943	917	580	559	295	
Arithmetic SD (s)	117	80	56	27	7	
Arithmetic RSD (%)	12.4	8.79	9.75	4.92	2.39	
Number of Sample Measurements (N)	2	2	2	2	2	

\*Denotes a statistical Outlier.



## Results for Event #2, 2016 Additional Elements in Serum: Platinum (Pt)

Serum Pt (µg/L)						
Lab Code	Method	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10
110	ICP-MS	1.32	0.33	0.04	0.52	1.79
596	HR-ICP-MS	1.52	0.321	0.0789	0.587	1.92
Summary Statistics						
		SE16-06	SE16-07	SE16-08	SE16-09	SE16-10
Arithmetic Mean ( $\bar{x}$ )		1.42	0.325	0.059	0.553	1.85
Arithmetic SD (s)		0.14	0.006	0.027	0.047	0.09
Arithmetic RSD (%)		9.95	1.95	46.2	8.55	4.95
Number of Sample Measurements (N)		2	2	2	2	2

\*Denotes a statistical Outlier.



## Results for Event #2, 2016 Additional Elements in Serum: Antimony (Sb)

Serum Sb ( $\mu\text{g/L}$ )						
Lab Code	Method	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10
103	DRC/CC-ICP-MS	8.34	4.53	3.51	11.7	1.40
110	ICP-MS	7.6	4.2	3.2	10.7	1.2
147	ICP-MS	6.81	3.87	2.97	10.3	1.16
597	DRC/CC-ICP-MS	6.58	3.68	2.85	8.92	1.26

Summary Statistics						
	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10	
Arithmetic Mean ( $\bar{x}$ )	7.33	4.07	3.13	10.4	1.25	
Arithmetic SD (s)	0.80	0.37	0.29	1.10	0.10	
Arithmetic RSD (%)	10.9	9.19	9.27	11	8.36	
Number of Sample Measurements (N)	4	4	4	4	4	

\*Denotes a statistical Outlier.



## Results for Event #2, 2016 Additional Elements in Serum: Tin (Sn)

Serum Sn ( $\mu\text{g/L}$ )						
Lab Code	Method	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10
110	ICP-MS	5.6	1.7	1.0	4.1	2.8
147	ICP-MS	5.07	1.57	0.852	4.01	2.60
596	ICP-MS	5.42	1.59	0.812	3.59	2.67

Summary Statistics						
	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10	
Arithmetic Mean ( $\bar{x}$ )	5.36	1.62	0.888	3.90	2.69	
Arithmetic SD (s)	0.26	0.07	0.099	0.27	0.10	
Arithmetic RSD (%)	5.02	4.32	11.1	6.97	3.77	
Number of Sample Measurements (N)	3	3	3	3	3	

\*Denotes a statistical Outlier.



## Results for Event #2, 2016 Additional Elements in Serum: Strontium (Sr)

Serum Sr (µg/L)						
Lab Code	Method	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10
103	DRC/CC-ICP-MS	28.0	28.3	17.8	18.3	35.2
200	ICP-MS	26.3	27.2	16.6	16.6	33.3

Summary Statistics						
	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10	
Arithmetic Mean ( $\bar{x}$ )	27.1	27.7	17.2	17.4	34.2	
Arithmetic SD (s)	1.2	0.7	0.8	1.2	1.3	
Arithmetic RSD (%)	4.42	2.80	4.93	6.88	3.92	
Number of Sample Measurements (N)	2	2	2	2	2	

\*Denotes a statistical Outlier.



## Results for Event #2, 2016 Additional Elements in Serum: Thallium (TI)

Serum TI ( $\mu\text{g/L}$ )						
Lab Code	Method	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10
103	DRC/CC-ICP-MS	1.99	5.42	3.98	1.06	3.41
110	ICP-MS	2.2	5.7	4.3	1.2	3.7
147	ICP-MS	1.93	5.48	4.21	1.16	3.45
596	HR-ICP-MS	2.12	5.24	3.44	1.02	3.18

Summary Statistics						
	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10	
Arithmetic Mean ( $\bar{x}$ )	2.06	5.46	3.98	1.11	3.43	
Arithmetic SD (s)	0.12	0.18	0.38	0.08	0.21	
Arithmetic RSD (%)	5.94	3.47	9.69	7.57	6.20	
Number of Sample Measurements (N)	4	4	4	4	4	

\*Denotes a statistical Outlier.



## Results for Event #2, 2016 Additional Elements in Serum: Uranium (U)

Serum U (µg/L)						
Lab Code	Method	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10
103	DRC/CC-ICP-MS	0.0231	0.0225	0.00958	0.0412	0.0139
110	ICP-MS	0.024	0.026	0.014	0.05	0.017
147	ICP-MS	0.025	0.0255	0.0145	0.0424	<0.0145
596	HR-ICP-MS	0.032	0.0314	0.0159	0.0458	0.0246

Summary Statistics						
	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10	
Arithmetic Mean ( $\bar{x}$ )	0.026	0.026	0.013	0.044	0.018	
Arithmetic SD (s)	0.004	0.003	0.002	0.003	0.005	
Arithmetic RSD (%)	15.9	14.0	20.2	8.80	29.7	
Number of Sample Measurements (N)	4	4	4	4	3	

\*Denotes a statistical Outlier.





## Results for Event #2, 2016 Additional Elements in Serum: Vanadium (V)

Serum V (µg/L)						
Lab Code	Method	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10
110	DRC/CC-ICP-MS	3.3	*0.8	6.5	1.7	4.4
147	DRC/CC-ICP-MS	2.71	0.462	5.92	1.24	4.29
485	HR-ICP-MS	3.01	0.420	6.22	1.27	4.63
597	DRC/CC-ICP-MS	2.52	0.41	5.56	1.09	3.93

Summary Statistics						
	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10	
Arithmetic Mean ( $\bar{x}$ )	2.88	0.430	6.05	1.32	4.31	
Arithmetic SD (s)	0.34	0.027	0.40	0.26	0.29	
Arithmetic RSD (%)	11.8	6.40	6.66	19.7	6.76	
Number of Sample Measurements (N)	4	3	4	4	4	

\*Denotes a statistical Outlier.



## Results for Event #2, 2016 Additional Elements in Serum: Tungsten (W)

Serum W (µg/L)						
Lab Code	Method	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10
110	ICP-MS	3.4	2.3	0.2	1.3	0.9
147	ICP-MS	3.79	2.37	0.197	1.31	0.896
200	ICP-MS	3.7	2.6	0.4	1.4	*1.2
596	HR-ICP-MS	4.32	2.42	0.132	1.33	0.910

Summary Statistics						
	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10	
Arithmetic Mean ( $\bar{x}$ )	3.80	2.42	0.232	1.33	0.902	
Arithmetic SD (s)	0.38	0.12	0.116	0.04	0.007	
Arithmetic RSD (%)	10.0	5.29	50.0	3.37	0.799	
Number of Sample Measurements (N)	4	4	4	4	3	

\*Denotes a statistical Outlier.



## Results for Event #2, 2016 Additional Elements in Serum

### Serum Ag (µg/L)

Lab Code	Method	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10
147	ICP-MS	<0.248	<0.248	<0.248	<0.248	<0.248

### Serum B (µg/L)

Lab Code	Method	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10
200	ICP-MS	117	67	36	42	52

### Serum Bi (µg/L)

Lab Code	Method	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10
147	ICP-MS	<0.200	<0.200	<0.200	<0.200	<0.200

### Serum I (µg/L)

Lab Code	Method	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10
147	ICP-MS	47.0	47.2	35.1	35.1	184

### Serum Li (µg/L)

Lab Code	Method	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10
147	ICP-MS	0.458	0.459	0.396	0.457	2.53

### Serum Te (µg/L)

Lab Code	Method	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10
147	ICP-MS	<0.0880	<0.0880	<0.0880	<0.0880	<0.0880
596	HR-ICP-MS	<LOD	<LOD	<LOD	<LOD	<LOD

### Serum Th (µg/L)

Lab Code	Method	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10
147	ICP-MS	<0.00789	<0.00789	<0.00789	<0.00789	<0.00789

### Serum Ti (µg/L)

Lab Code	Method	SE16-06	SE16-07	SE16-08	SE16-09	SE16-10
485	HR-ICP-MS	97.20	48.1	18.9	15.0	9.13



Department  
of Health

Wadsworth  
Center

## References

1. ISO/FDIS-13528 (2005) Statistical methods for use in proficiency testing by interlaboratory comparisons. International Organization for Standardization, Geneva.
2. Taylor A, Angerer J, Arnaud J, Claeys F, Jones RL, Mazarrasa O, Mairiaux E, Menditto A, Parsons PJ, Patriarca M, Pineau A, Valkonen S, Weber J-P, Weykamp C. Occupational and environmental laboratory medicine: A network of EQAS organisers. Accreditation and Quality Assurance. 2006;11(8-9):435-9. PubMed PMID: 086NJ-0011.