



**Department  
of Health**

**Wadsworth  
Center**

# **New York State Biomonitoring Program for Trace Elements**

## **Event #2, 2023**

### **Trace Elements in Whole Blood, Urine, and Serum**

## **July, 2023**

**Wadsworth Center**  
NEW YORK STATE DEPARTMENT OF HEALTH  
*Trace Elements Laboratory*



**Event #2, 2023:  
Trace Elements in Whole Blood, Urine, and Serum**

7/26/2023

Dear Laboratory Director,

This report summarizes performance for the second biomonitoring proficiency test (PT) event of 2023 for Trace Elements in Whole Blood, Urine, and Serum. One of the key goals of this PT program is to achieve harmonization of biomonitoring data for trace elements. In this report, we summarize the responses to our recent survey request. Please refer to the attachment at the end of the report for more details.

**Target Value Assignment and Performance Evaluation:**

For these PT materials, target values have been assigned for a limited number of trace elements that are gradable under criteria set by the NYS DOH Biomonitoring PT program. See assay-specific narratives for details. Data for additional trace elements are reported and are included here in order to characterize the PT materials more completely. Participant data and descriptive statistics are provided for educational purposes. No target value or acceptable range is implied.

Where the data permit, robust statistics were used to assign target values based on Algorithm A as defined by ISO 13528:2005E *Statistical methods for use in proficiency testing by inter-laboratory comparisons* [1]. Acceptable ranges for the graded elements are based on consensus criteria and/or those set by the NYS DOH's PT program. For example, some are fixed based on US regulatory guidelines (Pb, Cd) while for other elements the criteria are based on a consensus of the Network of PT scheme organizers for trace elements in occupational and environmental laboratory medicine [2]. Quality specifications are element and matrix specific; full details are provided under each element specific narrative.

A confidential, three-digit code number assigned by PT program staff identifies all laboratory participants.

Samples for the next PT event (Event #3, 2023) will be shipped September 6, 2023. Comments about this report may be directed to [trel@health.ny.gov](mailto:trel@health.ny.gov).

Sincerely,

Patrick J. Parsons, PhD  
Chief, Inorganic and Nuclear Chemistry,  
Division of Environmental Sciences  
Wadsworth Center

Kayla Mehigan  
Coordinator, Biomonitoring PT Program,  
Division of Environmental Sciences  
Wadsworth Center



**Department  
of Health**

**Wadsworth  
Center**

**Event #2, 2023**

**Trace Elements in  
Whole Blood**

**Wadsworth Center**  
NEW YORK STATE DEPARTMENT OF HEALTH  
*Trace Elements Laboratory*



**Event #2, 2023:  
Trace Elements in Whole Blood**

**PT Materials**

Human whole blood was purchased from Zen-Bio, Inc. and preserved with K<sub>2</sub>EDTA. The company certifies that this material was "non-reactive" for HBsAg, HBV DNA, HIV-1,2 Ab, HIV-1 RNA, HCV Ab, HCV RNA, and STS. Units of whole blood were filtered into polypropylene containers through cheesecloth to remove particulates and supplemented with arsenic (As), cadmium (Cd), cobalt (Co), chromium (Cr), mercury (Hg), manganese (Mn), lead (Pb), barium (Ba), beryllium (Be), copper (Cu), molybdenum (Mo), nickel (Ni), platinum (Pt), antimony (Sb), selenium (Se), tin (Sn), titanium (Ti), thallium (Tl), uranium (U), vanadium (V), tungsten (W), and zinc (Zn). Whole blood samples were homogenized overnight prior to aliquoting 2-mL into polypropylene vials. PT samples were stored at -80°C until the week of the PT event, when they were thawed at 4°C prior to circulation to laboratories

**Graded Elements**

Seven elements in whole blood are formally graded: As, Cd, Co, Cr, Hg, Mn, and Pb. Target values for the graded elements are assigned to these pools based on (a) the robust mean calculated from data reported by all laboratories, or (b) if a robust mean is not possible, the arithmetic mean after outlier deletion.

**Additional Elements**

An additional 22 elements were reported by at least one participant: Al, Ba, Be, Bi, Cs, Cu, Mg, Mo, Ni, Pt, Sb, Se, Sn, Sr, Te, Th, Ti, Tl, U, V, W, and Zn. These data are included here to provide a more complete characterization of the PT materials. All results reported by participant laboratories are tabulated and organized by lab code. The PT data are graphed for visual comparison purposes for all elements where at least five laboratories reported a value greater than the LOD. A statistical summary table is provided for samples where at least two comparable values were reported as above the LOD.

The summary statistics for the additional elements are provided for educational purposes only, i.e., no acceptable response is implied. However, it is expected that each laboratory would wish to investigate a potential source of bias if warranted by these data. Future events might result in additional elements becoming graded if a consensus can be reached regarding desired quality specifications.



## Results for Event #2, 2023: Summary Statistics

Whole Blood As (µg/L)					
	BE23-06	BE23-07	BE23-08	BE23-09	BE23-10
<b>Target (Arithmetic Mean (<math>\bar{x}</math>))</b>	15.5	10.5	2.37	45.8	5.5
<b>Upper Limit</b>	21.5	16.5	8.37	55.0	11.5
<b>Lower Limit</b>	9.5	4.5	0.00	36.6	0.0
<b>Arithmetic SD (s)</b>	0.9	0.5	0.14	2.3	0.3
<b>Arithmetic RSD (%)</b>	5.8	4.8	5.9	5.0	5.9
<b>Number of Sample Measurements (N)</b>	6	6	6	6	6

The acceptable range is based on quality specifications:  $\pm 6 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 6 \mu\text{g/L}$  at concentrations less than or equal to  $30 \mu\text{g/L}$ . These quality specifications were established by New York State Department of Health's Wadsworth Center, the PT Program organizer.



## Results for Event #2, 2023: Performance of Participating Laboratories






Whole Blood As (µg/L)						
Lab Code	Method	BE23-06	BE23-07	BE23-08	BE23-09	BE23-10
	<b>Target</b>	<b>15.5</b>	<b>10.5</b>	<b>2.37</b>	<b>45.8</b>	<b>5.5</b>
103	ICP-MS/MS	16.4	10.7	2.44	46.4	5.55
110	ICP-MS/MS	16.5	10.1	2.48	47.3	5.72
247	ICP-MS/MS	14.7	9.87	2.22	41.8	5.05
264	ICP-MS	14.44	10.78	2.42	47.00	5.69
293	DRC/CC-ICP-MS	15.73	11.1	2.49	47.92	5.7
391	ICP-MS	*1.60 ↓	*1.4 ↓	*0.81	*2.09 ↓	*0.5
597	ICP-MS/MS	15.0	10.6	2.18	44.4	5.07

Based on the grading criteria for As in Whole Blood, 91% of results were satisfactory, with 1 of the 7 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #2, 2023: Summary Figures

## Whole Blood As



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the arithmetic mean of all laboratories.

Gray area = acceptable range based on quality specifications:

±6 µg/L or ±20% around the target value, whichever is greater; thus, it is fixed at ±6 µg/L at concentrations less than or equal to 30 µg/L.



## Results for Event #2, 2023: Summary Statistics

Whole Blood Cd (µg/L)					
	BE23-06	BE23-07	BE23-08	BE23-09	BE23-10
<b>Target (Robust Mean (x*))</b>	3.25	0.69	8.1	0.98	13.0
<b>Upper Limit</b>	4.25	1.69	9.3	1.98	14.6
<b>Lower Limit</b>	2.25	0.00	6.9	0.00	11.4
<b>Robust SD (s*)</b>	0.17	0.04	0.5	0.05	0.8
<b>Robust RSD (%)</b>	5.2	5.4	6.2	5.1	6.2
<b>Number of Sample Measurements (N)</b>	11	11	12	11	12
<b>Standard Uncertainty (u)</b>	0.06	0.01	0.2	0.02	0.3

The acceptable range is based on quality specifications:  $\pm 1 \mu\text{g/L}$  or  $\pm 15\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1 \mu\text{g/L}$  at concentrations less than or equal to  $6.7 \mu\text{g/L}$ . These quality specifications are based on those used by US OSHA for occupational exposure.





## Results for Event #2, 2023: Performance of Participating Laboratories






Whole Blood Cd (µg/L)						
Lab Code	Method	BE23-06	BE23-07	BE23-08	BE23-09	BE23-10
	Target	3.25	0.69	8.1	0.98	13.0
103	ICP-MS/MS	3.21	0.671	8.41	1.01	13.3
107	ICP-MS/MS	3.230	0.669	8.029	0.982	12.895
110	ICP-MS	3.33	0.70	8.35	0.99	13.5
116	ICP-MS/MS	3.28	<1.50	7.24	<1.50	12.0
247	ICP-MS/MS	2.91	0.674	7.34	0.825	12.1
264	ICP-MS	2.92	0.61	7.72	0.93	12.15
293	DRC/CC-ICP-MS	3.21	0.71	8.100	1.0	12.90
391	ICP-MS	<0.00 ↓	6.91 ↑	10.261 ↑	0.4	16.75 ↑
597	ICP-MS/MS	3.21	0.685	8.03	0.914	13.1
605	ICP-MS	3.35	0.663	8.41	0.992	13.6
606	ICP-MS/MS	3.47	0.816	8.11	1.16	12.7
686	ICP-MS	3.44	0.729	8.67	1.02	13.7

Based on the grading criteria for Cd in Whole Blood, 93% of results were satisfactory, with 1 of the 12 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #2, 2023: Summary Figures

## Whole Blood Cd



**Legend:**

○ HHEAR Labs ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

±1 µg/L or ±15% around the target value, whichever is greater; thus, it is fixed at ±1 µg/L at concentrations less than or equal to 6.7 µg/L.



## Results for Event #2, 2023: Summary Statistics

	Whole Blood Co (µg/L)				
	BE23-06	BE23-07	BE23-08	BE23-09	BE23-10
<b>Target (Arithmetic Mean (<math>\bar{x}</math>))</b>	2.24	0.58	5.96	1.61	0.98
<b>Upper Limit</b>	3.74	2.08	7.46	3.11	2.48
<b>Lower Limit</b>	0.74	0.00	4.46	0.11	0.00
<b>Arithmetic SD (s)</b>	0.13	0.03	0.21	0.09	0.20
<b>Arithmetic RSD (%)</b>	5.8	4.6	3.5	5.6	20
<b>Number of Sample Measurements (N)</b>	7	7	8	8	8

The acceptable range is based on quality specifications:  $\pm 1.5 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1.5 \mu\text{g/L}$  at concentrations less than or equal to  $7.5 \mu\text{g/L}$ . These quality specifications were established based on discussions with the US FDA, and represent a consensus from a network of Trace Element PT program organizers



### Results for Event #2, 2023: Performance of Participating Laboratories






Whole Blood Co (µg/L)						
Lab Code	Method	BE23-06	BE23-07	BE23-08	BE23-09	BE23-10
	<b>Target</b>	<b>2.24</b>	<b>0.58</b>	<b>5.96</b>	<b>1.61</b>	<b>0.98</b>
103	ICP-MS/MS	2.30	0.600	6.04	1.68	0.892
110	ICP-MS/MS	2.38	0.57	6.22	1.71	0.91
247	ICP-MS/MS	2.19	0.599	5.66	1.57	0.831
255	ICP-MS	2.1	0.53	5.7	1.5	0.82
264	ICP-MS	2.07	0.59	6.11	1.70	0.90
293	DRC/CC-ICP-MS	2.39	0.61	6.15	1.64	0.90
391	ICP-MS	*0.77	*5.32 ↑	5.81	1.48	1.38
597	ICP-MS/MS	2.28	0.578	6.03	1.61	1.17

Based on the grading criteria for Co in Whole Blood, 98% of results were satisfactory, with 0 of the 8 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #2, 2023: Summary Figures

## Whole Blood Co



**Legend:**  
 ○ HHEAR Labs    ● Other Labs  
 Horizontal purple line = assigned target value based on the arithmetic mean of all laboratories.  
 Gray area = acceptable range based on quality specifications:  
 $\pm 1.5 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1.5 \mu\text{g/L}$  at concentrations less than or equal to  $7.5 \mu\text{g/L}$ .



## Results for Event #2, 2023: Summary Statistics

Whole Blood Cr (µg/L)					
	BE23-06	BE23-07	BE23-08	BE23-09	BE23-10
Target (Arithmetic Mean ( $\bar{x}$ ))	5.5	0.67	1.45	2.4	12.9
Upper Limit	7.5	2.67	3.45	4.4	15.5
Lower Limit	3.5	0.00	0.00	0.4	10.3
Arithmetic SD (s)	0.3	0.11	0.09	0.6	0.6
Arithmetic RSD (%)	5.5	16	6.2	25	4.7
Number of Sample Measurements (N)	7	5	6	8	8

The acceptable range is based on quality specifications:  $\pm 2 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 2 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ . These quality specifications were established based on discussions with the US FDA, and represent a consensus from a network of Trace Element PT program organizers



## Results for Event #2, 2023: Performance of Participating Laboratories






Whole Blood Cr (µg/L)						
Lab Code	Method	BE23-06	BE23-07	BE23-08	BE23-09	BE23-10
Target		5.5	0.67	1.45	2.4	12.9
103	ICP-MS/MS	5.15	<1.50	<1.50	2.06	13.3
110	ICP-MS/MS	5.82	0.83	1.56	2.59	13.2
247	ICP-MS/MS	5.15	0.653	1.44	2.37	11.9
255	ICP-MS	5.8	<1.0	1.4	2.6	13
264	ICP-MS	5.36	*1.46	1.46	3.35	13.87
293	DRC/CC-ICP-MS	5.80	0.69	1.53	2.58	13.31
391	ICP-MS	*0.37 ↓	0.61	*0.83	1.10	12.42
597	ICP-MS/MS	5.43	0.545	1.31	2.46	12.4

Based on the grading criteria for Cr in Whole Blood, 98% of results were satisfactory, with 0 of the 8 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #2, 2023: Summary Figures

## Whole Blood Cr



**Legend:**  
 ○ HHEAR Labs    ● Other Labs  
 Horizontal purple line = assigned target value based on the arithmetic mean of all laboratories.  
 Gray area = acceptable range based on quality specifications:  
 $\pm 2 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 2 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ .





## Results for Event #2, 2023: Summary Statistics

	Whole Blood Hg (µg/L)				
	BE23-06	BE23-07	BE23-08	BE23-09	BE23-10
<b>Target (Robust Mean (x*))</b>	19.8	1.62	0.86	10.3	3.7
<b>Upper Limit</b>	25.7	4.62	3.86	13.4	6.7
<b>Lower Limit</b>	13.9	0.00	0.00	7.2	0.7
<b>Robust SD (s*)</b>	2.2	0.21	0.14	1.6	0.8
<b>Robust RSD (%)</b>	11	13	16	16	22
<b>Number of Sample Measurements (N)</b>	12	11	10	12	12
<b>Standard Uncertainty (u)</b>	0.8	0.08	0.06	0.6	0.3

The acceptable range is based on quality specifications:  $\pm 3 \mu\text{g/L}$  or  $\pm 30\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 3 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ . These quality specifications were established by New York State Department of Health's Wadsworth Center, the PT Program organizer.



### Results for Event #2, 2023: Performance of Participating Laboratories






Whole Blood Hg (µg/L)						
Lab Code	Method	BE23-06	BE23-07	BE23-08	BE23-09	BE23-10
	Target	19.8	1.62	0.86	10.3	3.7
103	ICP-MS/MS	20.3	1.71	0.913	10.9	4.04
107	ICP-MS/MS	20.53	1.53	0.76	10.56	3.87
110	ICP-MS	18.6	1.19	0.57	9.88	2.95
116	ICP-MS/MS	20.0	<1.50	<1.50	11.0	4.15
247	ICP-MS/MS	18.6	1.59	0.946	9.81	3.82
264	ICP-MS	17.21	1.88	0.93	9.04	3.19
293	DRC/CC-ICP-MS	14.9	1.26	0.7	8.40	2.79
391	CV-AAS	20.8	2.44	*-0.7	7.76	2.39
597	ICP-MS/MS	18.5	1.44	0.763	9.72	3.52
605	ICP-MS	22.0	1.70	0.994	12.2	4.61
606	ICP-MS/MS	25.0	1.69	0.880	12.8	4.87
686	ICP-MS	21.6	1.65	1.01	11.6	4.33

Based on the grading criteria for Hg in Whole Blood, 98% of results were satisfactory, with 0 of the 12 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



## Results for Event #2, 2023: Summary Figures

### Whole Blood Hg



**Legend:**

○ HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

$\pm 3 \mu\text{g/L}$  or  $\pm 30\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 3 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ .



## Results for Event #2, 2023: Summary Statistics

Whole Blood Mn (µg/L)					
	BE23-06	BE23-07	BE23-08	BE23-09	BE23-10
<b>Target (Arithmetic Mean (<math>\bar{x}</math>))</b>	7.2	21.7	21.8	8.9	22.4
<b>Upper Limit</b>	10.2	25.4	25.5	11.9	26.2
<b>Lower Limit</b>	4.2	18.0	18.1	5.9	18.6
<b>Arithmetic SD (s)</b>	0.6	2.3	2.3	0.7	2.2
<b>Arithmetic RSD (%)</b>	8.6	11	11	7.9	9.8
<b>Number of Sample Measurements (N)</b>	6	8	8	7	8

The acceptable range is based on quality specifications:

$\pm 3 \mu\text{g/L}$  or  $\pm 17\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 3 \mu\text{g/L}$  at concentrations less than or equal to  $17.7 \mu\text{g/L}$ . These quality specifications were recently proposed by a network of Trace Element PT program organizers (Praamsma M, et al. An assessment of clinical laboratory performance for the determination of manganese in blood and urine. Clinical Chemistry Laboratory Medicine 2016; 54(12): 1921-1928).



### Results for Event #2, 2023: Performance of Participating Laboratories






Whole Blood Mn (µg/L)						
Lab Code	Method	BE23-06	BE23-07	BE23-08	BE23-09	BE23-10
	Target	7.2	21.7	21.8	8.9	22.4
107	ICP-MS/MS	7.52	21.82	21.86	9.32	22.77
110	ICP-MS	7.3	20.9	21.5	9.7	22.8
247	ICP-MS/MS	6.92	20.3	20.3	8.46	21.1
264	ICP-MS	*14.57 ↑	25.65 ↑	26.32 ↑	*13.61 ↑	27.14 ↑
293	DRC/CC-ICP-MS	6.2	18.18	18.4	8.2	20.65
391	ICP-MS	*19.8 ↑	24.37	23.0	8.0	20.31
597	ICP-MS/MS	7.25	21.0	20.4	9.21	21.5
606	ICP-MS/MS	8.05	21.7	22.4	9.55	22.8

Based on the grading criteria for Mn in Whole Blood, 85% of results were satisfactory, with 1 of the 8 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



## Results for Event #2, 2023: Summary Figures

### Whole Blood Mn



**Legend:**

○ HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the arithmetic mean of all laboratories. Gray area = acceptable range based on quality specifications:

$\pm 3 \mu\text{g/L}$  or  $\pm 17\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 3 \mu\text{g/L}$  at concentrations less than or equal to  $17.7 \mu\text{g/L}$ .



## Results for Event #2, 2023: Summary Statistics

Whole Blood Pb (µg/dL)					
	BE23-06	BE23-07	BE23-08	BE23-09	BE23-10
<b>Target (Robust Mean (x*))</b>	0.72	5.9	2.51	18.1	1.14
<b>Upper Limit</b>	2.72	7.9	4.51	20.1	3.14
<b>Lower Limit</b>	0.00	3.9	0.51	16.1	0.00
<b>Robust SD (s*)</b>	0.10	0.3	0.19	0.9	0.10
<b>Robust RSD (%)</b>	14	5.1	7.6	5.2	8.8
<b>Number of Sample Measurements (N)</b>	8	13	11	13	11
<b>Standard Uncertainty (u)</b>	NA	0.1	0.07	0.3	0.04

The acceptable range is based on quality specifications:  $\pm 2 \mu\text{g/dL}$  or  $\pm 10\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 2 \mu\text{g/dL}$  at concentrations less than or equal to  $20 \mu\text{g/dL}$ . These quality specifications are recommended by the Clinical Laboratory Standards Institute (CLSI, C40-A2) and have been proposed for use in proficiency testing programs approved under CLIA by the Centers for Medicare and Medicaid Services (CMS) in the USA. (<https://clsi.org/standards/products/clinical-chemistry-and-toxicology/documents/c40/>)

An arithmetic mean, SD, RSD and n are provided for sample BE23-06.



### Results for Event #2, 2023: Performance of Participating Laboratories

Whole Blood Pb (µg/dL)						
Lab Code	Method	BE23-06	BE23-07	BE23-08	BE23-09	BE23-10
Target		0.72	5.9	2.51	18.1	1.14
103	ICP-MS/MS	0.676	6.02	2.57	18.7	1.16
107	ICP-MS/MS	0.697	5.980	2.501	18.337	1.142
110	ICP-MS	0.63	5.76	2.35	17.5	1.08
116	ICP-MS/MS	<3.00	5.85	<3.00	17.5	<3.00
247	ICP-MS/MS	0.62	5.57	2.34	16.8	1.03
264	ICP-MS	0.65	5.24	2.22	16.10	1.01
293	DRC/CC-ICP-MS	0.83	6.00	2.48	18.4	1.24
343	ASV-LeadCare	<3.3	4.0	<3.3	17.5	<3.3
391	ETAAS-Z	0.91	6.87	3.20	21.9	↑ 1.12
597	ICP-MS/MS	0.728	6.00	2.50	18.1	1.14
605	ICP-MS	<1.00	6.21	2.71	19.2	1.27
606	ICP-MS/MS	<1.00	5.90	2.50	18.1	1.13
686	ICP-MS	<1.00	6.18	2.65	18.7	1.27






Based on the grading criteria for Pb in Whole Blood, 98% of results were satisfactory, with 0 of the 13 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.





# Results for Event #2, 2023: Summary Figures

## Whole Blood Pb



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

±2 µg/dL or ±10% around the target value, whichever is greater; thus, it is fixed at ±2 µg/dL at concentrations less than or equal to 20 µg/dL.



## Results for Event #2, 2023: Laboratory Data and Summary Statistics

### Whole Blood Mo (µg/L)

Lab Code	Method	BE23-06	BE23-07	BE23-08	BE23-09	BE23-10
103	ICP-MS/MS	3.95	<1.50	<1.50	7.20	5.09
110	ICP-MS/MS	4.08	1.48	0.65	7.04	5.14
264	ICP-MS	3.24	1.29	0.91	3.85	2.78
442	DRC/CC-ICP-MS	3.76	1.44	0.66	6.72	4.84
597	ICP-MS/MS	3.71	1.31	0.628	6.21	4.53

### Summary Statistics






	BE23-06	BE23-07	BE23-08	BE23-09	BE23-10
Arithmetic Mean ( $\bar{x}$ )	3.7	1.38	0.71	6.2	4.5
Arithmetic SD (s)	0.3	0.09	0.13	1.4	1.0
Arithmetic RSD (%)	8.5	6.5	18	23	22
Number of Sample Measurements (N)	5	4	4	5	5

\*Denotes a statistical Outlier.



## Results for Event #2, 2023: Summary Figures

### Whole Blood Mo



**Legend:**

○ HHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #2, 2023: Laboratory Data and Summary Statistics

### Whole Blood Sb (µg/L)

Lab Code	Method	BE23-06	BE23-07	BE23-08	BE23-09	BE23-10
103	ICP-MS/MS	1.02	4.24	0.578	8.28	2.89
110	ICP-MS/MS	1.35	5.82	0.78	10.4	3.66
264	ICP-MS	1.86	5.90	0.65	10.25	*5.81
293	DRC/CC-ICP-MS	1.3	5.9	0.7	8.9	3.2
442	DRC/CC-ICP-MS	1.81	6.24	0.880	10.2	3.88
597	ICP-MS/MS	1.37	6.55	0.779	10.2	3.62

### Summary Statistics






	BE23-06	BE23-07	BE23-08	BE23-09	BE23-10
Arithmetic Mean ( $\bar{x}$ )	1.5	5.8	0.73	9.7	3.5
Arithmetic SD (s)	0.3	0.8	0.11	0.9	0.4
Arithmetic RSD (%)	22	14	15	9.3	11
Number of Sample Measurements (N)	6	6	6	6	5

\*Denotes a statistical Outlier.



# Results for Event #2, 2023: Summary Figures

## Whole Blood Sb



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #2, 2023: Laboratory Data and Summary Statistics

### Whole Blood Se (µg/L)

Lab Code	Method	BE23-06	BE23-07	BE23-08	BE23-09	BE23-10
103	ICP-MS/MS	159	132	366	192	219
107	ICP-MS/MS	160.7	132.9	355.5	187.5	218.7
110	ICP-MS/MS	157	120	356	187	213
247	ICP-MS/MS	162	138	360	183	224
264	ICP-MS	145	136	366	201	229
293	DRC/CC-ICP-MS	129	110	294	153	169
597	ICP-MS/MS	147	125	340	179	203

### Summary Statistics






	BE23-06	BE23-07	BE23-08	BE23-09	BE23-10
Arithmetic Mean ( $\bar{x}$ )	151	128	348	183	211
Arithmetic SD (s)	12	10	25	15	20
Arithmetic RSD (%)	7.9	7.8	7.2	8.2	9.5
Number of Sample Measurements (N)	7	7	7	7	7

\*Denotes a statistical Outlier.



# Results for Event #2, 2023: Summary Figures

## Whole Blood Se



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #2, 2023: Laboratory Data and Summary Statistics

Whole Blood TI (µg/L)						
Lab Code	Method	BE23-06	BE23-07	BE23-08	BE23-09	BE23-10
103	ICP-MS/MS	0.209	1.42	0.140	3.01	0.448
110	ICP-MS/MS	0.22	1.32	0.16	3.01	0.41
264	ICP-MS	0.24	1.34	0.15	2.89	0.42
293	DRC/CC-ICP-MS	0.25	1.39	0.160	2.89	0.43
597	ICP-MS/MS	0.219	1.42	0.152	2.99	0.430
Summary Statistics						
		BE23-06	BE23-07	BE23-08	BE23-09	BE23-10
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>		0.228	1.38	0.152	2.96	0.428
<b>Arithmetic SD (s)</b>		0.017	0.05	0.008	0.06	0.014
<b>Arithmetic RSD (%)</b>		7.5	3.6	5.3	2.1	3.3
<b>Number of Sample Measurements (N)</b>		5	5	5	5	5






\*Denotes a statistical Outlier.





## Results for Event #2, 2023: Summary Figures

### Whole Blood TI



#### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #2, 2023: Laboratory Data and Summary Statistics

### Whole Blood Ba (µg/L)

Lab Code	Method	BE23-06	BE23-07	BE23-08	BE23-09	BE23-10
110	ICP-MS/MS	2.09	3.35	1.49	7.35	4.15
597	ICP-MS/MS	2.36	3.63	1.30	7.44	3.56

### Summary Statistics

	BE23-06	BE23-07	BE23-08	BE23-09	BE23-10
Arithmetic Mean ( $\bar{x}$ )	2.2	3.5	1.40	7.39	3.9
Arithmetic SD (s)	0.2	0.2	0.13	0.06	0.4
Arithmetic RSD (%)	8.5	5.7	9.3	0.81	10
Number of Sample Measurements (N)	2	2	2	2	2

\*Denotes a statistical Outlier.



## Results for Event #2, 2023: Laboratory Data and Summary Statistics

Whole Blood Be (µg/L)						
Lab Code	Method	BE23-06	BE23-07	BE23-08	BE23-09	BE23-10
110	ICP-MS/MS	0.65	4.76	2.23	5.41	0.84
597	ICP-MS/MS	0.511	5.31	2.11	6.24	0.816

Summary Statistics						
	BE23-06	BE23-07	BE23-08	BE23-09	BE23-10	
Arithmetic Mean ( $\bar{x}$ )	0.58	5.0	2.17	5.8	0.83	
Arithmetic SD (s)	0.10	0.4	0.08	0.6	0.02	
Arithmetic RSD (%)	17	7.7	3.7	10	2.1	
Number of Sample Measurements (N)	2	2	2	2	2	

\*Denotes a statistical Outlier.



## Results for Event #2, 2023: Laboratory Data and Summary Statistics

### Whole Blood Cs (µg/L)

Lab Code	Method	BE23-06	BE23-07	BE23-08	BE23-09	BE23-10
110	ICP-MS/MS	3.34	1.17	1.07	1.93	2.11
597	ICP-MS/MS	3.20	1.25	1.03	1.90	2.04

### Summary Statistics

	BE23-06	BE23-07	BE23-08	BE23-09	BE23-10
Arithmetic Mean ( $\bar{x}$ )	3.27	1.21	1.05	1.92	2.08
Arithmetic SD (s)	0.10	0.06	0.03	0.02	0.05
Arithmetic RSD (%)	3.1	4.7	2.9	1.1	2.4
Number of Sample Measurements (N)	2	2	2	2	2

\*Denotes a statistical Outlier.



## Results for Event #2, 2023: Laboratory Data and Summary Statistics

### Whole Blood Cu ( $\mu\text{g/L}$ )

Lab Code	Method	BE23-06	BE23-07	BE23-08	BE23-09	BE23-10
110	ICP-MS/MS	658	2290	1010	942	782
247	ICP-MS/MS	620	2326	945	851	737
597	ICP-MS/MS	577	2180	881	844	669

### Summary Statistics

	BE23-06	BE23-07	BE23-08	BE23-09	BE23-10
Arithmetic Mean ( $\bar{x}$ )	620	2270	950	880	730
Arithmetic SD (s)	40	80	60	50	60
Arithmetic RSD (%)	6.5	3.5	6.3	5.7	8.2
Number of Sample Measurements (N)	3	3	3	3	3

\*Denotes a statistical Outlier.



## Results for Event #2, 2023: Laboratory Data and Summary Statistics

### Whole Blood Ni (µg/L)

Lab Code	Method	BE23-06	BE23-07	BE23-08	BE23-09	BE23-10
103	ICP-MS/MS	1.02	<1.50	1.78	5.64	3.43
110	ICP-MS/MS	1.67	0.91	2.26	5.92	3.45
597	ICP-MS/MS	1.00	0.555	1.79	5.73	2.91

### Summary Statistics

	BE23-06	BE23-07	BE23-08	BE23-09	BE23-10
Arithmetic Mean ( $\bar{x}$ )	1.2	NA	1.9	5.76	3.3
Arithmetic SD (s)	0.4	NA	0.3	0.14	0.3
Arithmetic RSD (%)	33	NA	16	2.4	9.1
Number of Sample Measurements (N)	3	NA	3	3	3

\*Denotes a statistical Outlier.

Statistical data was not calculated for BE23-07 based on a lack of consensus among participating labs.



## Results for Event #2, 2023: Laboratory Data and Summary Statistics

### Whole Blood Pt (µg/L)

Lab Code	Method	BE23-06	BE23-07	BE23-08	BE23-09	BE23-10
110	ICP-MS/MS	0.975	3.34	1.39	0.537	6.03
293	DRC/CC-ICP-MS	0.75	3.00	1.11	0.30	4.71

### Summary Statistics

	BE23-06	BE23-07	BE23-08	BE23-09	BE23-10
Arithmetic Mean ( $\bar{x}$ )	0.9	3.2	1.3	NA	5.4
Arithmetic SD (s)	0.2	0.2	0.2	NA	0.9
Arithmetic RSD (%)	19	6.3	16	NA	17
Number of Sample Measurements (N)	2	2	2	NA	2

\*Denotes a statistical Outlier.

Statistical data was not calculated for BE23-09 based on a lack of consensus among participating labs.



## Results for Event #2, 2023: Laboratory Data and Summary Statistics

### Whole Blood Sn (µg/L)

Lab Code	Method	BE23-06	BE23-07	BE23-08	BE23-09	BE23-10
110	ICP-MS/MS	5.86	0.50	3.39	0.36	1.06
597	ICP-MS/MS	5.66	0.385	3.15	0.350	0.998

### Summary Statistics

	BE23-06	BE23-07	BE23-08	BE23-09	BE23-10
Arithmetic Mean ( $\bar{x}$ )	5.8	0.44	3.3	0.355	1.03
Arithmetic SD (s)	0.1	0.08	0.2	0.007	0.04
Arithmetic RSD (%)	2.4	18	5.2	2.0	3.9
Number of Sample Measurements (N)	2	2	2	2	2

\*Denotes a statistical Outlier.





## Results for Event #2, 2023: Laboratory Data and Summary Statistics

Whole Blood Sr ( $\mu\text{g/L}$ )						
Lab Code	Method	BE23-06	BE23-07	BE23-08	BE23-09	BE23-10
103	ICP-MS/MS	38.9	46.4	20.7	25.2	35.4
110	ICP-MS/MS	39.1	45.4	21.2	24.8	36.5
597	ICP-MS/MS	37.5	46.6	19.9	24.1	36.1
Summary Statistics						
		BE23-06	BE23-07	BE23-08	BE23-09	BE23-10
Arithmetic Mean ( $\bar{x}$ )		38.5	46.1	20.6	24.7	36.0
Arithmetic SD (s)		0.9	0.6	0.7	0.6	0.6
Arithmetic RSD (%)		2.3	1.3	3.4	2.4	1.7
Number of Sample Measurements (N)		3	3	3	3	3

\*Denotes a statistical Outlier.



## Results for Event #2, 2023: Laboratory Data and Summary Statistics

Whole Blood Ti (µg/L)						
Lab Code	Method	BE23-06	BE23-07	BE23-08	BE23-09	BE23-10
200	DRC/CC-ICP-MS	3.7	3.7	4.2	7.2	6.0
597	ICP-MS/MS	3.84	4.43	4.78	8.54	6.40

Summary Statistics						
	BE23-06	BE23-07	BE23-08	BE23-09	BE23-10	
Arithmetic Mean ( $\bar{x}$ )	3.77	4.1	4.5	7.9	6.2	
Arithmetic SD (s)	0.10	0.5	0.4	0.9	0.3	
Arithmetic RSD (%)	2.7	12	8.9	11	4.8	
Number of Sample Measurements (N)	2	2	2	2	2	

\*Denotes a statistical Outlier.



## Results for Event #2, 2023: Laboratory Data and Summary Statistics

### Whole Blood U (µg/L)

Lab Code	Method	BE23-06	BE23-07	BE23-08	BE23-09	BE23-10
103	ICP-MS/MS	0.0593	0.175	<0.0500	0.0743	0.138
110	ICP-MS/MS	0.0664	0.185	0.0310	0.0772	0.142
597	ICP-MS/MS	0.0633	0.183	0.0346	0.0730	0.130

### Summary Statistics

	BE23-06	BE23-07	BE23-08	BE23-09	BE23-10
Arithmetic Mean ( $\bar{x}$ )	0.063	0.181	0.033	0.075	0.137
Arithmetic SD (s)	0.004	0.005	0.003	0.002	0.006
Arithmetic RSD (%)	6.3	2.8	9.1	2.9	4.4
Number of Sample Measurements (N)	3	3	2	3	3

\*Denotes a statistical Outlier.



## Results for Event #2, 2023: Laboratory Data and Summary Statistics

Whole Blood V (µg/L)						
Lab Code	Method	BE23-06	BE23-07	BE23-08	BE23-09	BE23-10
110	ICP-MS/MS	2.32	0.27	0.49	1.17	3.12
597	ICP-MS/MS	2.18	0.251	0.443	1.18	2.92

Summary Statistics						
	BE23-06	BE23-07	BE23-08	BE23-09	BE23-10	
Arithmetic Mean ( $\bar{x}$ )	2.25	0.261	0.47	1.175	3.0	
Arithmetic SD (s)	0.10	0.013	0.03	0.007	0.1	
Arithmetic RSD (%)	4.4	5.2	6.4	0.60	4.6	
Number of Sample Measurements (N)	2	2	2	2	2	

\*Denotes a statistical Outlier.



## Results for Event #2, 2023: Laboratory Data and Summary Statistics

Whole Blood W ( $\mu\text{g/L}$ )						
Lab Code	Method	BE23-06	BE23-07	BE23-08	BE23-09	BE23-10
110	ICP-MS/MS	0.15	1.34	5.01	0.62	2.15
200	ICP-MS	0.20	1.62	5.70	0.74	2.30
597	ICP-MS/MS	0.161	1.46	4.78	0.610	2.01
Summary Statistics						
		BE23-06	BE23-07	BE23-08	BE23-09	BE23-10
Arithmetic Mean ( $\bar{x}$ )		0.17	1.47	5.2	0.66	2.15
Arithmetic SD (s)		0.03	0.14	0.5	0.07	0.15
Arithmetic RSD (%)		18	9.5	9.6	11	6.7
Number of Sample Measurements (N)		3	3	3	3	3

\*Denotes a statistical Outlier.



## Results for Event #2, 2023: Laboratory Data and Summary Statistics

### Whole Blood Zn (µg/L)

Lab Code	Method	BE23-06	BE23-07	BE23-08	BE23-09	BE23-10
110	ICP-MS/MS	5220	5720	5760	5330	5270
247	ICP-MS/MS	4453	5690	5060	4569	4700
597	ICP-MS/MS	4960	5810	5370	5200	4880

### Summary Statistics

	BE23-06	BE23-07	BE23-08	BE23-09	BE23-10
Arithmetic Mean ( $\bar{x}$ )	4900	5740	5400	5000	5000
Arithmetic SD (s)	400	60	400	400	300
Arithmetic RSD (%)	8.2	1.1	7.4	8.1	5.9
Number of Sample Measurements (N)	3	3	3	3	3

\*Denotes a statistical Outlier.



Results for Event #2, 2023:  
Additional Elements in Whole Blood

Whole Blood Al (µg/L)

Lab Code	Method	BE23-06	BE23-07	BE23-08	BE23-09	BE23-10
597	ICP-MS/MS	4.02	6.73	2.92	4.28	4.29

Whole Blood Bi (µg/L)

Lab Code	Method	BE23-06	BE23-07	BE23-08	BE23-09	BE23-10
597	ICP-MS/MS	<0.0187	0.0219	<0.0187	<0.0187	<0.0187

Whole Blood Mg (µg/L)

Lab Code	Method	BE23-06	BE23-07	BE23-08	BE23-09	BE23-10
597	ICP-MS/MS	26800	27900	26700	28900	27600

Whole Blood Te (µg/L)

Lab Code	Method	BE23-06	BE23-07	BE23-08	BE23-09	BE23-10
110	ICP-MS/MS	0.01	0.00	0.01	0.01	0.00

Whole Blood Th (µg/L)

Lab Code	Method	BE23-06	BE23-07	BE23-08	BE23-09	BE23-10
597	ICP-MS/MS	0.0268	0.0230	0.0221	0.0152	0.0266



**Department  
of Health**

**Wadsworth  
Center**

**Event #2, 2023**

**Trace Elements in  
Urine**

**Wadsworth Center**  
NEW YORK STATE DEPARTMENT OF HEALTH  
*Trace Elements Laboratory*





## Event #2, 2023: Trace Elements in Urine

### PT Materials

Urine was collected from volunteer donors into polyethylene containers and stored at 4°C. Following collection, urine was acidified to 1% (v/v) with nitric acid and mixed with a sulfamic acid solution (stock solution contained 200 mg/mL sulfamic acid and 10% (v/v) Triton-X 100) to a final concentration of 1% (v/v) to stabilize Hg. Urine was stored frozen at -80°C pending further preparation. The urine was thawed at room temperature and precipitated salts removed by centrifugation. Urine supernatants were combined into five separate pools. Each urine pool was supplemented with arsenic (As), barium (Ba), beryllium (Be), cadmium (Cd), cobalt (Co), chromium (Cr), mercury (Hg), manganese (Mn), lead (Pb), thallium (Tl), uranium (U), aluminum (Al), cesium (Cs), copper (Cu), molybdenum (Mo), nickel (Ni), platinum (Pt), antimony (Sb), selenium (Se), tin (Sn), strontium (Sr), tellurium (Te), titanium (Ti), vanadium (V), tungsten (W), and zinc (Zn). PT samples were stored at -80°C until the week of the PT event, when they were thawed at 4°C prior to circulation to laboratories for analysis.

### Graded Elements

Eleven elements in urine are formally graded: As, Ba, Be, Cd, Co, Cr, Hg, Mn, Pb, Tl, and U. Target values for the graded elements are assigned to these pools based on (a) the robust mean calculated from data reported by all laboratories, or (b) if a robust mean is not possible, the arithmetic mean after outlier deletion.

### Additional Elements

An additional 23 elements were reported by at least one participant: Ag, Al, B, Bi, Cs, Cu, Fe, I, Li, Mg, Mo, Ni, Pt, Sb, Se, Sn, Sr, Te, Th, Ti, V, W, and Zn. These data are included here to provide a more complete characterization of the PT materials. All results reported by participant laboratories are tabulated and organized by lab code. The PT data are graphed for visual comparison purposes for all elements where at least five laboratories reported a value greater than the LOD. A statistical summary table is provided for samples where at least two comparable values were reported as above the LOD.

The summary statistics for the additional elements are provided for educational purposes only, i.e., no acceptable response is implied. However, it is expected that each laboratory would wish to investigate a potential source of bias if warranted by these data. Future events might result in additional elements becoming graded if a consensus can be reached regarding desired quality specifications.



## Results for Event #2, 2023: Summary Statistics

	Urine As ( $\mu\text{g/L}$ )				
	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
<b>Target (Robust Mean (<math>x^*</math>))</b>	4.0	50.5	10.1	5.4	32.5
<b>Upper Limit</b>	10.0	60.6	16.1	11.4	39.0
<b>Lower Limit</b>	0.0	40.4	4.1	0.0	26.0
<b>Robust SD (<math>s^*</math>)</b>	0.4	2.7	0.6	0.4	1.6
<b>Robust RSD (%)</b>	9.2	5.3	5.9	7.4	4.9
<b>Number of Sample Measurements (N)</b>	14	15	15	14	15
<b>Standard Uncertainty (<math>u</math>)</b>	0.1	0.9	0.2	0.1	0.5

The acceptable range is based on quality specifications:  $\pm 6 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 6 \mu\text{g/L}$  at concentrations less than or equal to  $30 \mu\text{g/L}$ . These quality specifications are based on the same criteria used by the US Centers for Disease Control Prevention (CDC) for public health labs participating in the Laboratory Response Network (LRN) PT program for Toxic Metals.



### Results for Event #2, 2023: Performance of Participating Laboratories






		Urine As (µg/L)				
Lab Code	Method	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
Target		4.0	50.5	10.1	5.4	32.5
103	ICP-MS/MS	3.90	52.5	10.4	5.30	33.2
107	DRC/CC-ICP-MS	3.79	48.54	9.58	4.88	32.22
110	DRC/CC-ICP-MS	4.49	52.6	10.7	5.76	34.0
116	ICP-MS/MS	12.2 ↑	54.1	15.8	15.2 ↑	40.1 ↑
147	ICP-MS	2.54	49.0	9.44	4.12	32.1
220	DRC/CC-ICP-MS	4.06	51.5	10.5	5.64	33.7
264	ICP-MS	4.66	45.15	9.32	5.93	28.78
293	DRC/CC-ICP-MS	3.81	50.99	10.06	5.08	31.18
391	ICP-MS	5.903	59.751	10.35	5.575	40.007 ↑
399	DRC/CC-ICP-MS	3.87	49.5	10.2	5.24	32.3
442	ICP-MS/MS	3.63	47.6	9.51	5.32	31.2
597	ICP-MS/MS	3.85	47.7	9.57	5.18	31.0
605	ICP-MS	3.96	50.3	10.1	5.28	32.4
606	ICP-MS/MS	4.37	51.7	10.6	5.58	33.1
686	DRC/CC-ICP-MS	<6.00	50.2	10.1	<6.00	31.8

Based on the grading criteria for As in Urine, 95% of results were satisfactory, with 1 of the 15 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #2, 2023: Summary Figures

## Urine As



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

±6 µg/L or ±20% around the target value, whichever is greater; thus, it is fixed at ±6 µg/L at concentrations less than or equal to 30 µg/L.



## Results for Event #2, 2023: Summary Statistics

	Urine Ba (µg/L)				
	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
<b>Target (Robust Mean (x*))</b>	1.33	7.8	0.66	3.76	0.55
<b>Upper Limit</b>	2.33	9.4	1.66	4.76	1.55
<b>Lower Limit</b>	0.33	6.2	0.00	2.76	0.00
<b>Robust SD (s*)</b>	0.10	0.5	0.05	0.23	0.05
<b>Robust RSD (%)</b>	7.5	6.4	7.6	6.1	9.1
<b>Number of Sample Measurements (N)</b>	12	12	12	12	10
<b>Standard Uncertainty (u)</b>	0.03	0.2	0.02	0.08	0.02

The acceptable range is based on quality specifications:  $\pm 1 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1 \mu\text{g/L}$  at concentrations less than or equal to  $5 \mu\text{g/L}$ . These quality specifications are based on the same criteria used by the US Centers for Disease Control Prevention (CDC) for public health labs participating in the Laboratory Response Network (LRN) PT program for Toxic Metals.



## Results for Event #2, 2023: Performance of Participating Laboratories






		Urine Ba (µg/L)				
Lab Code	Method	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
Target		1.33	7.8	0.66	3.76	0.55
107	ICP-MS	1.406	8.152	0.665	3.969	0.529
110	ICP-MS	1.54	8.58	0.73	4.29	0.62
116	ICP-MS/MS	1.18	7.05	0.690	3.58	0.528
147	ICP-MS	1.30	7.95	0.666	3.80	0.659
220	ICP-MS	1.35	8.19	0.659	3.94	0.514
264	ICP-MS	1.21	6.61	0.57	3.33	0.47
293	DRC/CC-ICP-MS	1.34	7.46	0.63	3.69	0.52
399	ICP-MS/MS	1.38	8.28	0.710	3.82	0.577
597	ICP-MS/MS	1.25	7.62	0.671	3.72	0.536
605	ICP-MS	1.33	7.82	0.619	3.71	<0.6
606	ICP-MS/MS	1.29	7.44	0.605	3.50	0.601
686	ICP-MS	1.45	8.04	0.722	3.91	<0.600

Based on the grading criteria for Ba in Urine, 100% of results were satisfactory, with 0 of the 12 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #2, 2023: Summary Figures

## Urine Ba



**Legend:**  
○ HHEAR Labs    ● Other Labs  
Horizontal purple line = assigned target value based on the robust mean of all laboratories.  
Gray area = acceptable range based on quality specifications:  
±1 µg/L or ±20% around the target value, whichever is greater; thus, it is fixed at ±1 µg/L at concentrations less than or equal to 5 µg/L.



## Results for Event #2, 2023: Summary Statistics

	Urine Be (µg/L)				
	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
<b>Target (Robust Mean (x*))</b>	2.06	0.867	0.548	1.11	4.03
<b>Upper Limit</b>	3.06	1.867	1.548	2.11	5.03
<b>Lower Limit</b>	1.06	0.000	0.000	0.11	3.03
<b>Robust SD (s*)</b>	0.07	0.012	0.013	0.04	0.15
<b>Robust RSD (%)</b>	3.4	1.4	2.4	3.3	3.7
<b>Number of Sample Measurements (N)</b>	11	11	11	11	11
<b>Standard Uncertainty (u)</b>	0.03	0.004	0.005	0.01	0.06

The acceptable range is based on quality specifications:  $\pm 1 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1 \mu\text{g/L}$  at concentrations less than or equal to  $5 \mu\text{g/L}$ . These quality specifications are based on the same criteria used by the US Centers for Disease Control Prevention (CDC) for public health labs participating in the Laboratory Response Network (LRN) PT program for Toxic Metals.





## Results for Event #2, 2023: Performance of Participating Laboratories






		Urine Be (µg/L)				
Lab Code	Method	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
Target		2.06	0.867	0.548	1.11	4.03
107	ICP-MS	2.117	0.860	0.519	1.078	3.887
110	ICP-MS	2.03	0.84	0.55	1.10	4.02
116	ICP-MS/MS	1.97	0.868	0.474	1.06	3.82
147	ICP-MS	2.04	0.874	0.534	1.08	3.91
220	ICP-MS	2.11	0.871	0.555	1.14	4.19
264	ICP-MS	2.17	0.88	0.58	1.17	4.26
293	ICP-MS	2.07	0.85	0.54	1.13	4.06
399	ICP-MS/MS	2.05	0.872	0.552	1.09	3.98
597	ICP-MS/MS	2.01	0.942	0.547	1.10	3.95
605	ICP-MS	2.00	0.861	0.557	1.08	4.10
686	ICP-MS	2.13	0.868	0.558	1.15	4.12

Based on the grading criteria for Be in Urine, 100% of results were satisfactory, with 0 of the 11 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #2, 2023: Summary Figures

## Urine Be



**Legend:**  
 ○ HHEAR Labs    ● Other Labs  
 Horizontal purple line = assigned target value based on the robust mean of all laboratories.  
 Gray area = acceptable range based on quality specifications:  
 $\pm 1 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1 \mu\text{g/L}$  at concentrations less than or equal to  $5 \mu\text{g/L}$ .



## Results for Event #2, 2023: Summary Statistics

	Urine Cd (µg/L)				
	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
<b>Target (Robust Mean (x*))</b>	1.97	3.05	0.202	0.74	5.06
<b>Upper Limit</b>	2.97	4.05	1.202	1.74	6.06
<b>Lower Limit</b>	0.97	2.05	0.000	0.00	4.06
<b>Robust SD (s*)</b>	0.14	0.23	0.023	0.07	0.26
<b>Robust RSD (%)</b>	7.1	7.5	11	9.5	5.1
<b>Number of Sample Measurements (N)</b>	16	16	13	15	16
<b>Standard Uncertainty (u)</b>	0.04	0.07	0.008	0.02	0.08

The acceptable range is based on quality specifications:  $\pm 1 \mu\text{g/L}$  or  $\pm 15\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1 \mu\text{g/L}$  at concentrations less than or equal to  $6.6 \mu\text{g/L}$ . These quality specifications are based on the same criteria used by the US Centers for Disease Control Prevention (CDC) for public health labs participating in the Laboratory Response Network (LRN) PT program for Toxic Metals.



### Results for Event #2, 2023: Performance of Participating Laboratories






Lab Code	Method	Urine Cd (µg/L)				
		UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
	<b>Target</b>	<b>1.97</b>	<b>3.05</b>	<b>0.202</b>	<b>0.74</b>	<b>5.06</b>
103	ICP-MS/MS	2.01	3.19	0.216	0.790	5.18
107	DRC/CC-ICP-MS	1.929	2.938	0.167	0.687	5.079
110	ICP-MS	2.16	3.29	0.23	0.82	5.43
116	ICP-MS/MS	1.80	2.86	<0.200	0.665	4.96
147	ICP-MS	2.01	3.21	0.208	0.767	5.07
220	ICP-MS	1.81	2.66	0.197	0.718	4.84
264	ICP-MS	2.07	2.68	0.17	0.83	4.37
293	DRC/CC-ICP-MS	2.09	3.33	0.19	0.8	5.38
324	ICP-MS	2.112	3.070	<1	<1	5.076
391	ICP-MS	0.935 ↓	1.87 ↓	0.228	0.625	2.847 ↓
399	DRC/CC-ICP-MS	2.00	3.21	0.222	0.770	5.27
442	ICP-MS/MS	1.91	3.15	0.16	0.678	5.24
597	ICP-MS/MS	1.93	3.03	0.199	0.688	4.89
605	ICP-MS	1.88	3.06	0.203	0.726	4.94
606	ICP-MS/MS	1.91	2.94	0.215	0.741	4.91
686	ICP-MS	2.10	3.23	<0.240	0.794	5.32

Based on the grading criteria for Cd in Urine, 96% of results were satisfactory, with 1 of the 16 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #2, 2023: Summary Figures

## Urine Cd



**Legend:**  
 ○ HHEAR Labs    ● Other Labs  
 Horizontal purple line = assigned target value based on the robust mean of all laboratories.  
 Gray area = acceptable range based on quality specifications:  
 $\pm 1 \mu\text{g/L}$  or  $\pm 15\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1 \mu\text{g/L}$  at concentrations less than or equal to  $6.6 \mu\text{g/L}$ .



## Results for Event #2, 2023: Summary Statistics

	Urine Co (µg/L)				
	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
<b>Target (Robust Mean (x*))</b>	0.54	3.79	2.02	0.90	5.8
<b>Upper Limit</b>	2.04	5.29	3.52	2.40	7.3
<b>Lower Limit</b>	0.00	2.29	0.52	0.00	4.3
<b>Robust SD (s*)</b>	0.03	0.23	0.11	0.05	0.3
<b>Robust RSD (%)</b>	4.8	6.1	5.4	5.6	5.3
<b>Number of Sample Measurements (N)</b>	12	13	13	12	13
<b>Standard Uncertainty (u)</b>	0.01	0.08	0.04	0.02	0.1

The acceptable range is based on quality specifications:  $\pm 1.5 \mu\text{g/L}$  or  $\pm 15\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1.5 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ . These quality specifications were established based on discussions with the US FDA, and represent a consensus from a network of Trace Element PT program organizers



### Results for Event #2, 2023: Performance of Participating Laboratories






		Urine Co (µg/L)				
Lab Code	Method	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
Target		0.54	3.79	2.02	0.90	5.8
103	ICP-MS/MS	0.530	3.92	2.06	0.877	5.86
107	ICP-MS	0.516	3.736	1.961	0.859	5.721
110	ICP-MS	0.70	4.05	2.19	1.00	6.16
147	ICP-MS	0.556	3.55	1.86	0.878	5.42
220	ICP-MS	0.766	4.07	2.23	1.09	6.22
264	ICP-MS	0.53	3.31	1.74	0.95	5.09
293	DRC/CC-ICP-MS	0.54	4.06	2.07	0.92	6.14
324	ICP-MS	<1	3.718	2.010	<1	5.693
391	ICP-MS	0.413	3.068	1.403	0.458	3.853 ↓
399	DRC/CC-ICP-MS	0.523	3.73	1.98	0.880	5.70
597	ICP-MS/MS	0.548	3.83	2.04	0.870	5.98
605	ICP-MS	0.525	3.83	2.02	0.868	5.81
606	ICP-MS/MS	0.599	3.78	2.11	0.932	5.86

Based on the grading criteria for Co in Urine, 98% of results were satisfactory, with 0 of the 13 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #2, 2023: Summary Figures

## Urine Co



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

±1.5 µg/L or ±15% around the target value, whichever is greater; thus, it is fixed at ±1.5 µg/L at concentrations less than or equal to 10 µg/L.





### Results for Event #2, 2023: Summary Statistics

	Urine Cr (µg/L)				
	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
<b>Target (Robust Mean (x*))</b>	7.3	0.86	3.2	4.1	5.8
<b>Upper Limit</b>	10.3	3.86	6.2	7.1	8.8
<b>Lower Limit</b>	4.3	0.00	0.2	1.1	2.8
<b>Robust SD (s*)</b>	0.8	0.17	0.4	0.5	0.7
<b>Robust RSD (%)</b>	11	20	11	12	12
<b>Number of Sample Measurements (N)</b>	11	11	11	11	11
<b>Standard Uncertainty (u)</b>	0.3	0.07	0.1	0.2	0.3

The acceptable range is based on quality specifications:  $\pm 3 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 3 \mu\text{g/L}$  at concentrations less than or equal to  $15 \mu\text{g/L}$ . These quality specifications were established based on discussions with the US FDA, and represent a consensus from a network of Trace Element PT program organizers



## Results for Event #2, 2023: Performance of Participating Laboratories






Urine Cr ( $\mu\text{g/L}$ )						
Lab Code	Method	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
	Target	7.3	0.86	3.2	4.1	5.8
103	ICP-MS/MS	7.08	1.22	3.07	3.86	5.52
107	DRC/CC-ICP-MS	6.96	0.72	3.09	3.69	5.75
110	ICP-MS	8.13	0.93	3.68	4.55	6.51
116	ICP-MS/MS	6.67	0.590	2.95	3.75	5.48
147	DRC/CC-ICP-MS	8.27	0.900	3.68	4.70	7.07
264	ICP-MS	8.01	0.74	2.96	4.59	5.11
293	DRC/CC-ICP-MS	7.67	0.8	3.42	4.2	6.17
324	ICP-MS	7.180	1.080	3.330	4.156	5.845
391	ICP-MS	5.566	1.113	2.45	2.716	4.196
597	ICP-MS/MS	7.53	0.732	3.26	4.06	5.72
605	ICP-MS	6.63	0.832	3.59	3.79	6.51

Based on the grading criteria for Cr in Urine, 100% of results were satisfactory, with 0 of the 11 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #2, 2023: Summary Figures

## Urine Cr



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

±3 µg/L or ±20% around the target value, whichever is greater; thus, it is fixed at ±3 µg/L at concentrations less than or equal to 15 µg/L.



## Results for Event #2, 2023: Summary Statistics

	Urine Hg (µg/L)				
	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
<b>Target (Robust Mean (x*))</b>	5.2	3.22	25.7	2.24	0.53
<b>Upper Limit</b>	8.2	6.22	33.4	5.24	3.53
<b>Lower Limit</b>	2.2	0.22	18.0	0.00	0.00
<b>Robust SD (s*)</b>	0.6	0.20	0.6	0.23	0.06
<b>Robust RSD (%)</b>	12	6.2	2.3	10	11
<b>Number of Sample Measurements (N)</b>	12	12	12	12	10
<b>Standard Uncertainty (u)</b>	0.2	0.07	0.2	0.08	0.02

The acceptable range is based on quality specifications:  $\pm 3 \mu\text{g/L}$  or  $\pm 30\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 3 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ . These quality specifications were established by New York State Department of Health's Wadsworth Center, the PT Program organizer.



## Results for Event #2, 2023: Performance of Participating Laboratories






		Urine Hg (µg/L)				
Lab Code	Method	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
	<b>Target</b>	<b>5.2</b>	<b>3.22</b>	<b>25.7</b>	<b>2.24</b>	<b>0.53</b>
103	ICP-MS/MS	5.53	3.42	26.9	2.34	0.540
107	DRC/CC-ICP-MS	4.63	3.15	25.36	1.92	0.52
110	ICP-MS	5.29	3.02	25.9	2.12	0.391
116	ICP-MS/MS	4.66	3.14	23.3	1.83	0.536
147	ICP-MS	4.95	3.21	25.7	2.19	0.554
264	ICP-MS	6.69	3.37	25.35	2.45	0.44
293	DRC/CC-ICP-MS	5.21	3.26	26.04	2.27	0.59
391	ICP-MS	3.131	1.703	13.226 ↓	0.832	0.253
442	ICP-MS/MS	5.25	3.01	25.7	2.34	0.547
597	ICP-MS/MS	5.67	3.37	25.4	2.39	0.649
605	ICP-MS	5.42	3.38	27.6	2.27	<1.00
606	ICP-MS/MS	5.79	3.44	28.5	2.68	<1.00

Based on the grading criteria for Hg in Urine, 98% of results were satisfactory, with 0 of the 12 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #2, 2023: Summary Figures

## Urine Hg



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

±3 µg/L or ±30% around the target value, whichever is greater; thus, it is fixed at ±3 µg/L at concentrations less than or equal to 10 µg/L.



## Results for Event #2, 2023: Summary Statistics

	Urine Mn (µg/L)				
	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
<b>Target (Robust Mean (x*))</b>	4.9	8.2	3.39	1.45	0.63
<b>Upper Limit</b>	6.1	10.3	4.24	2.00	1.18
<b>Lower Limit</b>	3.7	6.2	2.54	0.90	0.08
<b>Robust SD (s*)</b>	0.3	0.6	0.19	0.12	0.08
<b>Robust RSD (%)</b>	6.5	7.3	5.6	8.3	13
<b>Number of Sample Measurements (N)</b>	13	13	13	13	12
<b>Standard Uncertainty (u)</b>	0.1	0.2	0.07	0.04	0.03

The acceptable range is based on quality specifications:  $\pm 0.55 \mu\text{g/L}$  or  $\pm 25\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 0.55 \mu\text{g/L}$  at concentrations less than or equal to  $2.2 \mu\text{g/L}$ . Quality specifications for Mn are consistent with those used by other External Quality Assessment Schemes for trace elements. (Praamsma M, et al. An assessment of clinical laboratory performance for the determination of manganese in blood and urine. Clinical Chemistry and Laboratory Medicine.2016; 54(12): 1921-1928).



### Results for Event #2, 2023: Performance of Participating Laboratories

Lab Code	Method	Urine Mn (µg/L)				
		UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
	<b>Target</b>	<b>4.9</b>	<b>8.2</b>	<b>3.39</b>	<b>1.45</b>	<b>0.63</b>
103	ICP-MS/MS	4.98	8.55	3.52	1.35	0.619
107	DRC/CC-ICP-MS	4.432	7.275	3.007	1.316	0.545
110	DRC/CC-ICP-MS	5.02	8.60	3.52	1.45	0.63
116	ICP-MS/MS	4.28	7.25	2.91	1.37	0.524
147	DRC/CC-ICP-MS	5.38	9.62	3.90	1.69	0.786
220	DRC/CC-ICP-MS	4.96	8.43	3.42	1.49	0.645
264	ICP-MS	5.02	7.24	2.93	1.55	0.56
293	DRC/CC-ICP-MS	5	8.73	3.53	1.46	0.71
324	ICP-MS	4.756	7.901	3.326	1.450	<1
391	ICP-MS	3.721	6.772	2.436 ↓	0.846 ↓	0.602
399	DRC/CC-ICP-MS	NR	NR	NR	NR	NR
597	ICP-MS/MS	4.74	8.33	3.45	1.44	0.702
605	ICP-MS	4.89	8.38	3.40	1.44	0.619
606	ICP-MS/MS	5.33	8.60	3.53	1.65	0.664






Based on the grading criteria for Mn in Urine, 97% of results were satisfactory, with 1 of the 14 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.





# Results for Event #2, 2023: Summary Figures

## Urine Mn



**Legend:**  
○ HHEAR Labs    ● Other Labs  
Horizontal purple line = assigned target value based on the robust mean of all laboratories.  
Gray area = acceptable range based on quality specifications:  
±0.55 µg/L or ±25% around the target value, whichever is greater; thus, it is fixed at ±0.55 µg/L at concentrations less than or equal to 2.2 µg/L.



## Results for Event #2, 2023: Summary Statistics

	Urine Pb (µg/L)				
	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
<b>Target (Robust Mean (x*))</b>	9.10	1.36	0.45	4.94	2.17
<b>Upper Limit</b>	10.92	2.36	1.45	5.94	3.17
<b>Lower Limit</b>	7.28	0.36	0.00	3.94	1.17
<b>Robust SD (s*)</b>	0.25	0.05	0.04	0.12	0.10
<b>Robust RSD (%)</b>	2.7	3.7	8.0	2.4	4.6
<b>Number of Sample Measurements (N)</b>	16	16	15	16	16
<b>Standard Uncertainty (u)</b>	0.08	0.02	0.01	0.04	0.03

The acceptable range is based on quality specifications:  $\pm 1 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1 \mu\text{g/L}$  at concentrations less than or equal to  $5 \mu\text{g/L}$ . These quality specifications are based on the same criteria used by the US Centers for Disease Control Prevention (CDC) for public health labs participating in the Laboratory Response Network (LRN) PT program for Toxic Metals.



### Results for Event #2, 2023: Performance of Participating Laboratories






Lab Code	Method	Urine Pb (µg/L)				
		UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
	<b>Target</b>	<b>9.10</b>	<b>1.36</b>	<b>0.45</b>	<b>4.94</b>	<b>2.17</b>
103	ICP-MS/MS	9.48	1.30	0.360	4.96	2.07
107	ICP-MS	9.287	1.380	0.469	5.016	2.229
110	DRC/CC-ICP-MS	8.91	1.37	0.46	4.83	2.17
116	ICP-MS/MS	8.31	1.27	0.390	4.63	2.01
147	ICP-MS	8.95	1.34	0.485	4.89	2.11
220	ICP-MS	8.81	1.36	0.429	4.90	2.15
264	ICP-MS	9.53	1.55	0.51	5.15	2.30
293	DRC/CC-ICP-MS	9.18	1.4	0.46	5.02	2.29
324	ICP-MS	9.239	1.369	<1	5.071	2.174
391	ICP-MS	5.612 ↓	1.065	0.282	2.693 ↓	1.26
399	ICP-MS/MS	9.22	1.46	0.490	5.06	2.26
442	ICP-MS/MS	9.04	1.34	0.448	4.95	2.11
597	ICP-MS/MS	9.19	1.37	0.467	4.93	2.20
605	ICP-MS	8.98	1.35	0.433	4.85	2.16
606	ICP-MS/MS	9.13	1.33	0.473	4.93	2.11
686	ICP-MS	9.15	1.49	0.449	4.99	2.27

Based on the grading criteria for Pb in Urine, 98% of results were satisfactory, with 1 of the 16 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #2, 2023: Summary Figures

## Urine Pb



**Legend:**  
 ○ HHEAR Labs    ● Other Labs  
 Horizontal purple line = assigned target value based on the robust mean of all laboratories.  
 Gray area = acceptable range based on quality specifications:  
 $\pm 1 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1 \mu\text{g/L}$  at concentrations less than or equal to  $5 \mu\text{g/L}$ .



## Results for Event #2, 2023: Summary Statistics

	Urine TI ( $\mu\text{g/L}$ )				
	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
<b>Target (Robust Mean (<math>x^*</math>))</b>	2.02	1.50	3.33	0.862	0.300
<b>Upper Limit</b>	2.42	1.80	4.00	1.062	0.500
<b>Lower Limit</b>	1.62	1.20	2.66	0.662	0.100
<b>Robust SD (<math>s^*</math>)</b>	0.05	0.05	0.12	0.021	0.009
<b>Robust RSD (%)</b>	2.5	3.3	3.6	2.4	2.9
<b>Number of Sample Measurements (N)</b>	13	13	13	13	13
<b>Standard Uncertainty (<math>u</math>)</b>	0.02	0.02	0.04	0.007	0.003

The acceptable range is based on quality specifications:  $\pm 0.2 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 0.2 \mu\text{g/L}$  at concentrations less than or equal to  $1 \mu\text{g/L}$ . These quality specifications are based on the same criteria used by the US Centers for Disease Control Prevention (CDC) for public health labs participating in the Laboratory Response Network (LRN) PT program for Toxic Metals.



### Results for Event #2, 2023: Performance of Participating Laboratories






Lab Code	Method	Urine TI (µg/L)				
		UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
	<b>Target</b>	<b>2.02</b>	<b>1.50</b>	<b>3.33</b>	<b>0.862</b>	<b>0.300</b>
103	ICP-MS/MS	2.06	1.52	3.32	0.859	0.307
107	ICP-MS	2.041	1.488	3.337	0.846	0.299
110	ICP-MS	2.03	1.73	3.34	0.85	0.31
147	ICP-MS	2.02	1.56	3.43	0.893	0.323
220	ICP-MS	1.92	1.47	3.21	0.840	0.295
264	ICP-MS	2.12	1.54	3.49	0.90	0.31
293	DRC/CC-ICP-MS	2.14	1.62	3.58	0.91	0.32
391	ICP-MS	0.866 ↓	0.681 ↓	1.806 ↓	0.366 ↓	0.066 ↓
399	ICP-MS/MS	2.02	1.50	3.34	0.871	0.295
597	ICP-MS/MS	1.99	1.44	3.25	0.855	0.298
605	ICP-MS	2.01	1.49	3.29	0.854	0.296
606	ICP-MS/MS	1.98	1.47	3.24	0.872	0.293
686	ICP-MS	2.05	1.51	3.36	0.866	0.293

Based on the grading criteria for TI in Urine, 92% of results were satisfactory, with 1 of the 13 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #2, 2023: Summary Figures

## Urine TI



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the robust mean of all laboratories.

Gray area = acceptable range based on quality specifications:

$\pm 0.2 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 0.2 \mu\text{g/L}$  at concentrations less than or equal to  $1 \mu\text{g/L}$ .



## Results for Event #2, 2023: Summary Statistics

	Urine U (µg/L)				
	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
<b>Target (Robust Mean (x*))</b>	0.165	0.044	0.194	0.068	0.0074
<b>Upper Limit</b>	0.198	0.074	0.233	0.098	0.0374
<b>Lower Limit</b>	0.132	0.014	0.155	0.038	0.0000
<b>Robust SD (s*)</b>	0.006	0.003	0.011	0.004	0.0013
<b>Robust RSD (%)</b>	3.6	6.8	5.7	5.8	18
<b>Number of Sample Measurements (N)</b>	14	14	14	14	11
<b>Standard Uncertainty (u)</b>	0.002	0.001	0.004	0.001	0.0005

The acceptable range is based on quality specifications:  $\pm 0.03 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 0.03 \mu\text{g/L}$  at concentrations less than or equal to  $0.15 \mu\text{g/L}$ . These quality specifications are based on the same criteria used by the US Centers for Disease Control Prevention (CDC) for public health labs participating in the Laboratory Response Network (LRN) PT program for Toxic Metals.





### Results for Event #2, 2023: Performance of Participating Laboratories






Lab Code	Method	Urine U (µg/L)				
		UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
	<b>Target</b>	<b>0.165</b>	<b>0.044</b>	<b>0.194</b>	<b>0.068</b>	<b>0.0074</b>
103	ICP-MS/MS	0.170	0.0464	0.220	0.0721	<0.0200
107	ICP-MS	0.1777	0.0464	0.2019	0.0678	0.0077
110	ICP-MS	0.1658	0.0458	0.203	0.0720	0.0071
116	ICP-MS/MS	0.162	0.0430	0.188	0.0700	<0.0150
147	ICP-MS	0.162	0.0433	0.194	0.0674	0.00810
220	ICP-MS	0.167	0.044	0.203	0.069	0.006
264	ICP-MS	0.17	0.04	0.20	0.07	0.01
324	ICP-MS	0.1473	0.0376	0.1699	0.0576	0.0060
391	ICP-MS	0.134	0.097 ↑	0.127 ↓	0.08	0.05 ↑
399	ICP-MS/MS	0.152	0.046	0.187	0.065	0.007
597	ICP-MS/MS	0.165	0.0411	0.1864	0.0643	0.00617
605	ICP-MS	0.166	0.045	0.197	0.063	0.007
606	ICP-MS/MS	0.171	0.047	0.196	0.067	0.008
686	ICP-MS	0.166	0.0418	0.192	0.0668	<0.0150

Based on the grading criteria for U in Urine, 96% of results were satisfactory, with 1 of the 14 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #2, 2023: Summary Figures

## Urine U



### Legend:

- HHEAR Labs
- Other Labs
- Horizontal purple line = assigned target value based on the robust mean of all laboratories.
- Gray area = acceptable range based on quality specifications:  $\pm 0.03 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 0.03 \mu\text{g/L}$  at concentrations less than or equal to  $0.15 \mu\text{g/L}$ .



## Results for Event #2, 2023: Laboratory Data and Summary Statistics






Urine AI (µg/L)						
Lab Code	Method	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
147	ICP-MS	24.5	13.8	13.9	18.6	*13.8
264	ICP-MS	24.29	6.58	9.37	17.93	5.52
293	DRC/CC-ICP-MS	24.82	9.17	11.87	16.46	7.82
324	ICP-MS	20.723	11.385	11.209	15.661	9.352
597	ICP-MS/MS	20.6	7.12	9.71	13.3	6.39
Summary Statistics						
		UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>		23	10	11.2	16.4	7
<b>Arithmetic SD (s)</b>		2	3	1.8	2.1	2
<b>Arithmetic RSD (%)</b>		9.1	31	16	13	23
<b>Number of Sample Measurements (N)</b>		5	5	5	5	4

\*Denotes a statistical Outlier.



## Results for Event #2, 2023: Summary Figures

### Urine AI



#### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #2, 2023: Laboratory Data and Summary Statistics






Urine Cs (µg/L)						
Lab Code	Method	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
107	ICP-MS	10.69	1.14	3.02	1.44	4.07
110	ICP-MS	12.0	1.26	3.37	1.62	4.60
147	ICP-MS	10.2	1.16	3.15	1.48	4.25
220	ICP-MS	10.7	1.16	3.13	1.47	4.22
264	ICP-MS	8.74	0.89	2.47	1.19	3.30
399	ICP-MS/MS	10.6	1.17	3.10	1.48	4.15
597	ICP-MS/MS	10.0	1.12	2.94	1.35	4.00
605	ICP-MS	10.3	1.13	2.97	1.41	4.01
606	ICP-MS/MS	10.1	1.07	2.88	1.40	3.81
Summary Statistics						
		UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>		10.4	1.12	3.00	1.43	4.0
<b>Arithmetic SD (s)</b>		0.9	0.10	0.25	0.12	0.4
<b>Arithmetic RSD (%)</b>		8.7	8.9	8.3	8.4	8.6
<b>Number of Sample Measurements (N)</b>		9	9	9	9	9

\*Denotes a statistical Outlier.



## Results for Event #2, 2023: Summary Figures

### Urine Cs



#### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #2, 2023: Laboratory Data and Summary Statistics

### Urine Cu (µg/L)

Lab Code	Method	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
110	ICP-MS	17.0	*24.2	66.4	42.7	106
116	ICP-MS/MS	10.2	12.0	55.1	37.8	90.9
147	ICP-MS	*20.8	17.3	64.2	46.7	100
264	ICP-MS	11.36	11.83	50.40	41.48	82.99
293	DRC/CC-ICP-MS	10.81	12.71	59.12	38.14	95.36
324	ICP-MS	11.585	13.090	58.721	39.752	96.080
391	ICP-MS	7.024	9.372	46.559	21.493	76.995
597	ICP-MS/MS	10.6	12.6	57.2	36.7	91.4

### Summary Statistics






	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
Arithmetic Mean ( $\bar{x}$ )	11	13	57	38	92
Arithmetic SD (s)	3	2	7	7	9
Arithmetic RSD (%)	26	19	12	18	9.8
Number of Sample Measurements (N)	7	7	8	8	8

\*Denotes a statistical Outlier.



## Results for Event #2, 2023: Summary Figures

### Urine Cu



#### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.





## Results for Event #2, 2023: Laboratory Data and Summary Statistics

### Urine Mo (µg/L)

Lab Code	Method	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
103	ICP-MS/MS	94.8	38.0	9.70	36.4	45.6
107	ICP-MS	97.59	37.63	9.44	36.24	45.57
110	ICP-MS	99.2	39.5	10.3	37.9	48.7
147	ICP-MS	90.2	35.7	9.39	34.5	43.9
220	ICP-MS	95.8	39.9	9.00	36.4	46.1
264	ICP-MS	73.83	27.04	6.72	27.28	33.18
293	DRC/CC-ICP-MS	93.29	37.98	10.34	36.95	45.37
324	ICP-MS	92.666	37.400	9.692	35.583	45.930
399	ICP-MS/MS	92.6	37.3	9.58	35.7	45.3
597	ICP-MS/MS	80.9	32.2	8.46	30.5	39.8
605	ICP-MS	91.4	36.2	9.01	34.7	43.5
606	ICP-MS/MS	94.4	37.6	9.67	36.2	45.3

### Summary Statistics






	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
<b>Robust Mean (x*)</b>	93	37.3	9.5	35.7	45.1
<b>Robust SD (s*)</b>	4	1.6	0.6	1.2	1.2
<b>Robust RSD (%)</b>	4.1	4.3	6.3	3.4	2.7
<b>Number of Sample Measurements (N)</b>	12	12	12	12	12
<b>Standard Uncertainty (u)</b>	1	0.6	0.2	0.4	0.4

\*Denotes a statistical Outlier.



# Results for Event #2, 2023: Summary Figures

## Urine Mo



### Legend:

- HHEAR Labs    ● Other Labs
- Horizontal purple line = robust mean of all laboratories.
- Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #2, 2023: Laboratory Data and Summary Statistics

### Urine Ni (µg/L)

Lab Code	Method	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
103	ICP-MS/MS	2.48	6.65	4.09	<0.600	2.44
107	DRC/CC-ICP-MS	2.09	6.11	3.30	0.43	2.28
110	ICP-MS	3.69	7.01	4.19	*1.65	3.33
147	ICP-MS	2.25	6.40	3.70	0.611	2.50
264	ICP-MS	2.47	5.80	3.23	0.77	2.34
293	DRC/CC-ICP-MS	2.64	7.01	4.08	0.9	2.83
324	ICP-MS	2.851	6.357	3.728	1.234	2.842
391	ICP-MS	1.602	4.997	2.548	0.444	1.848
442	DRC/CC-ICP-MS	2.43	6.85	3.88	0.596	2.78
597	ICP-MS/MS	2.75	6.92	4.18	1.08	3.05
605	ICP-MS	2.24	6.65	3.68	0.626	2.60

### Summary Statistics

	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
<b>Robust Mean (x*)</b>	2.5	6.5	3.7	0.8	2.6
<b>Robust SD (s*)</b>	0.4	0.4	0.4	0.3	0.4
<b>Robust RSD (%)</b>	15	6.2	11	43	15
<b>Number of Sample Measurements (N)</b>	11	11	11	10	11
<b>Standard Uncertainty (u)</b>	0.1	0.2	0.2	0.1	0.1






\*Denotes a statistical Outlier.

An arithmetic mean, SD, RSD and n are provided for sample UE23-09.



# Results for Event #2, 2023: Summary Figures

## Urine Ni



### Legend:

- HHEAR Labs    ● Other Labs
- Horizontal purple line = robust mean of all laboratories.
- Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #2, 2023: Laboratory Data and Summary Statistics

### Urine Pt (µg/L)

Lab Code	Method	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
107	ICP-MS	3.7000	1.0898	0.1797	0.3428	0.8389
110	ICP-MS	3.68	1.11	0.172	0.336	0.849
220	ICP-MS	3.61	1.14	0.187	0.339	0.861
264	ICP-MS	3.69	1.12	0.19	0.36	0.85
293	DRC/CC-ICP-MS	3.68	1.14	0.21	0.34	0.87
605	ICP-MS	3.55	1.08	0.165	0.326	0.813

### Summary Statistics






	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>	3.65	1.11	0.184	0.341	0.847
<b>Arithmetic SD (s)</b>	0.06	0.03	0.016	0.011	0.020
<b>Arithmetic RSD (%)</b>	1.6	2.2	8.7	3.2	2.4
<b>Number of Sample Measurements (N)</b>	6	6	6	6	6

\*Denotes a statistical Outlier.



## Results for Event #2, 2023: Summary Figures

### Urine Pt



#### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #2, 2023: Laboratory Data and Summary Statistics

### Urine Sb (µg/L)

Lab Code	Method	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
103	ICP-MS/MS	1.11	0.276	0.798	0.0727	1.95
107	ICP-MS	1.161	0.307	0.841	0.101	2.196
110	ICP-MS	1.30	0.394	0.921	0.096	2.49
147	ICP-MS	1.19	0.319	0.875	0.0902	2.19
220	ICP-MS	1.17	0.328	0.893	0.107	2.37
264	ICP-MS	1.06	0.25	0.74	0.08	1.99
293	DRC/CC-ICP-MS	1.29	0.35	1.02	0.12	2.34
324	ICP-MS	1.253	<1	<1	<1	2.315
399	ICP-MS/MS	1.22	0.318	0.881	0.094	2.33
597	ICP-MS/MS	1.15	0.318	0.822	0.118	2.14
605	ICP-MS	1.22	<0.800	0.818	<0.800	2.26
606	ICP-MS/MS	1.17	0.319	0.869	0.083	2.21

### Summary Statistics






	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
<b>Robust Mean (x*)</b>	1.19	0.318	0.86	0.096	2.24
<b>Robust SD (s*)</b>	0.07	0.019	0.06	0.018	0.15
<b>Robust RSD (%)</b>	5.9	5.8	7.3	19	6.7
<b>Number of Sample Measurements (N)</b>	12	10	11	10	12
<b>Standard Uncertainty (u)</b>	0.02	0.007	0.02	0.007	0.05

\*Denotes a statistical Outlier.



# Results for Event #2, 2023: Summary Figures

## Urine Sb



### Legend:

- HHEAR Labs    ● Other Labs
- Horizontal purple line = robust mean of all laboratories.
- Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.





## Results for Event #2, 2023: Laboratory Data and Summary Statistics

### Urine Se (µg/L)

Lab Code	Method	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
103	ICP-MS/MS	147	47.6	192	94.8	232
110	DRC/CC-ICP-MS	*169	57.2	200	103	234
147	ICP-MS	143	44.9	183	91.6	209
293	DRC/CC-ICP-MS	144.55	46.6	188.78	94	228.28
597	ICP-MS/MS	143	47.9	181	92.1	214

### Summary Statistics






	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>	144.4	49	189	95	223
<b>Arithmetic SD (s)</b>	1.9	5	8	5	11
<b>Arithmetic RSD (%)</b>	1.3	10	4.2	5.3	4.9
<b>Number of Sample Measurements (N)</b>	4	5	5	5	5

\*Denotes a statistical Outlier.



# Results for Event #2, 2023: Summary Figures

## Urine Se



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #2, 2023: Laboratory Data and Summary Statistics






Urine Sn (µg/L)						
Lab Code	Method	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
107	ICP-MS	3.64	3.26	0.69	2.09	0.60
110	ICP-MS	3.87	3.62	0.55	2.68	0.53
147	ICP-MS	2.91	2.85	0.441	1.94	0.397
220	ICP-MS	2.19	2.78	<0.4	1.38	<0.4
264	ICP-MS	2.42	2.53	0.54	1.57	0.49
324	ICP-MS	3.179	3.293	<1	2.199	<1
399	ICP-MS/MS	4.13	4.06	0.720	2.87	0.656
597	ICP-MS/MS	2.32	2.76	0.561	1.29	0.555
605	ICP-MS	3.27	3.11	<0.900	2.3	<0.900
Summary Statistics						
		UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
Arithmetic Mean ( $\bar{x}$ )		3.1	3.1	0.58	2.0	0.54
Arithmetic SD (s)		0.7	0.5	0.10	0.6	0.09
Arithmetic RSD (%)		23	16	17	30	17
Number of Sample Measurements (N)		9	9	6	9	6

\*Denotes a statistical Outlier.



## Results for Event #2, 2023: Summary Figures

### Urine Sn



#### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #2, 2023: Laboratory Data and Summary Statistics

### Urine Sr (µg/L)

Lab Code	Method	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
103	ICP-MS/MS	62.2	116	258	67.9	44.2
107	ICP-MS	65.3	120.4	280.9	71.4	46.4
200	ICP-MS	61.6	107	241	64.4	40.2
220	ICP-MS	61.7	110	253	69.5	45.5
264	ICP-MS	50.60	92.14	264.54	55.95	36.72
399	DRC/CC-ICP-MS	61.7	112	266	69.5	44.8
597	ICP-MS/MS	58.4	109	242	64.6	43.6
605	ICP-MS	61.8	115	256	69.0	45.1

### Summary Statistics






	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>	60	110	258	67	43
<b>Arithmetic SD (s)</b>	4	8	13	5	3
<b>Arithmetic RSD (%)</b>	6.7	7.3	5.1	7.5	7.4
<b>Number of Sample Measurements (N)</b>	8	8	8	8	8

\*Denotes a statistical Outlier.



# Results for Event #2, 2023: Summary Figures

## Urine Sr



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #2, 2023: Laboratory Data and Summary Statistics






Urine V (µg/L)						
Lab Code	Method	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
116	ICP-MS/MS	0.825	3.00	1.23	0.418	1.83
147	DRC/CC-ICP-MS	0.974	3.47	1.38	0.468	2.26
293	DRC/CC-ICP-MS	0.89	3.2	1.23	0.44	2.08
597	ICP-MS/MS	0.903	3.22	1.32	0.439	2.07
605	ICP-MS	0.868	3.40	1.32	0.321	2.08
Summary Statistics						
		UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>		0.89	3.26	1.30	0.42	2.06
<b>Arithmetic SD (s)</b>		0.05	0.18	0.07	0.06	0.15
<b>Arithmetic RSD (%)</b>		5.6	5.5	5.4	14	7.3
<b>Number of Sample Measurements (N)</b>		5	5	5	5	5

\*Denotes a statistical Outlier.



## Results for Event #2, 2023: Summary Figures

### Urine V



#### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.





## Results for Event #2, 2023: Laboratory Data and Summary Statistics

### Urine W (µg/L)

Lab Code	Method	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
107	ICP-MS	1.173	3.457	0.788	1.986	0.480
110	ICP-MS	1.14	3.67	0.834	2.11	0.481
147	ICP-MS	1.11	3.53	0.789	2.06	0.489
200	ICP-MS	1.23	2.78	0.66	1.95	0.37
220	ICP-MS	1.12	3.63	0.816	2.04	0.489
264	ICP-MS	1.10	3.46	0.78	1.97	0.42
324	ICP-MS	1.117	3.622	<1	1.997	<1
399	ICP-MS/MS	1.10	3.56	0.811	1.99	0.474
597	ICP-MS/MS	1.13	3.48	0.757	2.01	0.479
605	ICP-MS	1.10	3.40	0.770	1.94	0.444
606	ICP-MS/MS	1.09	3.49	0.799	1.97	0.439

### Summary Statistics






	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
<b>Robust Mean (x*)</b>	1.12	3.51	0.79	1.99	0.469
<b>Robust SD (s*)</b>	0.03	0.11	0.03	0.04	0.020
<b>Robust RSD (%)</b>	2.3	3.1	3.9	1.9	4.3
<b>Number of Sample Measurements (N)</b>	11	11	10	11	10
<b>Standard Uncertainty (u)</b>	0.01	0.04	0.01	0.01	0.008

\*Denotes a statistical Outlier.



# Results for Event #2, 2023: Summary Figures

## Urine W



### Legend:

- HHEAR Labs
- Other Labs
- Horizontal purple line = robust mean of all laboratories.
- Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #2, 2023: Laboratory Data and Summary Statistics

### Urine Zn (µg/L)

Lab Code	Method	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
110	ICP-MS	595	113	165	814	244
147	ICP-MS	617	84.3	168	843	243
264	ICP-MS	648.79	67.42	139.22	907.00	207.60
293	DRC/CC-ICP-MS	615.69	73.2	154.25	827.45	227.45
324	ICP-MS	616.319	75.362	156.103	835.937	231.962
391	ICP-MS	683.81	71.977	160.264	*106.021	239.173
597	ICP-MS/MS	579	66.7	147	772	211

### Summary Statistics






	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
Arithmetic Mean ( $\bar{x}$ )	622	79	156	830	229
Arithmetic SD (s)	35	16	10	40	15
Arithmetic RSD (%)	5.6	20	6.4	4.8	6.6
Number of Sample Measurements (N)	7	7	7	6	7

\*Denotes a statistical Outlier.



## Results for Event #2, 2023: Summary Figures

### Urine Zn



#### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #2, 2023: Laboratory Data and Summary Statistics

### Urine Bi (µg/L)

Lab Code	Method	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
147	ICP-MS	0.0794	0.0794	0.0794	0.0794	0.0794
264	ICP-MS	0.07	0.05	0.04	0.03	0.04
597	ICP-MS/MS	<0.0245	<0.0245	<0.0245	<0.0245	<0.0245

### Summary Statistics

	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
Arithmetic Mean ( $\bar{x}$ )	0.075	0.07	NA	NA	NA
Arithmetic SD (s)	0.007	0.02	NA	NA	NA
Arithmetic RSD (%)	9.3	32	NA	NA	NA
Number of Sample Measurements (N)	2	2	NA	NA	NA

\*Denotes a statistical Outlier.

Statistical data was not calculated for UE23-08, UE23-09 and UE23-10 based on a lack of consensus among participating labs.



## Results for Event #2, 2023: Laboratory Data and Summary Statistics

Urine I (µg/L)						
Lab Code	Method	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
110	ICP-MS	199	73.6	99.8	171	73.7
147	ICP-MS	197	73.2	96.6	178	69.4

Summary Statistics						
	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10	
Arithmetic Mean ( $\bar{x}$ )	198	73.4	98	175	72	
Arithmetic SD (s)	1	0.3	2	5	3	
Arithmetic RSD (%)	0.71	0.41	2.3	2.9	4.2	
Number of Sample Measurements (N)	2	2	2	2	2	

\*Denotes a statistical Outlier.



## Results for Event #2, 2023: Laboratory Data and Summary Statistics

### Urine Te (µg/L)

Lab Code	Method	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
110	ICP-MS	0.202	1.42	0.700	0.752	3.37
147	ICP-MS	0.218	1.42	0.685	0.698	3.37

### Summary Statistics

	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
Arithmetic Mean ( $\bar{x}$ )	0.210	1.42	0.693	0.72	3.37
Arithmetic SD (s)	0.011	0.00	0.011	0.04	0.00
Arithmetic RSD (%)	5.2	0.0	1.6	5.6	0.0
Number of Sample Measurements (N)	2	2	2	2	2

\*Denotes a statistical Outlier.



## Results for Event #2, 2023: Laboratory Data and Summary Statistics

### Urine Th (µg/L)

Lab Code	Method	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
147	ICP-MS	0.102	0.102	0.102	0.102	0.102
597	ICP-MS/MS	0.0491	0.0181	0.0244	0.0339	0.0103

### Summary Statistics

	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
Arithmetic Mean ( $\bar{x}$ )	NA	NA	NA	NA	NA
Arithmetic SD (s)	NA	NA	NA	NA	NA
Arithmetic RSD (%)	NA	NA	NA	NA	NA
Number of Sample Measurements (N)	NA	NA	NA	NA	NA

\*Denotes a statistical Outlier.

Statistical data was not calculated for UE23-06, UE23-07, UE23-08, UE23-09 and UE23-10 based on a lack of consensus among participating labs.





Results for Event #2, 2023:  
Additional Elements in Urine

Urine Ag (µg/L)

Lab Code	Method	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
147	ICP-MS	0.151	0.151	0.151	0.151	0.151

Urine B (µg/L)

Lab Code	Method	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
200	ICP-MS	1156	363	388	794	266

Urine Fe (µg/L)

Lab Code	Method	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
324	ICP-MS	12.405	5.800	4.853	8.253	6.193

Urine Li (µg/L)

Lab Code	Method	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
147	ICP-MS	14.2	4.77	5.88	11.6	3.12

Urine Mg (µg/L)

Lab Code	Method	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
597	ICP-MS/MS	38200	13300	12300	23500	14600

Urine Ti (µg/L)

Lab Code	Method	UE23-06	UE23-07	UE23-08	UE23-09	UE23-10
597	ICP-MS/MS	4.61	1.87	6.99	2.01	2.53



**Department  
of Health**

**Wadsworth  
Center**

**Event #2, 2023**

**Trace Elements in  
Serum**

**Wadsworth Center**  
NEW YORK STATE DEPARTMENT OF HEALTH  
*Trace Elements Laboratory*



## Event #2, 2023: Trace Elements in Serum

### PT Materials

Test materials were prepared from human serum obtained from Zen-Bio, Inc. The company certifies that these materials were tested by FDA approved methods and found to be negative for HIV 1Z2 and HIV-1 RNA, and non-reactive to HBsAg, HCV3 and STS. Units of serum were filtered into polypropylene containers through cheesecloth to remove particulates and supplemented with aluminum (Al), cobalt (Co), chromium (Cr), copper (Cu), selenium (Se), zinc (Zn), arsenic (As), beryllium (Be), cadmium (Cd), mercury (Hg), manganese (Mn), molybdenum (Mo), nickel (Ni), lead (Pb), platinum (Pt), antimony (Sb), tin (Sn), strontium (Sr), titanium (Ti), thallium (Tl), uranium (U), vanadium (V) and tungsten (W). PT samples were stored at -80°C until the week of the PT event, when they were thawed at 4°C prior to circulation to laboratories for analysis.

### Graded Elements

Six elements in serum are formally graded: Al, Co, Cr, Cu, Se, and Zn. Target values for the graded elements are assigned to these pools based on (a) the robust mean calculated from data reported by all laboratories, or (b) if a robust mean is not possible, the arithmetic mean after outlier deletion.

### Additional Elements

An additional 27 were reported by at least one participant: As, B, Ba, Be, Bi, Cd, Cs, Fe, Hg, I, Li, Mg, Mn, Mo, Ni, Pb, Pt, Sb, Sn, Sr, Te, Th, Ti, Tl, U, V, and W. These data are included here to provide a more complete characterization of the PT materials. All results reported by participant laboratories are tabulated and organized by lab code. The PT data are graphed for visual comparison purposes for all elements where at least five laboratories reported a value greater than the LOD. A statistical summary table is provided for samples where at least two comparable values were reported as above the LOD.

The summary statistics for the additional elements are provided for educational purposes only, i.e., no acceptable response is implied. However, it is expected that each laboratory would wish to investigate a potential source of bias if warranted by these data. Future events might result in additional elements becoming graded if a consensus can be reached regarding desired quality specifications.



## Results for Event #2, 2023: Summary Statistics

	Serum AI (µg/L)				
	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10
<b>Target (Arithmetic Mean (<math>\bar{x}</math>))</b>	34.3	21.2	9.1	75.4	13.9
<b>Upper Limit</b>	41.2	26.2	14.1	90.5	18.9
<b>Lower Limit</b>	27.4	16.2	4.1	60.3	8.9
<b>Arithmetic SD (s)</b>	0.9	1.1	2.2	2.9	1.5
<b>Arithmetic RSD (%)</b>	2.6	5.2	24	3.8	11
<b>Number of Sample Measurements (N)</b>	6	5	6	6	6

The acceptable range is based on quality specifications:  $\pm 5 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 5 \mu\text{g/L}$  at concentrations less than or equal to  $25 \mu\text{g/L}$ . These quality specifications were established by New York State Department of Health's Wadsworth Center, the PT Program organizer.



## Results for Event #2, 2023: Performance of Participating Laboratories






		Serum AI (µg/L)				
Lab Code	Method	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10
	<b>Target</b>	<b>34.3</b>	<b>21.2</b>	<b>9.1</b>	<b>75.4</b>	<b>13.9</b>
147	ETAAS-Z	34.4	21.6	10.7	75.8	16.0
264	ICP-MS	33.12	19.22	5.68	75.24	12.54
293	DRC/CC-ICP-MS	34.14	21.77	9.14	74.73	13.98
391	ETAAS-Z	33.54	*30.61 ↑	11.77	70.43	15.15
485	HR-ICP-MS	34.90	21.9	9.48	77.7	13.8
597	ICP-MS/MS	35.6	21.4	7.84	78.7	11.9

Based on the grading criteria for AI in Serum, 97% of results were satisfactory, with 0 of the 6 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



## Results for Event #2, 2023: Summary Figures

### Serum AI



#### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the arithmetic mean of all laboratories.  
Gray area = acceptable range based on quality specifications:

±5 µg/L or ±20% around the target value, whichever is greater; thus, it is fixed at ±5 µg/L at concentrations less than or equal to 25 µg/L.



## Results for Event #2, 2023: Summary Statistics

	Serum Co (µg/L)				
	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10
<b>Target (Arithmetic Mean (<math>\bar{x}</math>))</b>	20.5	1.93	0.750	2.54	9.5
<b>Upper Limit</b>	23.6	3.43	2.250	4.04	11.0
<b>Lower Limit</b>	17.4	0.43	0.000	1.04	8.0
<b>Arithmetic SD (s)</b>	0.6	0.06	0.025	0.16	0.5
<b>Arithmetic RSD (%)</b>	2.9	3.1	3.3	6.3	5.3
<b>Number of Sample Measurements (N)</b>	7	8	8	8	8

The acceptable range is based on quality specifications:  $\pm 1.5 \mu\text{g/L}$  or  $\pm 15\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 1.5 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ . These quality specifications were established based on discussions with the US FDA, and represent a consensus from a network of Trace Element PT program organizers



### Results for Event #2, 2023: Performance of Participating Laboratories

		Serum Co (µg/L)				
Lab Code	Method	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10
	<b>Target</b>	<b>20.5</b>	<b>1.93</b>	<b>0.750</b>	<b>2.54</b>	<b>9.5</b>
103	ICP-MS/MS	19.5	1.83	0.721	2.31	8.82
110	ICP-MS/MS	20.9	1.87	0.76	2.48	9.29
147	DRC/CC-ICP-MS	20.6	1.90	0.775	2.48	9.23
264	ICP-MS	21.40	2.01	0.75	2.60	9.90
293	DRC/CC-ICP-MS	20.6	1.97	0.77	2.63	9.61
442	DRC/CC-ICP-MS	20.1	1.93	0.705	2.45	8.97
485	HR-ICP-MS	20.7	1.98	0.77	2.53	9.44
597	ICP-MS/MS	*24.3 ↑	1.91	0.752	2.85	10.5






Based on the grading criteria for Co in Serum, 98% of results were satisfactory, with 0 of the 8 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.





# Results for Event #2, 2023: Summary Figures

## Serum Co



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the arithmetic mean of all laboratories. Gray area = acceptable range based on quality specifications:

±1.5 µg/L or ±15% around the target value, whichever is greater; thus, it is fixed at ±1.5 µg/L at concentrations less than or equal to 10 µg/L.



### Results for Event #2, 2023: Summary Statistics

	Serum Cr (µg/L)				
	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10
<b>Target (Arithmetic Mean (<math>\bar{x}</math>))</b>	3.1	7.9	4.60	1.9	0.87
<b>Upper Limit</b>	5.1	9.9	6.60	3.9	2.87
<b>Lower Limit</b>	1.1	5.9	2.60	0.0	0.00
<b>Arithmetic SD (s)</b>	0.4	0.3	0.23	0.3	0.15
<b>Arithmetic RSD (%)</b>	11	4.1	5.0	13	17
<b>Number of Sample Measurements (N)</b>	7	7	7	7	7

The acceptable range is based on quality specifications:  $\pm 2 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 2 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ . These quality specifications were established based on discussions with the US FDA, and represent a consensus from a network of Trace Element PT program organizers



## Results for Event #2, 2023: Performance of Participating Laboratories






		Serum Cr ( $\mu\text{g/L}$ )				
Lab Code	Method	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10
	<b>Target</b>	<b>3.1</b>	<b>7.9</b>	<b>4.60</b>	<b>1.9</b>	<b>0.87</b>
103	ICP-MS/MS	2.80	7.38	4.28	1.68	0.721
110	ICP-MS/MS	3.15	8.21	4.68	1.91	0.82
147	DRC/CC-ICP-MS	3.06	7.85	4.73	1.79	0.972
264	ICP-MS	2.54	7.78	4.30	1.55	0.70
293	DRC/CC-ICP-MS	3.19	8.39	4.89	2.13	0.88
485	HR-ICP-MS	3.12	7.90	4.60	1.90	0.86
597	ICP-MS/MS	3.67	8.02	4.72	2.28	1.14

Based on the grading criteria for Cr in Serum, 100% of results were satisfactory, with 0 of the 7 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



## Results for Event #2, 2023: Summary Figures

### Serum Cr



#### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the arithmetic mean of all laboratories.

Gray area = acceptable range based on quality specifications:

$\pm 2 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 2 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ .



### Results for Event #2, 2023: Summary Statistics

	Serum Cu (µg/L)				
	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10
<b>Target (Arithmetic Mean (<math>\bar{x}</math>))</b>	1313	1100	900	1950	1010
<b>Upper Limit</b>	1510	1270	1035	2240	1160
<b>Lower Limit</b>	1116	940	765	1660	860
<b>Arithmetic SD (s)</b>	29	50	22	170	60
<b>Arithmetic RSD (%)</b>	2.2	4.5	2.4	8.7	5.9
<b>Number of Sample Measurements (N)</b>	6	7	6	7	7

The acceptable range is based on quality specifications:  $\pm 95 \mu\text{g/L}$  or  $\pm 15\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 95 \mu\text{g/L}$  at concentrations less than or equal to  $635 \mu\text{g/L}$ . These quality specifications were established by New York State Department of Health's Wadsworth Center, the PT Program organizer.



### Results for Event #2, 2023: Performance of Participating Laboratories






		Serum Cu (µg/L)				
Lab Code	Method	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10
	<b>Target</b>	<b>1313</b>	<b>1100</b>	<b>900</b>	<b>1950</b>	<b>1010</b>
107	DRC/CC-ICP-MS	1318	1146	932	2044	1063
110	ICP-MS/MS	1330	1100	898	1900	986
147	DRC/CC-ICP-MS	1280	1109	894	1925	972
264	ICP-MS	1352	1147	913	1996	1038
293	DRC/CC-ICP-MS	1278	1074	865	1894	966
436	FAAS	1317	1000	*700 ↓	1651 ↓	953
597	ICP-MS/MS	*1530 ↑	1130	897	2210	1120

Based on the grading criteria for Cu in Serum, 91% of results were satisfactory, with 1 of the 7 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



## Results for Event #2, 2023: Summary Figures

### Serum Cu



#### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the arithmetic mean of all laboratories.

Gray area = acceptable range based on quality specifications:

$\pm 95 \mu\text{g/L}$  or  $\pm 15\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 95 \mu\text{g/L}$  at concentrations less than or equal to  $635 \mu\text{g/L}$ .



## Results for Event #2, 2023: Summary Statistics

	Serum Se (µg/L)				
	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10
<b>Target (Arithmetic Mean (<math>\bar{x}</math>))</b>	138	106	155	243	202
<b>Upper Limit</b>	166	127	186	292	242
<b>Lower Limit</b>	110	85	124	194	162
<b>Arithmetic SD (s)</b>	6	3	4	12	7
<b>Arithmetic RSD (%)</b>	4.3	3.0	2.7	4.9	3.5
<b>Number of Sample Measurements (N)</b>	8	8	8	8	8

The acceptable range is based on quality specifications:  $\pm 2$  µg/L or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 2$  µg/L at concentrations less than or equal to 10 µg/L. These quality specifications were established by New York State Department of Health's Wadsworth Center, the PT Program organizer.





## Results for Event #2, 2023: Performance of Participating Laboratories






Lab Code	Method	Serum Se (µg/L)				
		SE23-06	SE23-07	SE23-08	SE23-09	SE23-10
	<b>Target</b>	<b>138</b>	<b>106</b>	<b>155</b>	<b>243</b>	<b>202</b>
103	ICP-MS/MS	134	104	154	236	196
107	DRC/CC-ICP-MS	133.6	108.4	153.7	238.0	200.1
110	ICP-MS/MS	137	104	153	237	194
147	DRC/CC-ICP-MS	138	109	158	241	197
264	ICP-MS	138.59	107.27	154.26	246.68	203.28
293	DRC/CC-ICP-MS	139	109	159	251	208
436	A-7	130	100	160	227	201
597	ICP-MS/MS	150	104	147	265	214

Based on the grading criteria for Se in Serum, 100% of results were satisfactory, with 0 of the 8 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



# Results for Event #2, 2023: Summary Figures

## Serum Se



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the arithmetic mean of all laboratories.

Gray area = acceptable range based on quality specifications:

$\pm 2 \mu\text{g/L}$  or  $\pm 20\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 2 \mu\text{g/L}$  at concentrations less than or equal to  $10 \mu\text{g/L}$ .



## Results for Event #2, 2023: Summary Statistics

	Serum Zn (µg/L)				
	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10
<b>Target (Arithmetic Mean (<math>\bar{x}</math>))</b>	1160	1380	910	1580	630
<b>Upper Limit</b>	1330	1590	1050	1820	720
<b>Lower Limit</b>	990	1170	770	1340	540
<b>Arithmetic SD (s)</b>	100	60	40	120	50
<b>Arithmetic RSD (%)</b>	8.6	4.3	4.4	7.6	7.9
<b>Number of Sample Measurements (N)</b>	6	6	6	6	6

The acceptable range is based on quality specifications:  $\pm 15 \mu\text{g/L}$  or  $\pm 15\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 15 \mu\text{g/L}$  at concentrations less than or equal to  $100 \mu\text{g/L}$ . These quality specifications were established by New York State Department of Health's Wadsworth Center, the PT Program organizer.



## Results for Event #2, 2023: Performance of Participating Laboratories






		Serum Zn (µg/L)				
Lab Code	Method	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10
	<b>Target</b>	<b>1160</b>	<b>1380</b>	<b>910</b>	<b>1580</b>	<b>630</b>
107	DRC/CC-ICP-MS	1155	1453	968	1665	675
110	ICP-MS/MS	1150	1358	908	1560	605
147	DRC/CC-ICP-MS	1093	1333	882	1490	600
264	ICP-MS	1125	1363	893	1518	618
293	DRC/CC-ICP-MS	1078	1294	863	1464	582
597	ICP-MS/MS	1360 ↑	1450	953	1790	716

Based on the grading criteria for Zn in Serum, 97% of results were satisfactory, with 0 of the 6 laboratories reporting 2 or more of the 5 results outside of the acceptable ranges.



## Results for Event #2, 2023: Summary Figures

### Serum Zn



#### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = assigned target value based on the arithmetic mean of all laboratories.  
Gray area = acceptable range based on quality specifications:

$\pm 15 \mu\text{g/L}$  or  $\pm 15\%$  around the target value, whichever is greater; thus, it is fixed at  $\pm 15 \mu\text{g/L}$  at concentrations less than or equal to  $100 \mu\text{g/L}$ .



## Results for Event #2, 2023: Laboratory Data and Summary Statistics






Serum Mn (µg/L)						
Lab Code	Method	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10
103	ICP-MS/MS	3.32	14.0	2.25	7.52	5.57
110	ICP-MS/MS	3.51	13.8	2.36	7.46	5.5
147	DRC/CC-ICP-MS	3.62	14.7	2.49	7.75	5.78
264	ICP-MS	2.38	12.95	*1.46	6.36	4.61
293	DRC/CC-ICP-MS	3.430	14.56	2.39	8.19	5.780
597	ICP-MS/MS	4.00	14.4	2.43	8.61	6.38
Summary Statistics						
		SE23-06	SE23-07	SE23-08	SE23-09	SE23-10
<b>Arithmetic Mean (<math>\bar{x}</math>)</b>		3.4	14.1	2.38	7.6	5.6
<b>Arithmetic SD (s)</b>		0.5	0.6	0.09	0.8	0.6
<b>Arithmetic RSD (%)</b>		15	4.3	3.8	11	11
<b>Number of Sample Measurements (N)</b>		6	6	5	6	6

\*Denotes a statistical Outlier.



# Results for Event #2, 2023: Summary Figures

## Serum Mn



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area = ±2SD of the mean.

The mean and ±2SD of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #2, 2023: Laboratory Data and Summary Statistics

Serum Mo (µg/L)						
Lab Code	Method	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10
103	ICP-MS/MS	1.92	5.05	0.918	3.77	2.07
110	ICP-MS/MS	1.94	5.14	0.68	3.87	2.15
147	DRC/CC-ICP-MS	2.01	5.20	0.888	3.96	2.24
293	DRC/CC-ICP-MS	2.360	5.500	0.830	4.190	2.340
442	DRC/CC-ICP-MS	2.04	5.08	0.850	3.85	2.24
485	HR-ICP-MS	1.87	5.65	<1	4.03	2.24
597	ICP-MS/MS	2.03	4.72	0.652	3.98	2.21
Summary Statistics						
	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10	
Arithmetic Mean ( $\bar{x}$ )	2.02	5.2	0.80	3.95	2.21	
Arithmetic SD (s)	0.16	0.3	0.11	0.14	0.08	
Arithmetic RSD (%)	7.9	5.9	14	3.5	3.6	
Number of Sample Measurements (N)	7	7	6	7	7	






\*Denotes a statistical Outlier.





## Results for Event #2, 2023: Summary Figures

### Serum Mo



#### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #2, 2023: Laboratory Data and Summary Statistics

Serum Ni (µg/L)						
Lab Code	Method	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10
103	ICP-MS/MS	7.94	1.73	1.10	7.40	2.76
110	ICP-MS/MS	7.95	1.94	1.19	7.45	2.84
147	DRC/CC-ICP-MS	7.99	1.83	1.14	7.80	2.87
293	DRC/CC-ICP-MS	9.01	*2.62	*1.71	8.77	3.31
485	HR-ICP-MS	7.90	1.83	1.09	7.92	2.86
597	ICP-MS/MS	9.57	1.82	1.05	8.98	3.31






Summary Statistics						
	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10	
Arithmetic Mean ( $\bar{x}$ )	8.4	1.83	1.11	8.1	3.0	
Arithmetic SD (s)	0.7	0.07	0.05	0.7	0.3	
Arithmetic RSD (%)	8.3	3.8	4.5	8.6	8.4	
Number of Sample Measurements (N)	6	5	5	6	6	

\*Denotes a statistical Outlier.



# Results for Event #2, 2023: Summary Figures

## Serum Ni



### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #2, 2023: Laboratory Data and Summary Statistics

### Serum V (µg/L)

Lab Code	Method	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10
110	ICP-MS/MS	0.66	4.06	1.29	1.83	2.52
147	DRC/CC-ICP-MS	0.603	3.97	1.18	1.74	2.48
293	DRC/CC-ICP-MS	0.74	4.62	1.45	1.96	2.9
485	HR-ICP-MS	0.63	4.09	1.22	1.83	2.47
597	ICP-MS/MS	0.735	4.04	1.25	2.04	2.93

### Summary Statistics






	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10
Arithmetic Mean ( $\bar{x}$ )	0.67	4.2	1.28	1.88	2.7
Arithmetic SD (s)	0.06	0.3	0.10	0.12	0.2
Arithmetic RSD (%)	9.1	6.3	7.8	6.4	8.6
Number of Sample Measurements (N)	5	5	5	5	5

\*Denotes a statistical Outlier.



## Results for Event #2, 2023: Summary Figures

### Serum V



#### Legend:

○ HHEAR Labs    ● Other Labs

Horizontal purple line = arithmetic mean of all laboratories.

Gray area =  $\pm 2SD$  of the mean.

The mean and  $\pm 2SD$  of all laboratories are not intended to be quality specifications and are included for informational purposes only.



## Results for Event #2, 2023: Laboratory Data and Summary Statistics

Serum As ( $\mu\text{g/L}$ )						
Lab Code	Method	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10
103	ICP-MS/MS	25.1	1.80	5.53	2.70	13.1
110	ICP-MS/MS	25.0	1.78	5.32	2.69	12.5
147	DRC/CC-ICP-MS	24.2	1.80	5.44	2.76	12.7
597	ICP-MS/MS	26.6	1.72	5.01	3.05	14.2
Summary Statistics						
	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10	
Arithmetic Mean ( $\bar{x}$ )	25.2	1.77	5.3	2.80	13.1	
Arithmetic SD (s)	1.0	0.04	0.2	0.17	0.8	
Arithmetic RSD (%)	4.0	2.3	4.3	6.1	6.1	
Number of Sample Measurements (N)	4	4	4	4	4	

\*Denotes a statistical Outlier.



## Results for Event #2, 2023: Laboratory Data and Summary Statistics

### Serum Ba (µg/L)

Lab Code	Method	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10
110	ICP-MS/MS	1.80	1.10	1.21	3.64	3.52
147	ICP-MS	1.66	1.24	1.35	2.93	3.71
597	ICP-MS/MS	1.89	1.33	1.48	3.38	4.26

### Summary Statistics

	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10
Arithmetic Mean ( $\bar{x}$ )	1.78	1.22	1.35	3.3	3.8
Arithmetic SD (s)	0.12	0.12	0.14	0.4	0.4
Arithmetic RSD (%)	6.7	9.8	10	12	11
Number of Sample Measurements (N)	3	3	3	3	3

\*Denotes a statistical Outlier.



## Results for Event #2, 2023: Laboratory Data and Summary Statistics

Serum Be (µg/L)						
Lab Code	Method	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10
110	ICP-MS/MS	3.54	0.74	1.53	2.74	4.59
147	ICP-MS	3.23	0.800	1.35	2.69	4.24
293	ICP-MS	3.27	0.75	1.38	2.82	4.45
597	ICP-MS/MS	3.68	0.679	1.39	3.19	4.72
Summary Statistics						
	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10	
Arithmetic Mean ( $\bar{x}$ )	3.4	0.74	1.41	2.9	4.5	
Arithmetic SD (s)	0.2	0.05	0.08	0.2	0.2	
Arithmetic RSD (%)	6.4	6.8	5.7	7.9	4.7	
Number of Sample Measurements (N)	4	4	4	4	4	

\*Denotes a statistical Outlier.





## Results for Event #2, 2023: Laboratory Data and Summary Statistics

### Serum Bi (µg/L)

Lab Code	Method	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10
147	ICP-MS	0.0397	0.0397	0.0410	0.118	0.0563
597	ICP-MS/MS	<0.0237	<0.0237	0.0293	0.108	<0.0237

### Summary Statistics

	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10
Arithmetic Mean ( $\bar{x}$ )	NA	NA	0.035	0.113	NA
Arithmetic SD (s)	NA	NA	0.008	0.007	NA
Arithmetic RSD (%)	NA	NA	23	6.2	NA
Number of Sample Measurements (N)	NA	NA	2	2	NA

\*Denotes a statistical Outlier.

Statistical data was not calculated for SE23-06, SE23-07 and SE23-10 based on a lack of consensus among participating labs.



## Results for Event #2, 2023: Laboratory Data and Summary Statistics

Serum Cd (µg/L)						
Lab Code	Method	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10
103	ICP-MS/MS	3.29	0.351	1.42	0.717	2.50
110	ICP-MS/MS	3.32	0.34	1.38	0.72	2.46
147	ICP-MS	3.25	0.367	1.41	0.737	2.53
597	ICP-MS/MS	3.96	0.374	1.46	0.868	2.96
Summary Statistics						
	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10	
Arithmetic Mean ( $\bar{x}$ )	3.5	0.358	1.42	0.76	2.6	
Arithmetic SD (s)	0.3	0.015	0.03	0.07	0.2	
Arithmetic RSD (%)	8.6	4.2	2.1	9.2	8.8	
Number of Sample Measurements (N)	4	4	4	4	4	

\*Denotes a statistical Outlier.



## Results for Event #2, 2023: Laboratory Data and Summary Statistics

### Serum Cs (µg/L)

Lab Code	Method	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10
110	ICP-MS/MS	0.40	0.45	0.42	0.57	0.47
597	ICP-MS/MS	0.459	0.468	0.420	0.656	0.538

### Summary Statistics

	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10
Arithmetic Mean ( $\bar{x}$ )	0.43	0.459	0.42	0.61	0.50
Arithmetic SD (s)	0.04	0.013	0.00	0.06	0.05
Arithmetic RSD (%)	9.3	2.8	0.0	9.8	10
Number of Sample Measurements (N)	2	2	2	2	2

\*Denotes a statistical Outlier.



## Results for Event #2, 2023: Laboratory Data and Summary Statistics

### Serum Hg (µg/L)

Lab Code	Method	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10
103	ICP-MS/MS	3.29	1.04	1.60	2.95	6.76
110	ICP-MS/MS	3.17	0.79	1.49	2.92	6.31
597	ICP-MS/MS	3.68	0.890	1.51	3.31	7.21

### Summary Statistics

	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10
Arithmetic Mean ( $\bar{x}$ )	3.4	0.91	1.53	3.1	6.8
Arithmetic SD (s)	0.3	0.13	0.06	0.2	0.5
Arithmetic RSD (%)	8.8	14	3.9	7.2	7.4
Number of Sample Measurements (N)	3	3	3	3	3

\*Denotes a statistical Outlier.



## Results for Event #2, 2023: Laboratory Data and Summary Statistics

Serum Mg (µg/L)						
Lab Code	Method	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10
264	ICP-MS	21484.00	16923.00	19334.50	20654.18	18237.81
597	ICP-MS/MS	25200	16900	19500	23400	20200

Summary Statistics						
	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10	
Arithmetic Mean ( $\bar{x}$ )	23000	16912	19420	22000	19200	
Arithmetic SD (s)	3000	16	120	1900	1400	
Arithmetic RSD (%)	13	0.09	0.62	8.6	7.3	
Number of Sample Measurements (N)	2	2	2	2	2	

\*Denotes a statistical Outlier.



## Results for Event #2, 2023: Laboratory Data and Summary Statistics

Serum Pb (µg/L)						
Lab Code	Method	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10
103	ICP-MS/MS	3.72	10.4	2.45	5.42	1.16
110	ICP-MS/MS	3.85	9.80	2.37	6.02	1.16
597	ICP-MS/MS	4.58	10.8	2.42	6.37	1.47
Summary Statistics						
	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10	
Arithmetic Mean ( $\bar{x}$ )	4.1	10.3	2.41	5.9	1.3	
Arithmetic SD (s)	0.5	0.5	0.04	0.5	0.2	
Arithmetic RSD (%)	12	4.9	1.7	8.5	14	
Number of Sample Measurements (N)	3	3	3	3	3	

\*Denotes a statistical Outlier.



## Results for Event #2, 2023: Laboratory Data and Summary Statistics

Serum Pt (µg/L)						
Lab Code	Method	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10
110	ICP-MS/MS	1.24	0.101	0.365	0.529	0.820
264	ICP-MS	*0.87	<0.10	<0.10	*0.15	*0.49
293	DRC/CC-ICP-MS	1.32	0.11	0.39	0.57	0.84
Summary Statistics						
		SE23-06	SE23-07	SE23-08	SE23-09	SE23-10
Arithmetic Mean ( $\bar{x}$ )		1.28	0.106	0.38	0.55	0.80
Arithmetic SD (s)		0.06	0.006	0.02	0.03	0.01
Arithmetic RSD (%)		4.4	5.7	4.8	5.3	1.7
Number of Sample Measurements (N)		2	2	2	2	2

\*Denotes a statistical Outlier.



## Results for Event #2, 2023: Laboratory Data and Summary Statistics

Serum Sb (µg/L)						
Lab Code	Method	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10
103	ICP-MS/MS	0.662	1.31	2.19	0.850	4.36
110	ICP-MS/MS	0.65	1.25	2.25	0.88	4.16
147	ICP-MS	0.634	1.31	2.14	0.870	4.21
597	ICP-MS/MS	0.710	1.26	2.04	0.93	4.55
Summary Statistics						
	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10	
Arithmetic Mean ( $\bar{x}$ )	0.66	1.28	2.15	0.88	4.32	
Arithmetic SD (s)	0.03	0.03	0.09	0.03	0.18	
Arithmetic RSD (%)	4.5	2.3	4.2	3.4	4.2	
Number of Sample Measurements (N)	4	4	4	4	4	

\*Denotes a statistical Outlier.





## Results for Event #2, 2023: Laboratory Data and Summary Statistics

### Serum Sn (µg/L)

Lab Code	Method	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10
110	ICP-MS/MS	5.20	3.26	1.02	0.51	6.34
597	ICP-MS/MS	5.74	3.21	1.06	0.609	7.03

### Summary Statistics

	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10
Arithmetic Mean ( $\bar{x}$ )	5.5	3.23	1.04	0.56	6.7
Arithmetic SD (s)	0.4	0.04	0.03	0.07	0.5
Arithmetic RSD (%)	7.3	1.2	2.9	13	7.5
Number of Sample Measurements (N)	2	2	2	2	2

\*Denotes a statistical Outlier.



## Results for Event #2, 2023: Laboratory Data and Summary Statistics

Serum Sr ( $\mu\text{g/L}$ )						
Lab Code	Method	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10
103	ICP-MS/MS	391	237	180	449	239
200	ICP-MS	424	265	195	487	257
597	ICP-MS/MS	438	236	179	507	272

Summary Statistics						
	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10	
Arithmetic Mean ( $\bar{x}$ )	418	246	185	480	256	
Arithmetic SD (s)	24	16	9	30	17	
Arithmetic RSD (%)	5.7	6.5	4.9	6.3	6.6	
Number of Sample Measurements (N)	3	3	3	3	3	

\*Denotes a statistical Outlier.



## Results for Event #2, 2023: Laboratory Data and Summary Statistics

Serum Ti (µg/L)						
Lab Code	Method	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10
200	DRC/CC-ICP-MS	4.70	4.60	3.10	3.60	3.60
442	ICP-MS/MS	6.38	4.13	2.47	2.08	3.34
485	HR-ICP-MS	6.35	4.37	2.50	2.03	3.40
597	ICP-MS/MS	8.03	5.50	4.11	3.33	4.64

Summary Statistics					
	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10
Arithmetic Mean ( $\bar{x}$ )	6.4	4.7	3.0	2.8	3.7
Arithmetic SD (s)	1.4	0.6	0.8	0.8	0.6
Arithmetic RSD (%)	22	13	27	29	16
Number of Sample Measurements (N)	4	4	4	4	4

\*Denotes a statistical Outlier.



## Results for Event #2, 2023: Laboratory Data and Summary Statistics

Serum TI (µg/L)						
Lab Code	Method	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10
103	ICP-MS/MS	4.47	0.809	3.31	2.23	0.378
110	ICP-MS/MS	4.72	0.84	3.37	2.30	0.37
147	ICP-MS	4.49	0.815	3.28	2.20	0.363
597	ICP-MS/MS	5.64	0.869	3.49	2.72	0.442
Summary Statistics						
	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10	
Arithmetic Mean ( $\bar{x}$ )	4.8	0.83	3.36	2.4	0.39	
Arithmetic SD (s)	0.6	0.03	0.09	0.2	0.04	
Arithmetic RSD (%)	13	3.2	2.7	10	10	
Number of Sample Measurements (N)	4	4	4	4	4	

\*Denotes a statistical Outlier.



## Results for Event #2, 2023: Laboratory Data and Summary Statistics

### Serum U (µg/L)

Lab Code	Method	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10
103	ICP-MS/MS	0.0273	0.279	0.110	0.160	0.235
110	ICP-MS/MS	0.0266	0.276	0.101	0.157	0.236
597	ICP-MS/MS	0.0339	0.279	0.102	0.196	0.276

### Summary Statistics

	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10
Arithmetic Mean ( $\bar{x}$ )	0.029	0.278	0.104	0.17	0.25
Arithmetic SD (s)	0.004	0.002	0.005	0.02	0.02
Arithmetic RSD (%)	14	0.61	4.8	13	9.2
Number of Sample Measurements (N)	3	3	3	3	3

\*Denotes a statistical Outlier.



## Results for Event #2, 2023: Laboratory Data and Summary Statistics

### Serum W (µg/L)

Lab Code	Method	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10
110	ICP-MS/MS	1.19	0.24	0.57	2.50	0.73
200	ICP-MS	1.10	0.24	0.55	2.24	0.64
597	ICP-MS/MS	1.32	0.245	0.563	2.86	0.803

### Summary Statistics

	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10
Arithmetic Mean ( $\bar{x}$ )	1.20	0.242	0.561	2.5	0.72
Arithmetic SD (s)	0.11	0.003	0.010	0.3	0.08
Arithmetic RSD (%)	9.2	1.2	1.8	12	11
Number of Sample Measurements (N)	3	3	3	3	3

\*Denotes a statistical Outlier.



## Results for Event #2, 2023: Additional Elements in Serum

### Serum B (µg/L)

Lab Code	Method	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10
200	ICP-MS	82	70	30	89	39

### Serum Fe (µg/L)

Lab Code	Method	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10
264	ICP-MS	234.04	153.75	1226.39	627.00	735.61

### Serum I (µg/L)

Lab Code	Method	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10
147	ICP-MS	55.0	53.9	62.4	78.5	43.9

### Serum Li (µg/L)

Lab Code	Method	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10
147	ICP-MS	0.798	0.638	0.724	0.796	0.671

### Serum Te (µg/L)

Lab Code	Method	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10
110	ICP-MS/MS	0.01	0.01	0.00	0.01	0.00

### Serum Th (µg/L)

Lab Code	Method	SE23-06	SE23-07	SE23-08	SE23-09	SE23-10
597	ICP-MS/MS	<0.00892	<0.00892	<0.00892	<0.00892	<0.00892



Department  
of Health

Wadsworth  
Center

## References

1. ISO/FDIS-13528 (2005) Statistical methods for use in proficiency testing by interlaboratory comparisons. International Organization for Standardization, Geneva.
2. Taylor A, Angerer J, Arnaud J, Claeys F, Jones RL, Mazarrasa O, Mairiaux E, Menditto A, Parsons PJ, Patriarca M, Pineau A, Valkonen S, Weber J-P, Weykamp C. Occupational and environmental laboratory medicine: A network of EQAS organisers. Accreditation and Quality Assurance. 2006;11(8-9):435-9. PubMed PMID: 086NJ-0011.